

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO	<u>1613610</u>
SAMPLE DESCRIBED AS	24 hour sample Canister 54107	DATE RECEIVED	<u>05/16/2016</u>
SAMPLING LOCATION	Porter Ranch Castlebay Elementary School	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54107</u>
Pressure (Torr)	1398
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	418
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 5/19/16

Approved By: 

Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

AM
5-18-16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1613610</u>
24 hour sample Canister 54107	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.3	1.92	2.05	6.32	PASS
CH ₄ , ppmvC	0.3	2.02	2.04	0.71	PASS
CO ₂ , ppmvC	0.3	1.57	1.80	14.98	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.90	-6.36	PASS
NM/NEOC, ppmvC	0.4	2.03	2.18	7.65	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.3	1036	1062	2.47	PASS
CH ₄ , ppmvC	0.3	1068	1054	-1.31	PASS
CO ₂ , ppmvC	0.3	1022	1044	2.18	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1047	0.31	PASS
NM/NEOC, ppmvC	0.4	1024	1024	0.04	PASS

DATE ANALYZED 05/16/2016
REFERENCE NO: 16QM2AB
 QM2-101-103

WO #: 1613610



**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 5/16/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 5/16/16 Time: 10:20am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch / Castlebay Elem	54107	5/15/16 / 00:00 / 24 hours	-30"	+13

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>Ningning Ran</i>	SCAQMD Lab	5/16/2016	12:12

Remarks: 1:3 scheduled samples from trailer at Castlebay
Castlebay Lane Charter School – 19010 Castlebay Ln, Porter Ranch, CA