

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO	Cher Snyder, Assistant DEO Engineering & Compliance	LABORATORY NO	<u>1602724</u>
SAMPLE DESCRIBED AS	24 hour sample Canister 54572	DATE RECEIVED	<u>01/27/2016</u>
SAMPLING LOCATION	Castlebay Elementary Charter School 19010 Cotlebay Ln Porter Ranch, CA 91326	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54572</u>
Pressure (Torr)	740
CO, ppm	< 3
CH ₄ , ppm	3
CO ₂ , ppm	408
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 2/23/16

Approved By: 

Rudy Eden, Senior Manager
Laboratory Services
909-396-2391

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1602724</u>
24 hour sample Canister 54572	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC91340	MDL				
CO, ppmvC	0.3	10.40	10.30	-0.95	PASS
CH ₄ , ppmvC	0.3	10.17	10.88	6.95	PASS
CO ₂ , ppmvC	0.4	10.38	10.59	2.05	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	11.00	10.77	-2.05	PASS
NM/NEOC, ppmvC	0.2	10.64	10.61	-0.27	PASS


		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC135067	MDL				
CO, ppmvC	0.3	10100	10063	-0.36	PASS
CH ₄ , ppmvC	0.3	9950	10391	4.43	PASS
CO ₂ , ppmvC	0.4	10100	9998	-1.01	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	9940	10197	2.59	PASS
NM/NEOC, ppmvC	0.2	10000	10119	1.19	PASS

DATE ANALYZED	02/09/2016
REFERENCE NO:	16QM2AA
	QM2-101-73

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT
 INVOICE
 LAP AL
LABORATORY

WO #: 1602724



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: ~~1/21/16~~ 1-27-16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Robert Wimmer Date: 1/26/16 Time: 15:00

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch / Castlebay Elem	54169	1/22/16 / 00:00 / 24 hours	-30"	0*
		*Sample leaked enroute to hq		
Porter Ranch / Castlebay Elem	54572	1/25/16 / 00:00 / 24 hours	-30	12.5

Relinquished by	Received by	Firm/Agency	Date	Time
R Wimmer		SCAQMD Lab	1/27/16	915

Remarks: 1:3 scheduled samples from trailer at Castlebay
Castlebay Lane Charter School – 19010 Castlebay Ln, Porter Ranch, CA

Sample 54169 is inverted as received