

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS  
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

<b>TO</b>	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	<b>LABORATORY NO</b>	<u>1613328</u>
<b>SAMPLE DESCRIBED AS</b>	24 hour sample Canister 54645	<b>DATE RECEIVED</b>	<u>05/13/2016</u>
<b>SAMPLING LOCATION</b>	Highlands Community Pool Parking Lot	<b>FACILITY ID NO</b>	<u>NA</u>
		<b>REQUESTED BY</b>	<u>Sumner Wilson</u>
		<b>ST NO / PROJECT</b>	<u>NA</u>

---

Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54645</u>
Pressure (Torr)	740
CO, ppm	< 3
CH <sub>4</sub> , ppm	2
CO <sub>2</sub> , ppm	436
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 5/19/16

Approved By: *Solomon Teffera*

Solomon Teffera, Acting Senior Manager  
Laboratory Services  
909-396-2199

909-396-2199  
5.13.16

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT**  
**21865 Copley Drive, Diamond Bar CA 91765-1482**

**MONITORING & ANALYSIS**  
**QUALITY CONTROL SUMMARY**

Page 2 of 2

<b>SAMPLE DESCRIBED AS</b>	<b>LABORATORY NO</b>	<u>1613328</u>
24 hour sample Canister 54645	<b>REQUESTED BY</b>	<u>Sumner Wilson</u>

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

**QUALITY CONTROL -- End of run control recovery**

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.3	1.92	2.34	21.88	PASS
CH <sub>4</sub> , ppmvC	0.3	2.02	2.06	1.83	PASS
CO <sub>2</sub> , ppmvC	0.3	1.57	2.42	54.20	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	2.03	2.07	2.08	PASS
NM/NEOC, ppmvC	0.4	2.03	2.08	2.34	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.3	1036	1057	2.01	PASS
CH <sub>4</sub> , ppmvC	0.3	1068	1048	-1.89	PASS
CO <sub>2</sub> , ppmvC	0.3	1022	1042	1.93	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	1044	1046	0.15	PASS
NM/NEOC, ppmvC	0.4	1024	1022	-0.13	PASS

**DATE ANALYZED** 05/13/2016  
**REFERENCE NO:** 16QM2AB  
 QM2-102

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
SAMPLE ANALYSIS REQUEST**

DI  
 IN  
 L/  
LAB

WO #: 1613328



TO: SCAQMD LAB:  OTHER:  \_\_\_\_\_  
 SOURCE NAME: \_\_\_\_\_ Southern California Gas Co. \_\_\_\_\_ I.D. No. \_\_\_\_\_  
 Source Address: \_\_\_\_\_ 12801 Tampa Ave \_\_\_\_\_ City: \_\_\_\_\_ Porter Ranch \_\_\_\_\_  
 Mailing Address: \_\_\_\_\_ City: \_\_\_\_\_ Zip: \_\_\_\_\_ 91326 \_\_\_\_\_  
 Contact Person: \_\_\_\_\_ Title: \_\_\_\_\_ Tel: \_\_\_\_\_

Analysis Requested by: \_\_\_\_\_ Sumner Wilson \_\_\_\_\_ Date: \_\_\_\_\_ 5/13/16 \_\_\_\_\_  
 Approved by: \_\_\_\_\_ Jason Low \_\_\_\_\_ Office: \_\_\_\_\_ Budget #: \_\_\_\_\_ 44716 \_\_\_\_\_  
 REASON REQUESTED: Court/Hearing Board  Permit Pending  Hazardous/Toxic Spill   
 Suspected Violation Rule(s) \_\_\_\_\_ Other  \_\_\_\_\_

Sample Collected by: \_\_\_\_\_ Qian Zhou \_\_\_\_\_ Date: \_\_\_\_\_ 5/13/16 \_\_\_\_\_ Time: \_\_\_\_\_ 09:15am \_\_\_\_\_

**REQUESTED ANALYSIS: PAMS analysis**

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Highlands Community	54645	5-12-16 / 00:00 / 24 hours	-30"	+11.5

Relinquished by	Received by	Firm/Agency	Date	Time
<i>zhongqian</i>	<i>[Signature]</i>	SCAQMD Lab	5/13/16	11:57

Remarks 1:3 scheduled samples from trailer

Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)