

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO	<u>1615114</u>
SAMPLE DESCRIBED AS	24 hour sample Canister 54113	DATE RECEIVED	<u>05/31/2016</u>
SAMPLING LOCATION	Castlebay Elementary School	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54113</u>
Pressure (Torr)	1331
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	420
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 6/2/16

Approved By: 
Solomon Telfera, Acting Senior Manager
Laboratory Services
909-396-2199

1615114
6206

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1615114</u>
24 hour sample Canister 54113	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.09	8.68	PASS
CH ₄ , ppmvC	0.3	2.02	2.03	0.13	PASS
CO ₂ , ppmvC	0.4	1.57	1.85	17.99	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.96	-3.23	PASS
NM/NEOC, ppmvC	0.5	2.03	1.95	-3.88	PASS


		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1069	3.14	PASS
CH ₄ , ppmvC	0.3	1068	1065	-0.24	PASS
CO ₂ , ppmvC	0.4	1022	1052	2.93	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1051	0.66	PASS
NM/NEOC, ppmvC	0.5	1024	1033	0.92	PASS

DATE ANALYZED	06/01/2016
REFERENCE NO:	16QM2AC
	QM2-101-106

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRI
 INVOIC
 LAP AU
LABORA

WO #: 1615114



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 5/31/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 5/31/16 Time: 10:25am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch / Castlebay Elem	54113	5/30/16 / 00:00 / 24 hours	-30"	+11

Relinquished by	Received by	Firm/Agency	Date	Time
<i>z/ouqian</i>	<i>LN</i>	SCAQMD Lab	5/31/16	12:30

Remarks: 1:3 scheduled samples from trailer at Castlebay
Castlebay Lane Charter School – 19010 Castlebay Ln, Porter Ranch, CA
Left sampler, sn 3419