

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482


MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Jason Low, Ph.D., ADEO Monitoring & Analysis	LABORATORY NO	1631915
SAMPLE DESCRIBED AS	24 hour sample Canister 54715	DATE RECEIVED	11/18/2016
SAMPLING LOCATION	Porter Ranch Community School	FACILITY ID NO	NA
		REQUESTED BY	Sumner Wilson
		ST NO / PROJECT	NA
		REPORT GENERATED	

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54715</u>
Pressure (Torr)	760
CO, ppm	6
CH ₄ , ppm	2
CO ₂ , ppm	424
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 01/27/17 Reviewed By: 
Joan Niertit, Principal A.Q. Chemist
Laboratory Services

Date Approved: 2-7-17 Approved By: 
Aaron Katzenstein, Ph.D., Acting Senior Manager
Laboratory Services
909-396-2219

Handwritten notes:
760
1-20-17

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1631915</u>
24 hour sample Canister 54715	REQUESTED BY	<u>Sumner Wilson</u>
	DATE ANALYZED	<u>12/1/2016</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.03	5.45	PASS
CH ₄ , ppmvC	0.3	2.02	2.03	0.35	PASS
CO ₂ , ppmvC	0.4	1.57	4.91	213.01	FAIL
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	2.07	1.84	PASS
NM/NEOC, ppmvC	0.5	2.03	1.99	-1.98	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1039	0.28	PASS
CH ₄ , ppmvC	0.3	1068	1036	-3.04	PASS
CO ₂ , ppmvC	0.4	1022	1029	0.71	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1024	-1.94	PASS
NM/NEOC, ppmvC	0.5	1024	998	-2.52	PASS

REFERENCE NO: 16QM2AD
 QM2-101-139

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

WO #: 1631915

D.
 IN
 L.
LAB



TO: SCAQMD LAB: OTHER: _____
 SOURCE NAME: Southern California Gas Co. I.D. No. _____
 Source Address: 12801 Tampa Ave City: Porter Ranch
 Mailing Address: _____ City: _____ Zip: 91326
 Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 11/18/16
 Approved by: Jason Low Office: _____ Budget #: 44716
 REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
 Suspected Violation Rule(s) _____ Other _____

Sample Collected by: Kayla Jordan Date: 11/18/16 Time: 11:30am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch Community Elementary School (PRCS)	54715	11/14/16 / 00:00 / 24 hours	<-30"	9"

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Kayla Jordan</i>	<i>[Signature]</i>	SCAQMD Lab	11/18/2016	14:59

Remarks: 1:6 scheduled samples from trailer at PRCS
 Porter Ranch Community School Elementary – 12450 Mason Ave, Porter Ranch, CA 91326
 GPS (34.293369, -118.580505)
 Right Sampler, sn 3900