

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS  
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO Jason Low, Ph.D., ADEO  
Monitoring & Analysis

LABORATORY NO 1633127

DATE RECEIVED 12/01/2017

SAMPLE DESCRIBED AS  
24 hour sample  
Canister 54680

FACILITY ID NO NA

REQUESTED BY Sumner Wilson

SAMPLING LOCATION  
Reseda Station  
18328 Gault St  
Los Angeles, CA 91335

ST NO / PROJECT NA

REPORT GENERATED 1/13/2017

Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54680</u>
Pressure (Torr)	677
CO, ppm	< 3
CH <sub>4</sub> , ppm	3
CO <sub>2</sub> , ppm	543
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 1-13-17

Reviewed By: Joan Niertit For JEN  
Joan Niertit, Principal A.Q. Chemist  
Laboratory Services

Date Approved: 1-20-17

Approved By: Aaron Katzenstein  
Aaron Katzenstein, Ph.D., Sr. Manager  
Laboratory Services  
909-396-2219

2017  
1-13-17

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT**  
**21865 Copley Drive, Diamond Bar CA 91765-1482**

**MONITORING & ANALYSIS**  
**QUALITY CONTROL SUMMARY**

Page 2 of 2

<b>SAMPLE DESCRIBED AS</b>	<b>LABORATORY NO</b>	<u>1633127</u>
24 hour sample Canister 54680	<b>REQUESTED BY</b>	<u>Sumner Wilson</u>
	<b>DATE ANALYZED</b>	<u>1/3/2017</u>

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

**QUALITY CONTROL -- End of run control recovery**

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.10	9.17	PASS
CH <sub>4</sub> , ppmvC	0.3	2.02	2.03	0.39	PASS
CO <sub>2</sub> , ppmvC	0.4	1.57	1.98	26.06	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	2.03	1.99	-2.04	PASS
NM/NEOC, ppmvC	0.5	2.03	2.17	6.76	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1052	1.58	PASS
CH <sub>4</sub> , ppmvC	0.3	1068	1049	-1.73	PASS
CO <sub>2</sub> , ppmvC	0.4	1022	1034	1.19	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	1044	1035	-0.87	PASS
NM/NEOC, ppmvC	0.5	1024	1010	-1.36	PASS

**REFERENCE NO:** 16QM2AD  
 QM2-101-133



**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
SAMPLE ANALYSIS REQUEST**

DISTR  
 INVOI  
 LAP A  
 LABOR

WO #: 1633127



TO: SCAQMD LAB:  OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. \_\_\_\_\_

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: \_\_\_\_\_ City: \_\_\_\_\_ Zip: 91326

Contact Person: \_\_\_\_\_ Title: \_\_\_\_\_ Tel: \_\_\_\_\_

Analysis Requested by: Sumner Wilson Date: 12/1/16

Approved by: Jason Low Office: \_\_\_\_\_ Budget #: 44716

REASON REQUESTED: Court/Hearing Board  Permit Pending  Hazardous/Toxic Spill

Suspected Violation Rule(s) \_\_\_\_\_ Other

Sample Collected by: Kayla Jordan Date: 12/1/16 Time: 11:15am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Reseda Station	54680	11/26/16 / 00:00 / 24 hours	<-30"	0"

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Kayla Jordan</i>	<i>V. Vergara</i>	SCAQMD Lab	12/1/16	13:54

Remarks: 1:3 scheduled samples from station at Reseda.  
 Reseda Station – 18328 Gault St, Los Angeles, CA 91335  
 GPS (34.199225, -118.532743). Right sampler sn 6113