

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO Jason Low, Ph.D., ADEO
Monitoring & Analysis

LABORATORY NO 1633125

DATE RECEIVED 12/01/2016

SAMPLE DESCRIBED AS
24 hour sample
Canister 54645

FACILITY ID NO NA

REQUESTED BY Sumner Wilson

SAMPLING LOCATION
Highlands Community Center

ST NO / PROJECT NA

REPORT GENERATED 1/20/2017

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54645</u>
Pressure (Torr)	1269
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	420
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 1-20-17

Reviewed By: John McKeyle For JFN
Joan Niertit, Principal A.Q. Chemist
Laboratory Services

Date Approved: 1/26/17

Approved By: [Signature]
Aaron Katzenstein, Ph.D., Sr. Manager
Laboratory Services
909-396-2219

1-20-17

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482**

**MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY**

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1633125</u>
24 hour sample Canister 54645	REQUESTED BY	<u>Sumner Wilson</u>
	DATE ANALYZED	<u>1/3/2017</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.10	9.17	PASS
CH ₄ , ppmvC	0.3	2.02	2.03	0.39	PASS
CO ₂ , ppmvC	0.4	1.57	1.98	26.06	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.99	-2.04	PASS
NM/NEOC, ppmvC	0.5	2.03	2.17	6.76	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1052	1.58	PASS
CH ₄ , ppmvC	0.3	1068	1049	-1.73	PASS
CO ₂ , ppmvC	0.4	1022	1034	1.19	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1035	-0.87	PASS
NM/NEOC, ppmvC	0.5	1024	1010	-1.36	PASS

REFERENCE NO: 16QM2AD
QM2-101-133

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT INFORMATION
 IN
 LAB
WO #: 1633125



TO: SCAQMD LAB: OTHER: _____
 SOURCE NAME: Southern California Gas Co. I.D. No. _____
 Source Address: 12801 Tampa Ave City: Porter Ranch
 Mailing Address: _____ City: _____ Zip: 91326
 Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 12/1/16
 Approved by: Jason Low Office: _____ Budget #: 44716
 REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
 Suspected Violation Rule(s) _____ Other

Sample Collected by: Kayla Jordan Date: 12/1/16 Time: 10:15am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Highlands Community	54645	11/26/16 / 00:00 / 24 hours	<-30"	11"

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Kayla Jordan</i>	<i>V. Vergara</i>	SCAQMD Lab	12/1/16	13:54

Remarks 1:3 scheduled samples from trailer.
 Highlands community pool parking lot. Address: 12378 High Glen Way, Northridge CA 91326 (across from 12377)
 Right Sampler, sn 5323