Field Evaluation of Perkin Elmer - ELM



Air Quality Sensor Performance Evaluation Center

Background

- From 07/22/2015 to 09/25/2015, three Perkin Elmer ELM monitors were deployed in Rubidoux and run side-by-side SCAQMD's Federal Reference Method (FRM) and Federal Equivalent Method (FEM) instruments measuring the same pollutants
- <u>ELM (3 units tested)</u>:
 Metal-Oxide gas sensors (non-FRM)
 Light-Scattering PM sensors (non-FEM)
 Each unit measures: NO, NO₂, O₃, PM10, Temp and RH
 Unit cost: ~\$5,200
 Time resolution: 1-min

≻Units IDs: 1088, 1177, 1197



- <u>SCAQMD FRM/FEM instruments</u>:
 - ➢NO_X instrument; cost: ~\$11,000
 - Time resolution: 1-min
 - $>O_3$ instrument; cost: ~\$13,000
 - Time resolution; 1-min
 - Meteorological station (wind speed, wind direction temperature, relative humidity, and pressure); cost: ~\$5,000
 - Time resolution: 1-min
 - MetOne BAM (reference method); Cost: ~\$20,000
 - Beta-attenuation monitor (FEM); Measures PM_{2.5}
 - Time resolution: 1-hr
 - ➢GRIMM (reference method); Cost: ~\$25,000 and up
 - Optical particle counter (FEM); Uses proprietary algorithms to calculate total PM, PM_{2.5}, and PM₁ from particle number measurements
 - Time resolution: 1-min

Data validation & recovery

- Basic QA/QC procedures were used to validate the collected data (i.e., obvious outliers, negative values, and invalid data-points were eliminated from the data-set)
- Data recovery for all three ELM sensors was ~100%

ELM; intra-model variability

 With the exception of PM₁₀, modest-to-low intra-model variability was observed for all measured pollutants and meteorological variables



Data validation & recovery

- Basic QA/QC procedures were used to validate the collected FEM data (i.e. obvious outliers, negative values and invalid data-points were eliminated from data-set)
- PM_{10} data recovery for the GRIMM and BAM instruments was ~100%

Equivalent Methods; BAM vs GRIMM

• Very good correlation between the two equivalent methods R² = 0.81



ELM vs FEM GRIMM (PM₁₀; 5-min mean)



- ELM PM₁₀ measurements do not seem to track well the PM10 diurnal variations recorded by the GRIMM (FEM) instrument
- All ELM units show very poor correlation with the corresponding FEM data (R²<0.15)



200

ELM vs FEM GRIMM (PM₁₀; 1-hr mean)



- ELM PM₁₀ measurements do not seem to track well the PM10 diurnal variations recorded by the GRIMM (FEM) instrument
- All ELM units show very poor correlation with the corresponding FEM data (R²<0.17)





ELM vs FEM BAM (PM₁₀; 1-hr mean)



- ELM PM₁₀ measurements do not seem to track well the PM10 diurnal variations recorded by the BAM (FEM) instrument
- All ELM units show very poor correlation with the corresponding FEM data (R²<0.18)







7

ELM vs FRM (NO₂; 5-min mean)



- ELM NO₂ measurements do not seem to track the NO₂ diurnal variations recorded by the FRM instrument
- Very poor correlation with FRM measurement data (R²~0.0)
- Potential interference w/ ambient ozone and/or RH (to be investigated during chamber experiments)







ELM vs FRM (O₃; 5-min mean)



- ELM Ozone measurements correlate very well with the corresponding FRM measurements (0.89
 R²<0.96)
- For units 1177 and 1199 the baseline is substantially higher than 0







Temperature (5-min mean)



- ELM Temp measurements show excellent correlation with the corresponding Station temp data (0.94 < R² < 0.95)
- ELM temp data are slightly overestimated



Relative Humidity (5-min mean)



 ELM Relative Humidity measurements show excellent correlation with the corresponding Station RH data (0.90 < R² < 0.97)







Discussion

- Overall, the three ELM monitors were reliable (i.e. no down time over a period of about two months) and they showed modest to low intra-model variability for all measured pollutants (except PM₁₀) and meteorological variables
- The ELM ozone sensors showed excellent correlation with a substantially more expensive FRM ozone instrument (0.89 < R² < 0.96)
- The ELM PM₁₀ and NO₂ sensors correlated very poorly with the corresponding FEM instruments data (0.0 < R² < 0.15)
- NO₂ sensor measurements might have been affected by a potential interference with ozone and/or relative humidity. This will be thoroughly examined during laboratory testing
- Temperature and relative humidity correlated very well (0.90 < R² < 0.97) with the corresponding weather station data
- No sensor calibration had been performed prior to the beginning of this field testing
- Laboratory chamber testing is necessary to fully evaluate the performance of these sensors under controlled temperature/relative humidity conditions and known gaseous concentrations
- All results are still preliminary