# Laboratory Evaluation: Aeroqual S500-GSS



Air Quality Sensor Performance Evaluation Center

## Background

Three **Aeroqual S500-GSS** sensors (units IDs: 1, 2, 3) were evaluated in the South Coast AQMD Laboratory under controlled Volatile Organic Compound (VOC) concentrations and climate conditions. The sensor measurements were compared with a reference instrument (Thermo Fisher Scientific, Model 55i; hereinafter **Thermo 55i**).

#### Aeroqual S500-GSS (3 units tested):

- VOC Sensor Gas Sensitive Semiconductors (GSS, Aeroqual, non-FEM)
  - > TVOC output range: 0-25 ppm
  - Accuracy of Factory Calibration: < ±0.1 ppm + 10%
  - Minimum Detection Limit: 0.1ppm
  - Measurement interval: 1-min
- Each unit measures: VOC (ppb)
- ➤ Unit cost: ~\$1,650
- ➤ Units IDs: 1, 2, 3



Aeroqual S500-GSS

thermoscertic
 thermos

#### Thermo 55i

#### <u>Reference Instrument\*:</u>


- Thermo Fisher 55i
  Measures: methane (CH₄) and non-methane hydrocarbon (NMHC)
  > Unit cost: ~\$27,000
  > Specifications:
  > Measurement ranges: 0-50 ppm
  > Limit of Detection (LOD): 50 ppb
  > Analysis time: ~70 seconds
  > Accuracy: ±1% of range
  > Repeatability: ±2% of measured value or 50 ppb (whichever is larger)
  > Drift: ±2% of span over 24 hours
  - Ambient operating temperature: 15-35 °C
  - Sample temperature: ambient to 35 °C

\*An Agilent GC-FID is typically also used for reference measurements in the AQ-SPEC VOC sensor laboratory testing protocol. However, it was not needed for this evaluation as the testing was prematurely completed.

# Phase 1: Transient Plume Detection

| Testing Phase<br>#1       | Method   | Parameters Evaluated  |
|---------------------------|--|---|
| Transient Plume Detection | 5 VOC plume events at various concentrations in randomized order | <ul><li> Response time</li><li> % of peak detection</li></ul> |

### Aeroqual S500-GSS vs Thermo 55i



- Sensor response to a transient event was determined to be successful if the sensor peak amplitude was at least 3 times the standard deviation of 20 consecutive pre-peak sensor values.
- The Aeroqual S500-GSS sensors responded to three of the five transient events (circled in the plot), which had VOC concentrations of at least 6 ppm as recorded by the Thermo 55i; there was no apparent response to VOC concentration events lower than 6 ppm by the Aeroqual S500-GSS sensors.

### **VOC Concentrations for Subsequent Test Phases**

| Testing Phase                     | VOC Concentration   |
|-----------------------------------|---|
| 1. Transient Event Detection      | 5 events at various concentrations up to 10 ppm in randomized order                                     |
| 2. Initial Concentration Ramping  | Concentration ramping from 15 ppb to 8 ppm  |
| 3. Effect of Temperature and RH   | Concentration held at 4 ppm while varying T and RH conditions   |
| 4. Effect of gaseous interferents | Concentrations held at either 4 ppm (with CO or CO2 as interferent) or 200 ppb (with O3 as interferent) |
| 5. Outdoor Simulation             | Concentrations from 200 to 400 ppb  |
| 6. Final Concentration Ramping    | Concentration ramping from 15 ppb to 8 ppm  |

The VOC concentrations used in most subsequent testing phases are less than 6 ppm. Based on the results from the Phase 1 Transient Event Detection test, the sensors did not proceed further to Phase 2 testing, which exposes sensors to VOC blend concentrations of 0.06, 0.2, 0.4, 1.6, 2, 4, 6, and 8 ppm.

## Discussion

- > Data Recovery: The Aeroqual S500-GSS sensors showed 100% data recovery for all experiments
- Phase 1: Transient Event Detection
  - The sensors responded to three of the five transient events. The sensors did not appear to show responses to VOC concentrations lower than 6 ppm as recorded by Thermo 55i.
  - The VOC concentrations required in the subsequent testing phases are generally lower than 6 ppm; based on the results from the Transient Event Detection test, the sensors will not proceed to further testing.