

# AQ-SPEC

## Air Quality Sensor Performance Evaluation Center

### Evaluation Summary

#### Sensor Description

Manufacturer/Model:  
Elitech/  
Temtop M2000 2<sup>nd</sup>  
Generation

Pollutants:  
PM<sub>2.5</sub> and PM<sub>10</sub> mass  
concentration

Time Resolution:  
1 min.

Type: Optical



#### Additional Information

##### Field evaluation report:

<http://www.aqmd.gov/aq-spec/evaluations/field>

##### Lab evaluation report:

<http://www.aqmd.gov/aq-spec/evaluations/laboratory>

##### AQ-SPEC website:

<http://www.aqmd.gov/aq-spec>

- Overall, the accuracy of the Temtop M2000 2<sup>nd</sup> Generation sensors was fairly constant (55% to 70%) over the PM<sub>2.5</sub> mass concentration range tested. The Temtop M2000 sensors overestimated PM<sub>2.5</sub> measurements from FEM GRIMM in the laboratory experiments at 20 °C and 40% RH.
- The Temtop M2000 sensors showed high precision for all test combinations (PM concentrations, T and RH) for PM<sub>2.5</sub> mass concentrations
- The Temtop M2000 sensors (IDs: 1, 2, and 3) showed low intra-model variability in both the field and laboratory evaluations.
- Data recovery was 100% from all units in both the field and laboratory evaluations.
- For PM<sub>2.5</sub>, Temtop M2000 sensors showed strong correlations with the FEM GRIMM and FEM T640 from the field ( $0.76 < R^2 < 0.83$ ) and showed very weak correlations with the GRIMM and T640 for PM<sub>10</sub> ( $0.17 < R^2 < 0.28$ ). The Temtop M2000 sensors showed very strong correlations with the FEM GRIMM in the laboratory evaluations ( $R^2 > 0.99$  for PM<sub>2.5</sub>).
- The same three Temtop M2000 units were tested both in the field (1<sup>st</sup> stage of testing) and in the laboratory (2<sup>nd</sup> stage of testing)



### Field Evaluation Highlights

- Deployment period 03/27/2020 to 06/04/2020 : the three Temtop M2000 sensors showed strong correlations with the corresponding FEM GRIMM and FEM T640 PM<sub>2.5</sub> mass concentrations and showed very weak correlations with the corresponding GRIMM and T640 PM<sub>10</sub> mass concentrations
- The units exhibited low intra-model variability and data recovery for PM<sub>2.5</sub> and PM<sub>10</sub> was ~100% from all units.

5-min mean, all ref. instr.

PM<sub>2.5</sub>:  $0.76 < R^2 < 0.83$

PM<sub>10</sub>:  $0.17 < R^2 < 0.28$



Coefficient of Determination ( $R^2$ ) quantifies how the three sensors followed the PM<sub>2.5</sub> concentration change by the reference instruments.

An  $R^2$  approaching the value of 1 reflects a near perfect agreement, whereas a value of 0 indicates a complete lack of correlation.

# Laboratory Evaluation Highlights

## Accuracy (PM<sub>2.5</sub>)

$$A (\%) = 100 - \frac{|\bar{X} - \bar{R}|}{\bar{R}} * 100$$


Steady state #	Sensor Mean (µg/m <sup>3</sup> )	FEM GRIMM (µg/m <sup>3</sup> )	Accuracy (%)
1	10.9	8.4	70.3
2	19.7	13.7	56.6
3	63.7	45.3	59.3
4	161.4	117.7	62.9
5	379.2	261.5	55.0

Accuracy was evaluated by a concentration ramping experiment at 20 °C and 40%. The sensor's readings at each ramping steady state are compared to the reference instrument.

A negative % means sensors' overestimation by more than two fold. The higher the positive value (close to 100%), the higher the sensor's accuracy.




## Precision (PM<sub>2.5</sub>)



100% represents high precision.

Sensor's ability to generate precise measurements of PM<sub>2.5</sub> concentration at low, medium, and high pollutant levels were evaluated under 9 combinations of T and RH, including extreme weather conditions like cold and dry (5 °C and 15%) cold and humid (5 °C and 65%), hot and humid (35 °C and 65%), or hot and dry (35 °C and 15%).

## Coefficient of Determination



The Temtop M2000 sensors showed very strong correlations with the corresponding FEM GRIMM PM<sub>2.5</sub> data ( $R^2 > 0.99$ ) at 20 °C and 40% RH.

## Climate Susceptibility

From the laboratory studies, temperature and relative humidity had minimal effect on the Temtop M2000 sensors' precision.

## Observed Interferents

N/A



All documents, reports, data, and other information provided in this document are for informational use only. Mention of trade names or commercial products does not constitute endorsement or recommendation. As a Government Agency, the South Coast AQMD and its AQ-SPEC program highly recommend interested entities to make use and purchase decisions based on the requirements of their study design, the technical aspects and features of their specific project applications.