



South Coast AQMD Site Survey Report for Compton

Last updated: May 13, 2023



AQS ID	ARB Number	Site Start Date	Reporting Agency and Agency Code
060371302	70112	01/2004	South Coast AQMD (0972)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
700 N. Bullis Road Compton, CA 90221	Los Angeles	South Coast	33.901389	-118.205000	22



Detailed Site Information

Local site name	Compton			
AQS ID	060371302			
GPS coordinates (decimal degrees)	Latitude: 33.901389, Longitude: -118.205000			
Street Address	700 N. Bullis Road, Compton, CA 90221			
County	Los Angeles			
Distance to roadways (meters)	14; 1680			
Traffic count (AADT, year)	1,000 / 2012; 710/105, 225,000, 2011			
Groundcover (e.g. asphalt, dirt, sand)	Asphalt			
Representative statistical area name (i.e. MSA, CBSA, other)	31080-Los Angeles-Long Beach-Anaheim, MSA			
Pollutant, POC	Carbon Monoxide, 1	Nitrogen Dioxide, 1	Ozone , 1	Lead, 1
Primary / QA Collocated / Other	N/A	N/A	N/A	Primary
Parameter code	42101	42602	44201	14129
Basic monitoring objective(s)	NAAQS	NAAQS	NAAQS	NAAQS
Site type(s)	Highest Concentration	Population Exposure	Population Exposure	non-source-oriented
Monitor (type)	SLAMS	SLAMS	SLAMS	SLAMS
Network Affiliation	N/A	N/A	N/A	N/A
Instrument manufacturer and model	Horiba APMA 370	Teledyne T200	Teledyne T400	TSP, A Sampler, Tisch HiVol +
Method code	158	099	087	110
FRM/FEM/ARM/ other	FRM	FRM	FEM	FRM
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	N/A	N/A	South Coast AQMD
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Spatial scale (e.g. micro, neighborhood)	Middle	Middle	Neighborhood	Neighborhood
Monitoring start date (MM/DD/YYYY)	01/2004	01/2004	01/2004	01/2004
Current sampling frequency (e.g.1:3, continuous)	1:1	1:1	1:1	1:6
Calculated sampling frequency (e.g. 1:3/1:1)	N/A	N/A	N/A	1:6
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	01/01-12/31
Probe height (meters)	4.1	4.1	4.1	2.2
Distance from supporting structure (meters)	1.5	1.5	1.5	1.2
Distance from obstructions on roof (meters)	N/A	N/A	N/A	N/A

Distance from obstructions not on roof (meters)	N/A	N/A	N/A	N/A
Distance from trees (meters)	17 m NW Height 8.8 m	17 m NW Height 8.8 m	17 m NW Height 8.8 m	18 m NW Height 8.8 m
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	N/A
Distance between collocated monitors (meters)	N/A	N/A	N/A	2.0
Unrestricted airflow (degrees)	360°	360°	360°	360°
Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	Teflon	Teflon	Teflon	N/A
Residence time for reactive gases (seconds)	6.5	11.6	11.8	N/A
Will there be changes within the next 18 months? (Y/N)	No	No	No	No
Is it suitable for comparison against the annual PM _{2.5} ? (Y/N)	N/A	N/A	N/A	N/A
Frequency of flow rate verification for manual PM samplers	N/A	N/A	N/A	Monthly
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	N/A	N/A
Frequency of one-point QC check for gaseous instruments	Nightly	Nightly	Nightly	N/A
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	05/19/2022	05/19/2022	05/19/2022	N/A
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	N/A	N/A	03/25/2022 10/28/2022

Pollutant, POC	24 Hour PM2.5, 1	Lead, 2	Continuous PM2.5, 3	
Primary / QA Collocated / Other	Primary	QA Collocated	Other	
Parameter code	88101	14129	88502	
Basic monitoring objective(s)	NAAQS	NAAQS	NAAQS	
Site type(s)	Population Exposure	non-source-oriented	Population Exposure	
Monitor (type)	SLAMS	SLAMS	SLAMS	
Network Affiliation	N/A	N/A	N/A	
Instrument manufacturer and model	Partisol 2025i	TSP, B Sampler, Tisch HiVol +	Met One BAM 1020	
Method code	145	110	733	
FRM/FEM/ARM/ other	FRM	FRM	FEM	
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Analytical Lab (i.e., weigh lab, toxics lab, other)	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Spatial scale (e.g. micro, neighborhood)	Neighborhood	Neighborhood	Neighborhood	
Monitoring start date (MM/DD/YYYY)	01/2004	05/2015	07/01/2020	
Current sampling frequency (e.g. 1:3, continuous)	Daily	1:6	1:1	
Calculated sampling frequency (e.g. 1:3/1:1)	Daily	1:6	N/A	
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	
Probe height (meters)	3.1	2.2	4.7	
Distance from supporting structure (meters)	2.1	1.2	2.1	
Distance from obstructions on roof (meters)	NA	N/A	N/A	
Distance from obstructions not on roof (meters)	N/A	N/A	N/A	
Distance from trees (meters)	20 m NW Height 8.8 m	17 m NW Height 8.8 m	14 m NW Height 8.8 m	
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	
Distance between collocated monitors (meters)	N/A	2.0	N/A	
Unrestricted airflow (degrees)	360°	360°	360°	

Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A	N/A	N/A	
Residence time for reactive gases (seconds)	N/A	N/A	N/A	
Will there be changes within the next 18 months? (Y/N)	No	No	No	
Is it suitable for comparison against the annual PM2.5? (Y/N)	Yes	N/A	Yes	
Frequency of flow rate verification for manual PM samplers	Monthly	Monthly	N/A	
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	Monthly	
Frequency of one-point QC check for gaseous instruments	N/A	N/A	N/A	
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A	N/A	N/A	
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	03/25/2022 09/06/2022	03/25/2022 09/06/2022	03/25/2022 09/06/2022	

Pollutant, POC	WS & D, 1/1	RH/T, 1/1	BP, 1	
Primary / QA Collocated / Other	N/A	N/A	N/A	
Parameter code	61101/61102	62201/62101	64101	
Basic monitoring objective(s)	Research	Research	Research	
Site type(s)	Meteorological	Meteorological	Meteorological	
Monitor (type)	SLAMS	SLAMS	SLAMS	
Network Affiliation	N/A	N/A	N/A	
Instrument manufacturer and model	RM Young 05305V	Rotronic HC2-S3	Met One 091	
Method code	065/065	063/063	015	
FRM/FEM/ARM/ other	N/A	N/A	N/A	
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	N/A	N/A	
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Spatial scale (e.g. micro, neighborhood)	Neighborhood	Neighborhood	Neighborhood	
Monitoring start date (MM/DD/YYYY)	01/2004	01/2004	01/2004	
Current sampling frequency (e.g.1:3, continuous)	Continuous	Continuous	Continuous	
Calculated sampling frequency (e.g. 1:3/1:1)	1:1	1:1	1:1	
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	
Probe height (meters)	10	5.5	3.5	
Distance from supporting structure (meters)	10	3.0	1.0	
Distance from obstructions on roof (meters)	N/A	N/A	N/A	
Distance from obstructions not on roof (meters)	N/A	N/A	N/A	
Distance from trees (meters)	16 m NW Height 8.8 m	16 m NW Height 8.8 m	16 m NW Height 8.8 m	
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	
Distance between collocated monitors (meters)	N/A	N/A	N/A	
Unrestricted airflow (degrees)	360°	360°	360°	

Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A	N/A	N/A	
Residence time for reactive gases (seconds)	N/A	N/A	N/A	
Will there be changes within the next 18 months? (Y/N)	No	No	No	
Is it suitable for comparison against the annual PM2.5? (Y/N)	N/A	N/A	N/A	
Frequency of flow rate verification for manual PM samplers	N/A	N/A	N/A	
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	N/A	
Frequency of one-point QC check for gaseous instruments	N/A	N/A	N/A	
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A	N/A	N/A	
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	N/A	N/A	

**Compton
Site Photos**



Looking North from the probe.



Looking East from the probe.



Looking South from the probe.



Looking West from the probe.

**Compton
Site Photos (Cont.)**



Looking at the probe from the North.



Looking at the probe from the East.



Looking at the probe from the South.



Looking at the probe from the West.