


South Coast AQMD Site Survey Report for Fontana

Last updated: May 12, 2023



AQS ID	ARB Number	Site Start Date	Reporting Agency and Agency Code
060712002	36197	08/1981	South Coast AQMD (0972)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
14360 Arrow Highway Fontana, CA 92335	San Bernardino	South Coast	34.100020	-117.491982	363



Detailed Site Information

Local site name	Fontana			
AQS ID	060712002			
GPS coordinates (decimal degrees)	Latitude: 34.100020, Longitude: -117.491982			
Street Address	14360 Arrow Highway, Fontana, CA 92335			
County	San Bernardino			
Distance to roadways (meters)	94			
Traffic count (AADT, year)	12,500 / 2012			
Groundcover (e.g. asphalt, dirt, sand)	Gravel			
Representative statistical area name (i.e. MSA, CBSA, other)	40140-Riverside-San Bernardino-Ontario, CA MSA			
Pollutant, POC	Carbon Monoxide, 1	Nitrogen Dioxide, 1	Ozone, 1	Sulfur Dioxide, 1
Primary / QA Collocated / Other	N/A	N/A	N/A	N/A
Primary / QA Collocated / Other	N/A	N/A	N/A	N/A
Parameter code	42101	42602	44201	42401
Basic monitoring objective(s)	NAAQS	NAAQS	NAAQS	NAAQS
Site type(s)	Population Exposure	Population Exposure	Population Exposure	Population Exposure
Monitor (type)	SLAMS	SLAMS	SLAMS	SLAMS
Network Affiliation	N/A	N/A	N/A	N/A
Instrument manufacturer and model	Horiba APMA 360	Teledyne T200	Teledyne T400	Thermo 43i
Method code	106	099	087	560
FRM/FEM/ARM/ other	FRM	FRM	FEM	FEM
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	N/A	N/A	N/A
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Spatial scale (e.g. micro, neighborhood)	Neighborhood	Urban	Urban	Neighborhood
Monitoring start date (MM/DD/YYYY)	08/1981	08/1981	08/1981	08/1981
Current sampling frequency (e.g. 1:3, continuous)	1:1	1:1	1:1	1:1
Calculated sampling frequency (e.g. 1:3/1:1)	N/A	N/A	N/A	N/A
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	01/01-12/31
Probe height (meters)	4.1	4.1	4.1	4.1
Distance from supporting structure (meters)	1.5	1.5	1.5	1.5

Distance from obstructions on roof (meters)	N/A	N/A	N/A	N/A
Distance from obstructions not on roof (meters)	N/A	N/A	N/A	N/A
Distance from trees (meters)	N/A	N/A	N/A	N/A
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	N/A
Distance between collocated monitors (meters)	N/A	N/A	N/A	N/A
Unrestricted airflow (degrees)	360°	360°	360°	360°
Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	Teflon	Teflon	Teflon	Teflon
Residence time for reactive gases (seconds)	5.3	6.4	5.9	13.7
Will there be changes within the next 18 months? (Y/N)	No	No	No	No
Is it suitable for comparison against the annual PM _{2.5} ? (Y/N)	N/A	N/A	N/A	N/A
Frequency of flow rate verification for manual PM samplers	N/A	N/A	N/A	N/A
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	N/A	N/A
Frequency of one-point QC check for gaseous instruments	Nightly	Nightly	Nightly	Nightly
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	02/10/2022	02/10/2022	02/10/2022	12/01/2022
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	N/A	N/A	N/A

Pollutant, POC	PM10, 2	Speciated PM2.5, 11	24 Hour PM2.5, 1	Continuous PM2.5, 3
Primary / QA Collocated / Other	Primary	Other	Primary	Other
Parameter code	81102	88502	88101	88502
Basic monitoring objective(s)	NAAQS	NAAQS	NAAQS	General Public Info
Site type(s)	Highest Concentration	Population Exposure	Population Exposure	Population Exposure
Monitor (type)	SLAMS	SLAMS	SLAMS	Other
Network Affiliation	N/A	N/A	N/A	N/A
Instrument manufacturer and model	Tisch	Met One SASS	Partisol 2025i	Met One BAM 1020
Method code	141	810	145	731
FRM/FEM/ARM/ other	FRM	Other	FRM	Non-FEM
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Analytical Lab (i.e., weigh lab, toxics lab, other)	South Coast AQMD	South Coast AQMD	South Coast AQMD	N/A
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Spatial scale (e.g. micro, neighborhood)	Neighborhood	Neighborhood	Neighborhood	Neighborhood
Monitoring start date (MM/DD/YYYY)	08/1981	02/20/2004	01/1985	05/10/2023
Current sampling frequency (e.g. 1:3, continuous)	1:6	1:6	1:3	1:1
Calculated sampling frequency (e.g. 1:3/1:1)	1:6	No CFR mandated sampling schedule.	1:3	N/A
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	01/01-12/31
Probe height (meters)	2.5	2.8	3.1	4.7
Distance from supporting structure (meters)	1.5	1.8	2.1	2.1 *Roof itself is supporting structure.
Distance from obstructions on roof (meters)	N/A	N/A	N/A	N/A
Distance from obstructions not on roof (meters)	N/A	N/A	N/A	N/A
Distance from trees (meters)	N/A	N/A	N/A	N/A
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	N/A
Distance between collocated monitors (meters)	N/A	N/A	N/A	N/A
Unrestricted airflow (degrees)	360°	360°	360°	360°

Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A	N/A	N/A	N/A
Residence time for reactive gases (seconds)	N/A	N/A	N/A	N/A
Will there be changes within the next 18 months? (Y/N)	No	No	No	No
Is it suitable for comparison against the annual PM2.5? (Y/N)	N/A	N/A	Yes	No
Frequency of flow rate verification for manual PM samplers	Monthly	Monthly	Monthly	N/A
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	N/A	Monthly
Frequency of one-point QC check for gaseous instruments	N/A	N/A	N/A	N/A
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A	N/A	N/A	N/A
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	04/28/2022 10/21/2022	05/12//2022 10/21/2022	04/28/2022 10/21/2022	None – New Installation moved from Upland (060711004)

Pollutant, POC	WS & D, 1/1	RH/T, 1/1	BP, 1	
Primary / QA Collocated / Other	N/A	N/A	N/A	
Parameter code	61101/61102	62201/62101	64101	
Basic monitoring objective(s)	Research	Research	Research	
Site type(s)	Meteorological	Meteorological	Meteorological	
Monitor (type)	SLAMS	SLAMS	SLAMS	
Network Affiliation	N/A	N/A	N/A	
Instrument manufacturer and model	RM Young 05305V	Elektronik EE181	Met One 091	
Method code	065/065	061/061	015	
FRM/FEM/ARM/ other	N/A	N/A	N/A	
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	N/A	N/A	
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Spatial scale (e.g. micro, neighborhood)	Neighborhood/Urban	Neighborhood/Urban	Neighborhood/Urban	
Monitoring start date (MM/DD/YYYY)	08/1981	08/1981	08/1981	
Current sampling frequency (e.g.1:3, continuous)	Continuous	Continuous	Continuous	
Calculated sampling frequency (e.g. 1:3/1:1)	1:1	1:1	1:1	
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	
Probe height (meters)	10	9.0	2	
Distance from supporting structure (meters)	10	9.0	2	
Distance from obstructions on roof (meters)	N/A	N/A	N/A	
Distance from obstructions not on roof (meters)	N/A	N/A	N/A	
Distance from trees (meters)	6	6	6	
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	
Distance between collocated monitors (meters)	N/A	N/A	N/A	
Unrestricted airflow (degrees)	360°	360°	360°	

Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A	N/A	N/A	
Residence time for reactive gases (seconds)	N/A	N/A	N/A	
Will there be changes within the next 18 months? (Y/N)	No	No	No	
Is it suitable for comparison against the annual PM _{2.5} ? (Y/N)	N/A	N/A	N/A	
Frequency of flow rate verification for manual PM samplers	N/A	N/A	N/A	
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	N/A	
Frequency of one-point QC check for gaseous instruments	N/A	N/A	N/A	
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A	N/A	N/A	
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	N/A	N/A	

**Fontana
Site Photos**



Looking North from the probe.



Looking East from the probe.



Looking South from the probe.



Looking West from the probe.

**Fontana
Site Photos (Cont.)**




Looking at the probe from the North.



Looking at the probe from the East.



Looking at the probe from the South.



Looking at the probe from the West.