



South Coast AQMD Site Survey Report for Pico Rivera #2

Last updated: May 16, 2023



AQS ID	ARB Number	Site Start Date	Reporting Agency and Agency Code			
060371602	70185	09/2005	South Coast AQMD (0972)			
Site Address		County	Air Basin	Latitude	Longitude	Elevation
4144 San Gabriel River Pkwy. Pico Rivera, CA 90660		Los Angeles	South Coast	34.010283	-118.068500	58m



Detailed Site Information

Local site name	Pico Rivera #2			
AQS ID	060371602			
GPS coordinates (decimal degrees)	Latitude: 34.010283, Longitude: -118.068500			
Street Address	4144 San Gabriel River Pkwy., Pico Rivera, CA			
County	Los Angeles			
Distance to roadways (meters)	43			
Traffic count (AADT, year)	20,000 / 2012; 605/Beverly, 255,000 2011			
Groundcover (e.g. asphalt, dirt, sand)	Asphalt			
Representative statistical area name (i.e. MSA, CBSA, other)	31080-Los Angeles, Long Beach-Anaheim MSA			
Pollutant, POC	Carbon Monoxide, 1	Nitrogen Dioxide, 1	Ozone, 1	Lead, 1
Primary / QA Collocated / Other	N/A	N/A	N/A	Primary
Parameter code	42101	42602	44201	14129
Basic monitoring objective(s)	NAAQS	NAAQS	NAAQS	NAAQS
Site type(s)	Population Exposure	Population Exposure	Population Exposure	non-source-oriented
Monitor (type)	SLAMS	SLAMS	SLAMS	SLAMS
Network affiliation	N/A	N/A	N/A	N/A
Instrument manufacturer and model	Horiba APMA 370	Teledyne T200	Thermo 49i	GMW TSP 1200
Method code	158	099	047	110
FRM/FEM/ARM/ other	FRM	FRM	FEM	FRM
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	N/A	N/A	South Coast AQMD
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Spatial scale (e.g. micro, neighborhood)	Neighborhood	Neighborhood	Neighborhood	Neighborhood
Monitoring start date (MM/DD/YYYY)	9/2005	9/2005	09/2005	09/2005
Current sampling frequency (e.g. 1:3, continuous)	1:1	1:1	1:1	1:6
Calculated sampling frequency (e.g. 1:3/1:1)	N/A	N/A	N/A	N/A
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	01/01-12/31
Probe height (meters)	4.4	4.4	4.4	2.2
Distance from supporting structure (meters)	1.8 Roof itself is supporting structure.	1.8 Roof itself is supporting structure.	1.8 Roof itself is supporting structure.	1.2
Distance from obstructions on roof (meters)	N/A	N/A	N/A	N/A

Distance from obstructions not on roof (meters)	N/A	N/A	N/A	N/A
Distance from trees (meters)	N/A	N/A	N/A	N/A
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	N/A
Distance between collocated monitors (meters)	N/A	N/A	N/A	N/A
Unrestricted airflow (degrees)	360°	360°	360°	360°
Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	Teflon	Teflon	Teflon	N/A
Residence time for reactive gases (seconds)	7.8	13.7	10.4	N/A
Will there be changes within the next 18 months? (Y/N)	No	No	No	No
Is it suitable for comparison against the annual PM _{2.5} ? (Y/N)	N/A	N/A	N/A	N/A
Frequency of flow rate verification for manual PM samplers	N/A	N/A	N/A	Monthly
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	N/A	N/A
Frequency of one-point QC check for gaseous instruments	Nightly	Nightly	Nightly	N/A
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	04/14/2022	04/14/2022	04/14/2022	N/A
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	N/A	N/A	03/02/2022 08/04/2022

Pollutant, POC	24 Hour PM2.5, 1			
Primary / QA Collocated / Other	Primary			
Parameter code	88101			
Basic monitoring objective(s)	NAAQS			
Site type(s)	Population Exposure,			
Monitor (type)	SLAMS			
Network affiliation	N/A			
Instrument manufacturer and model	Partisol 2025i			
Method code	145			
FRM/FEM/ARM/ other	FRM			
Collecting Agency	South Coast AQMD			
Analytical Lab (i.e., weigh lab, toxics lab, other)	South Coast AQMD			
Reporting Agency	South Coast AQMD			
Spatial scale (e.g. micro, neighborhood)	Neighborhood			
Monitoring start date (MM/DD/YYYY)	09/2005			
Current sampling frequency (e.g. 1:3, continuous)	1:6			
Calculated sampling frequency (e.g. 1:3/1:1)	1:3			
Sampling season (MM/DD-MM/DD)	01/01-12/31			
Probe height (meters)	3.1			
Distance from supporting structure (meters)	2.1			
Distance from obstructions on roof (meters)	N/A			
Distance from obstructions not on roof (meters)	N/A			
Distance from trees (meters)	N/A			
Distance to furnace or incinerator flue (meters)	N/A			
Distance between collocated monitors (meters)	N/A			
Unrestricted airflow (degrees)	360°			

Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A			
Residence time for reactive gases (seconds)	N/A			
Will there be changes within the next 18 months? (Y/N)	No			
Is it suitable for comparison against the annual PM2.5? (Y/N)	Yes			
Frequency of flow rate verification for manual PM samplers	Monthly			
Frequency of flow rate verification for automated PM analyzers	N/A			
Frequency of one-point QC check for gaseous instruments	N/A			
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A			
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	03/02/2022 08/04/2022			

Pollutant, POC	WS & D, 1/1	RH/T, 1/1	BP, 1	
Primary / QA Collocated / Other	N/A	N/A	N/A	
Parameter code	61101/61102	62201/62101	64101	
Basic monitoring objective(s)	Research	Research	Research	
Site type(s)	Meteorological	Meteorological	Meteorological	
Monitor (type)	SLAMS	SLAMS	SLAMS	
Network Affiliation	N/A	N/A	N/A	
Instrument manufacturer and model	RM Young 05305V	Rotronic HC2-S3	Met One 091	
Method code	065/065	063/063	015	
FRM/FEM/ARM/ other	N/A	N/A	N/A	
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	N/A	N/A	
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Spatial scale (e.g. micro, neighborhood)	Neighborhood	Neighborhood	Neighborhood	
Monitoring start date (MM/DD/YYYY)	09/2005	09/2005	09/2005	
Current sampling frequency (e.g.1:3, continuous)	Continuous	Continuous	Continuous	
Calculated sampling frequency (e.g. 1:3/1:1)	1:1	1:1	1:1	
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	
Probe height (meters)	10	9.0	3	
Distance from supporting structure (meters)	10	9.0	0.5	
Distance from obstructions on roof (meters)	N/A	N/A	N/A	
Distance from obstructions not on roof (meters)	35	35	35	
Distance from trees (meters)	N/A	N/A	N/A	
Distance to furnace or incinerator flue (meters)	N/A	N/A	9	
Distance between collocated monitors (meters)	N/A	N/A	N/A	
Unrestricted airflow (degrees)	270°	270°	270°	

Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A	N/A	N/A	
Residence time for reactive gases (seconds)	N/A	N/A	N/A	
Will there be changes within the next 18 months? (Y/N)	No	No	No	
Is it suitable for comparison against the annual PM _{2.5} ? (Y/N)	N/A	N/A	N/A	
Frequency of flow rate verification for manual PM samplers	N/A	N/A	N/A	
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	N/A	
Frequency of one-point QC check for gaseous instruments	N/A	N/A	N/A	
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A	N/A	N/A	
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	N/A	N/A	

Pico Rivera #2
Site Photos



Looking North from the probe.



Looking East from the probe.



Looking South from the probe.



Looking West from the probe.


**Pico Rivera #2
Site Photos (Cont.)**



Looking at the probe from the North.



Looking at the probe from the East.



Looking at the probe from the South.



Looking at the probe from the West.