


# South Coast AQMD


## Site Survey Report for San Bernardino

Last updated: June 23, 2023



AQS ID	ARB Number	Site Start Date	Reporting Agency and Agency Code
060719004	36203	05/1986	South Coast AQMD (0972)

Site Address	County	Air Basin	Latitude	Longitude	Elevation
24302 E. 4th Street San Bernardino, CA 92410	San Bernardino	South Coast	34.106678	-117.274063	316m



## Detailed Site Information

Local site name	San Bernardino			
AQS ID	060719004			
GPS coordinates (decimal degrees)	Latitude: 34.106678, Longitude: -117.274063			
Street Address	24302 E. 4 <sup>th</sup> Street, San Bernardino, CA 92410			
County	San Bernardino			
Distance to roadways (meters)	18			
Traffic count (AADT, year)	2,500 / 2012			
Groundcover (e.g. asphalt, dirt, sand)	Asphalt			
Representative statistical area name (i.e. MSA, CBSA, other)	40140-Riverside-San Bernardino-Ontario, CA MSA			
Pollutant, POC	Carbon Monoxide, 1	Nitrogen Dioxide, 1	Ozone, 1	Continuous PM10, 3
Primary / QA Collocated / Other	N/A	N/A	N/A	Primary
Parameter code	42101	42602	44201	81102
Basic monitoring objective(s)	NAAQS	NAAQS	NAAQS	NAAQS
Site type(s)	Population Exposure	Population Exposure	Highest Concentration	Population Exposure
Monitor (type)	SLAMS	SLAMS	SLAMS	SLAMS
Network affiliation	N/A	Vulnerable and susceptible population (aka RA40)	N/A	N/A
Instrument manufacturer and model	Teledyne T300EU	Teledyne T200	Teledyne T400	Thermo 5014i
Method code	093	099	087	150
FRM/FEM/ARM/ other	FRM	FRM	FEM	FEM
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	N/A	N/A	N/A
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	South Coast AQMD
Spatial scale (e.g. micro, neighborhood)	Urban	Urban	Neighborhood	Neighborhood
Monitoring start date (MM/DD/YYYY)	05/1986	05/1986	05/1986	09/01/2004
Current sampling frequency (e.g.1:3, continuous)	1:1	1:1	1:1	1;1
Calculated sampling frequency (e.g. 1:3/1:1)	N/A	N/A	N/A	N/A
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	01/01-12/31
Probe height (meters)	4.5	4.5	4.5	4.2
Distance from supporting structure (meters)	1.9	1.9	1.9	1.6
Distance from	N/A	N/A	N/A	N/A

obstructions on roof (meters)				
Distance from obstructions not on roof (meters)	N/A	N/A	N/A	N/A
Distance from trees (meters)	N/A	N/A	N/A	N/A
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	N/A
Distance between collocated monitors (meters)	N/A	N/A	N/A	2.6
Unrestricted airflow (degrees)	360°	360°	360°	360°
Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	Teflon	Teflon	Teflon	N/A
Residence time for reactive gases (seconds)	7.0	11.9	7.8	N/A
Will there be changes within the next 18 months? (Y/N)	No	No	No	No
Is it suitable for comparison against the annual PM2.5? (Y/N)	N/A	N/A	N/A	N/A
Frequency of flow rate verification for manual PM samplers	N/A	N/A	N/A	N/A
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	N/A	Monthly
Frequency of one-point QC check for gaseous instruments	Nightly	Nightly	Nightly	N/A
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	03/10/2022	03/10/2022	03/10/2022	N/A
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	N/A	N/A	04/22/2022 09/30/2022

Pollutant, POC	Lead, 2	24 Hour PM2.5, 1		
Primary / QA Collocated / Other	Primary	Primary		
Parameter code	14129	88101		
Basic monitoring objective(s)	NAAQS	NAAQS		
Site type(s)	non-source-oriented	Population Exposure		
Monitor (type)	SLAMS	SLAMS		
Network affiliation	N/A	N/A		
Instrument manufacturer and model	GMW 1200 TSP	Thermo 2025i Partisol PM2.5		
Method code	110	145		
FRM/FEM/ARM/ other	FRM	FRM		
Collecting Agency	South Coast AQMD	South Coast AQMD		
Analytical Lab (i.e., weigh lab, toxics lab, other)	South Coast AQMD	South Coast AQMD		
Reporting Agency	South Coast AQMD	South Coast AQMD		
Spatial scale (e.g. micro, neighborhood)	Neighborhood	Neighborhood		
Monitoring start date (MM/DD/YYYY)	09/1990	08/27/2008		
Current sampling frequency (e.g.1:3, continuous)	1:6	1:3		
Calculated sampling frequency (e.g. 1:3/1:1)	1:6	1:3		
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31		
Probe height (meters)	2.2	3.1		
Distance from supporting structure (meters)	1.2 *Stand itself is supporting structure.	2.1 *Stand itself is supporting structure.		
Distance from obstructions on roof (meters)	N/A	N/A		
Distance from obstructions not on roof (meters)	N/A	N/A		
Distance from trees (meters)	N/A	N/A		
Distance to furnace or incinerator flue (meters)	N/A	N/A		
Distance between collocated monitors (meters)	N/A	N/A		
Unrestricted airflow (degrees)	360°	360°		

Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A	N/A		
Residence time for reactive gases (seconds)	N/A	N/A		
Will there be changes within the next 18 months? (Y/N)	No	No		
Is it suitable for comparison against the annual PM <sub>2.5</sub> ? (Y/N)	N/A	Yes		
Frequency of flow rate verification for manual PM samplers	Monthly	Monthly		
Frequency of flow rate verification for automated PM analyzers	N/A	N/A		
Frequency of one-point QC check for gaseous instruments	N/A	N/A		
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A	N/A		
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	04/22/2022 09/30/2022	04/22/2022 09/30/2022		

Pollutant, POC	WS & D, 1/1	RH/T, 1/1		
Primary / QA Collocated / Other	N/A	N/A		
Parameter code	61101/61102	62201/62101		
Basic monitoring objective(s)	Research	Research		
Site type(s)	Meteorological	Meteorological		
Monitor (type)	SLAMS	SLAMS		
Network affiliation	N/A	N/A		
Instrument manufacturer and model	RM Young 05305V	Rotronic HC2-S3		
Method code	065/065	063/063		
FRM/FEM/ARM/ other	N/A	N/A		
Collecting Agency	South Coast AQMD	South Coast AQMD		
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	N/A		
Reporting Agency	South Coast AQMD	South Coast AQMD		
Spatial scale (e.g. micro, neighborhood)	Urban/Middle/ Neighborhood	Urban/Middle/ Neighborhood		
Monitoring start date (MM/DD/YYYY)	05/1986	05/1986		
Current sampling frequency (e.g. 1:3, continuous)	Continuous	Continuous		
Calculated sampling frequency (e.g. 1:3/1:1)	1:1	1:1		
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31		
Probe height (meters)	10	9		
Distance from supporting structure (meters)	10	9		
Distance from obstructions on roof (meters)	N/A	N/A		
Distance from obstructions not on roof (meters)	N/A	N/A		
Distance from trees (meters)	12	12		
Distance to furnace or incinerator flue (meters)	N/A	N/A		
Distance between collocated monitors (meters)	N/A	N/A		
Unrestricted airflow (degrees)	360°	360°		
Probe material for reactive gases	N/A	N/A		

(e.g. Pyrex, stainless steel, Teflon)				
Residence time for reactive gases (seconds)	N/A	N/A		
Will there be changes within the next 18 months? (Y/N)	No	No		
Is it suitable for comparison against the annual PM <sub>2.5</sub> ? (Y/N)	N/A	N/A		
Frequency of flow rate verification for manual PM samplers	N/A	N/A		
Frequency of flow rate verification for automated PM analyzers	N/A	N/A		
Frequency of one-point QC check for gaseous instruments	N/A	N/A		
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A	N/A		
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	N/A		



## San Bernardino Site Photos



**Looking North from the probe.**



**Looking East from the probe.**



**Looking South from the probe.**



**Looking West from the probe.**

**San Bernardino  
Site Photos (Cont.)**



**Looking at the probe from the North.**



**Looking at the probe from the East.**



**Looking at the probe from the South.**



**Looking at the probe from the West.**