

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO Cher Snyder, Assistant DEO Engineering & Compliance	LABORATORY NO	<u>1601347</u>
	DATE RECEIVED	<u>01/13/2016</u>
SAMPLE DESCRIBED AS Triggered Samples: Canisters: 54708, E3398, 54728, 54036	FACILITY ID NO	<u>NA</u>
	REQUESTED BY	<u>Sumner Wilson</u>
SAMPLING LOCATION Porter Ranch Community Elementary School Porter Ranch, CA 91326	ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister	Canister	Canister	Canister
Number	<u>54708</u>	<u>E3398</u>	<u>54728</u>	<u>54036</u>
Pressure (Torr)	740	740	740	740
CO, ppm	< 3	< 3	< 3	< 3
CH ₄ , ppm	9	9	15	17
CO ₂ , ppm	419	413	412	408
Ethane, ppmvC	< 1	< 1	< 1	1
NM/NEOC, ppmvC	< 1	< 1	< 1	< 1

Date Approved: 1/28/16

Approved By: 
Rudy Eden, Senior Manager
Laboratory Services
909-396-2391

1/28/16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1601347</u>
Triggered Samples: Canisters: 54708, E3398, 54728, 54036	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC91340	MDL				
CO, ppmvC	0.3	10.40	10.36	-0.39	PASS
CH ₄ , ppmvC	0.3	10.17	10.93	7.45	PASS
CO ₂ , ppmvC	0.4	10.38	10.56	1.75	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	11.00	10.58	-3.80	PASS
NM/NEOC, ppmvC	0.2	10.64	10.41	-2.11	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC135067	MDL				
CO, ppmvC	0.3	10100	10057	-0.43	PASS
CH ₄ , ppmvC	0.3	9950	10370	4.22	PASS
CO ₂ , ppmvC	0.4	10100	9992	-1.07	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	9940	10196	2.57	PASS
NM/NEOC, ppmvC	0.2	10000	10122	1.22	PASS

DATE ANALYZED 01/25/2016
REFERENCE NO: 16QM2AA
 QM2-101-68

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT INFORMATION
 INV
 LAF
 LABOI

WO #: 1601347



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 12/31/15

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Mike Koch Date: 1/12/16 Time: 14:00

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End vac
Porter Ranch Community Elementary School	54708	1-8-16 / 18:27 / 5 min	-30"	-2"
Porter Ranch Community Elementary School	E3398	1-10-16 / 15:50 / 5 min	-30"	0
Porter Ranch Community Elementary School	54728	1-10-16 / 22:47 / 5 min	-30'	0
Porter Ranch Community Elementary School	54036	1-10-16 / 23:19 / 5 min	-30'	-4'

Relinquished by	Received by	Firm/Agency	Date	Time
Mike Koch	<i>[Signature]</i>	SCAQMD Lab	1/13/16	13:41

Remarks: Samples collected by passive sampling via the XonTech 912 triggered by the Mocon NMHC.

Trigger is set to 20ppm

Porter Ranch Community School Elementary – 5 Sesnon Blvd, Porter Ranch, CA 91326

GPS (34.293369, -118.580505)