

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO Cher Snyder, Assistant DEO
Engineering & Compliance

LABORATORY NO 1607012

DATE RECEIVED 03/11/2016

SAMPLE DESCRIBED AS

24 hour sample
Canister # 54057

FACILITY ID NO NA

REQUESTED BY Sumner Wilson

SAMPLING LOCATION

Porter Ranch Community
Elementary School
Porter Ranch, CA 91326

ST NO / PROJECT NA

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane
non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54057</u>
Pressure (Torr)	740

CO, ppm	< 3
CH ₄ , ppm	3
CO ₂ , ppm	403
Ethane, ppmvC	NR
NM/NEOC, ppmvC	NR

NR: Not reported. Results may be biased low. Methane data accurate.

Date Approved: 3/25/2016

Approved By: 

Rudy Eden, Senior Manager
Laboratory Services
909-396-2391

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1607012</u>
24 hour sample Canister # 54057	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC135067	MDL				
CO, ppmvC	0.3	10100	9534	-5.60	FAIL
CH ₄ , ppmvC	0.3	9950	10082	1.33	PASS
CO ₂ , ppmvC	0.4	10100	9572	-5.23	FAIL
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	9940	9261	-6.83	FAIL
NM/NEOC, ppmvC	0.2	10000	NA	NA	FAIL

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC91340	MDL				
CO, ppmvC	0.3	10.40	9.72	-6.58	PASS
CH ₄ , ppmvC	0.3	10.17	10.39	2.15	PASS
CO ₂ , ppmvC	0.4	10.38	10.43	0.51	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.2	11.00	7.99	-27.38	FAIL
NM/NEOC, ppmvC	0.2	10.64	NA	NA	FAIL

Instrument passed pre-run quality control check, but marginally failed post-run for several compounds, not including methane. Problem was traced to an oven which does not affect methane results.

DATE ANALYZED	03/11/2016
REFERENCE NO:	16QM2AA
	QM2-101-87

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

I
 I
 I
LA.

WO #: 1607012



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 3/11/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 3/11/16 Time: 8:45am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Porter Ranch Community Elementary School (PRCS)-A	54057	3/10/16 / 00:00 / 24 hours	<-30"	+9.5
Porter Ranch Community Elementary School (PRCS)-B	54683	3/10/16 / 00:00 / 24 hours	<-30"	+8

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhou</i>	<i>[Signature]</i>	SCAQMD Lab	3/11/16	11:33

Remarks: 1:3 scheduled samples from trailer at PRCS

Porter Ranch Community School Elementary – 12450 Mason Ave, Porter Ranch, CA 91326

GPS (34.293369, -118.580505)

A: AQMD 0015735; B: AQMD 0015769