

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO	Cher Snyder, Assistant DEO Engineering & Compliance	LABORATORY NO	<u>1608813</u>
		DATE RECEIVED	<u>03/29/2016</u>
SAMPLE DESCRIBED AS	24 hour sample Canister # 54713	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
SAMPLING LOCATION	Reseda Station 18328 Gault St Los Angeles, CA 91335	ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54713</u>
Pressure (Torr)	740
CO, ppm	< 3
CH ₄ , ppm	3
CO ₂ , ppm	414
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 4/6/16

Approved By: 
Solomon Teffera, Acting Manager
Laboratory Services
909-396-2199

96
3/30/16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1608813</u>
24 hour sample Canister # 54713	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.3	1.92	2.09	8.78	PASS
CH ₄ , ppmvC	0.3	2.02	2.10	3.92	PASS
CO ₂ , ppmvC	0.3	1.57	1.94	23.58	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	2.17	7.03	PASS
NM/NEOC, ppmvC	0.4	2.03	2.12	4.67	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.3	1036	1054	1.76	PASS
CH ₄ , ppmvC	0.3	1068	1049	-1.78	PASS
CO ₂ , ppmvC	0.3	1022	1041	1.81	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1042	-0.17	PASS
NM/NEOC, ppmvC	0.4	1024	1017	-0.67	PASS

DATE ANALYZED 03/29/2016
REFERENCE NO: 16QM2AB
 QM2-101-94

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DIST
 INVC
 LAP
 LABOF

WO #: 1608813



TO: SCAQMD LAB: OTHER: _____
 SOURCE NAME: _____ Southern California Gas Co. _____ I.D. No. _____
 Source Address: _____ 12801 Tampa Ave _____ City: _____ Porter Ranch _____
 Mailing Address: _____ City: _____ Zip: _____ 91326 _____
 Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: _____ Sumner Wilson _____ Date: _____ 3/29/16 _____
 Approved by: _____ Jason Low _____ Office: _____ Budget #: _____ 44716 _____
 REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
 Suspected Violation Rule(s) _____ Other _____

Sample Collected by: _____ Qian Zhou _____ Date: _____ 3/29/16 _____ Time: _____ 10:25am _____

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Reseda Station	54713	3/28/16 / 00:00 / 24 hours	<-30"	+10

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>SL</i>	SCAQMD Lab	3/29/16	12:07

Remarks: 1:3 scheduled samples from station at Reseda
 Reseda Station – 18328 Gault St, Los Angeles, CA 91335
 GPS (34.199225, -118.532743)