

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS  
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

<b>TO</b>	Cher Snyder, Assistant DEO Engineering & Compliance	<b>LABORATORY NO</b>	<u>1609131</u>
<b>SAMPLE DESCRIBED AS</b>	24 hour sample Canister # 54557	<b>DATE RECEIVED</b>	<u>04/01/2016</u>
<b>SAMPLING LOCATION</b>	Reseda Station 18328 Gault St. Los Angeles, CA 91335	<b>FACILITY ID NO</b>	<u>NA</u>
		<b>REQUESTED BY</b>	<u>Sumner Wilson</u>
		<b>ST NO / PROJECT</b>	<u>NA</u>

---

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

Type	Canister
Number	<u>54557</u>
Pressure (Torr)	740
CO, ppm	< 3
CH <sub>4</sub> , ppm	3
CO <sub>2</sub> , ppm	431
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 4/1/16

Approved By: 

Solomon Teffera, Acting Senior Manager  
Laboratory Services  
909-396-2199

21865-1482-16

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT**  
**21865 Copley Drive, Diamond Bar CA 91765-1482**

**MONITORING & ANALYSIS**  
**QUALITY CONTROL SUMMARY**

Page 2 of 2

<b>SAMPLE DESCRIBED AS</b>	<b>LABORATORY NO</b>	<u>1609131</u>
24 hour sample Canister # 54557	<b>REQUESTED BY</b>	<u>Sumner Wilson</u>

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**


**QUALITY CONTROL -- End of run control recovery**

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.3	1.92	2.06	7.03	PASS
CH <sub>4</sub> , ppmvC	0.3	2.02	2.08	2.84	PASS
CO <sub>2</sub> , ppmvC	0.3	1.57	2.07	31.80	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	2.03	2.44	20.41	PASS
NM/NEOC, ppmvC	0.4	2.03	2.08	2.63	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.3	1036	1057	1.99	PASS
CH <sub>4</sub> , ppmvC	0.3	1068	1051	-1.57	PASS
CO <sub>2</sub> , ppmvC	0.3	1022	1043	2.09	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	1044	1045	0.10	PASS
NM/NEOC, ppmvC	0.4	1024	1022	-0.12	PASS

**DATE ANALYZED** 04/01/2016  
**REFERENCE NO:** 16QM2AB  
 QM2-101-94

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
SAMPLE ANALYSIS REQUEST**



WO #: 1609131



TO: SCAQMD LAB:  OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. \_\_\_\_\_

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: \_\_\_\_\_ City: \_\_\_\_\_ Zip: 91326

Contact Person: \_\_\_\_\_ Title: \_\_\_\_\_ Tel: \_\_\_\_\_

Analysis Requested by: Sumner Wilson Date: 4/1/16

Approved by: Jason Low Office: \_\_\_\_\_ Budget #: 44716

REASON REQUESTED: Court/Hearing Board  Permit Pending  Hazardous/Toxic Spill   
Suspected Violation Rule(s) \_\_\_\_\_ Other

Sample Collected by: Qian Zhou Date: 4/1/16 Time: 10:12am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Reseda Station	54557	3/31/16 / 00:00 / 24 hours	<-30"	+14

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>On L D oil</i>	SCAQMD Lab	4-1-16	12:05

Remarks: 1:3 scheduled samples from station at Reseda  
Reseda Station - 18328 Gault St, Los Angeles, CA 91335  
GPS (34.199225, -118.532743)