## SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT 21865 Copley Drive, Diamond Bar CA 91765-1482

### MONITORING & ANALYSIS REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO	Jason Low	LABORATORY NO	1620213
	Atmospheric Measurements Manager		
	Science and Technology Advancement	DATE RECEIVED	07/21/2016
SAM	PLE DESCRIBED AS		, <del></del>
	24 hour sample	FACILITY ID NO	NA
	Canister 54645		
		REQUESTED BY	Sumner Wilson
SAM	PLING LOCATION		
	Reseda Station	ST NO / PROJECT	NA
	18328 Gault St		
	Los Angeles, CA 91335		

Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	54645
Pressure (Torr)	760
CO, ppm	< 3
CH <sub>4</sub> , ppm	3
CO <sub>2</sub> , ppm	434
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 7/27//6

Approved By:

Solomon Teffera, Acting Senior Manager

Laboratory Services 909-396-2199

## SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT 21865 Copley Drive, Diamond Bar CA 91765-1482

### MONITORING & ANALYSIS QUALITY CONTROL SUMMARY

Page 2 of 2

#### SAMPLE DESCRIBED AS

LABORATORY NO

1620213

24 hour sample Canister 54645 **REQUESTED BY** 

Sumner Wilson

Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

### QUALITY CONTROL -- End of run control recovery

CC106783	MDL	Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CO, ppmvC	0.4	1.92	2.25	16.75	PASS
CH <sub>4</sub> , ppmvC	0.3	2.02	1.99	-1.88	PASS
CO <sub>2</sub> , ppmvC	0.4	1.57	2.46	57.06	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	2.03	1.97	-2.82	PASS
NM/NEOC, ppmvC	0.5	2.03	1.85	-8.60	PASS
CC12628	MDL	Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CO, ppmvC	0.4	1036	1064	2.67	PASS
CH <sub>4</sub> , ppmvC	0.3	1068	1061	-0.66	PASS
		7 7 7 7			IADD
CO <sub>2</sub> , ppmvC	0.4	1022	1045	2.21	PASS
CO <sub>2</sub> , ppmvC C <sub>2</sub> H <sub>4</sub> , ppmvC					
1 2000 1000 1000 1000 1000 1000 1000	0.4	1022	1045	2.21	PASS

DATE ANALYZED REFERENCE NO:

07/21/2016 16QM2AD QM2-101-117

# SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT SAMPLE ANALYSIS REQUEST

WO #: 1620213

TO: SCAQMD LAB: ⊠	OTHER:						
SOURCE NAME:	Southern Calif	fornia Gas Co	i. I.D	. No			
Source Address: 12801 Ta	ampa Ave		City:	Porter Rane	ch		
Mailing Address:		(	City:	Zip:	91326		
Analysis Requested by: Sumner Wilson Date: 7/21/16							
		v Office:					
REASON REQUESTED: 0				Hazardous/Toxi			
	Rule(s)			riazai dous/ 10x1	c spin 📋		
- Caspectou - Common			Other				
Sample Collected by:	Qian Zhou	Date:	7/21/16	Time:1	1:30am		
REQUESTED ANALYSIS: PAMS analysis							
City/Location	Can#		/ time/ duration		End Press		
Reseda Station	54645	7/20/16 /	00:00 / 24 hours	<-30"	+13		
			horities 1				
		<u> </u>					
Relinquished by Received		by	Firm/Agenc	y Date	Time		
zhougian (	Tale W Leas	2	SCAQMD La	1 7-27-1C	13:12		
	1 0						
Remarks: 1:3 scheduled samples fro							
Reseda Station – 18328 Gault St, Los GPS (34.199225, -118.532743). Rigil							