

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS  
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

**TO** Jason Low  
Atmospheric Measurements Manager  
Science and Technology Advancement

**LABORATORY NO** 1620534

**DATE RECEIVED** 07/26/2016

**SAMPLE DESCRIBED AS**  
24 hour sample  
Canister 53383

**FACILITY ID NO** NA

**REQUESTED BY** Sumner Wilson

**SAMPLING LOCATION**  
Reseda Station  
18328 Gault St  
Los Angeles, CA 91335

**ST NO / PROJECT** NA

---

---

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

Type	Canister
Number	<u>53383</u>
Pressure (Torr)	760
CO, ppm	7
CH <sub>4</sub> , ppm	2
CO <sub>2</sub> , ppm	446
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 8/2/16

Approved By: *Solomon Teffer*  
Solomon Teffer, Acting Senior Manager  
Laboratory Services  
909-396-2199

*Sumner Wilson*  
7/26/16

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT**  
**21865 Copley Drive, Diamond Bar CA 91765-1482**

**MONITORING & ANALYSIS**  
**QUALITY CONTROL SUMMARY**

Page 2 of 2

<b>SAMPLE DESCRIBED AS</b>	<b>LABORATORY NO</b>	<u>1620534</u>
24 hour sample Canister 53383	<b>REQUESTED BY</b>	<u>Sumner Wilson</u>

**Carbon monoxide (CO), methane (CH<sub>4</sub>), carbon dioxide (CO<sub>2</sub>), ethane (C<sub>2</sub>H<sub>6</sub>), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).**

**QUALITY CONTROL -- End of run control recovery**

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.09	8.43	PASS
CH <sub>4</sub> , ppmvC	0.3	2.02	2.00	-1.02	PASS
CO <sub>2</sub> , ppmvC	0.4	1.57	2.38	51.93	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	2.03	2.01	-0.86	PASS
NM/NEOC, ppmvC	0.5	2.03	2.12	4.58	PASS


		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1060	2.34	PASS
CH <sub>4</sub> , ppmvC	0.3	1068	1057	-1.02	PASS
CO <sub>2</sub> , ppmvC	0.4	1022	1042	1.95	PASS
C <sub>2</sub> H <sub>4</sub> , ppmvC	0.4	NA	NA	NA	NA
C <sub>2</sub> H <sub>6</sub> , ppmvC	0.4	1044	1043	-0.06	PASS
NM/NEOC, ppmvC	0.5	1024	1017	-0.65	PASS

<b>DATE ANALYZED</b>	07/27/2016
<b>REFERENCE NO:</b>	16QM2AD QM2-101-117

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT  
SAMPLE ANALYSIS REQUEST**

DIST  
 INVC  
 LAP  
 LABOR

WO #: 1620534



TO: SCAQMD LAB:  OTHER:   
 SOURCE NAME: Southern California Gas Co. I.D. No. \_\_\_\_\_  
 Source Address: 12801 Tampa Ave City: Porter Ranch  
 Mailing Address: \_\_\_\_\_ City: \_\_\_\_\_ Zip: 91326  
 Contact Person: \_\_\_\_\_ Title: \_\_\_\_\_ Tel: \_\_\_\_\_

Analysis Requested by: Sumner Wilson Date: 7/26/16  
 Approved by: Jason Low Office: \_\_\_\_\_ Budget #: 44716  
 REASON REQUESTED: Court/Hearing Board  Permit Pending  Hazardous/Toxic Spill   
 Suspected Violation Rule(s) \_\_\_\_\_ Other

Sample Collected by: Qian Zhou Date: 7/26/16 Time: 15:00pm

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Reseda Station	53383	7/23/16 / 00:00 / 24 hours	<-30"	+12

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>[Signature]</i>	SCAQMD Lab	7/26/2016	17:21

Remarks: 1:6 scheduled samples from station at Reseda.  
 Reseda Station - 18328 Gault St, Los Angeles, CA 91335  
 GPS (34.199225, -118.532743). Left sampler sn 4148