

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO Jason Low
ADEO
Science and Technology Advancement

LABORATORY NO 1630127

DATE RECEIVED 11/01/2016

SAMPLE DESCRIBED AS
24 hour sample
Canister 54717

FACILITY ID NO NA

REQUESTED BY Sumner Wilson


SAMPLING LOCATION
Reseda Station
18328 Gault St
Los Angeles, CA 91335

ST NO / PROJECT Southern California Gas Co.

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54717</u>
Pressure (Torr)	760
CO, ppm	6
CH ₄ , ppm	3
CO ₂ , ppm	467
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 11/15/16

Approved By: 
Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

11-15-16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1630127</u>
24 hour sample Canister 54717	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

			Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL					
CO, ppmvC	0.4		1.92	2.08	8.11	PASS
CH ₄ , ppmvC	0.3		2.02	1.94	-4.39	PASS
CO ₂ , ppmvC	0.4		1.57	8.06	413.62	FAIL
C ₂ H ₄ , ppmvC	0.4		NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4		2.03	1.94	-4.24	PASS
NM/NEOC, ppmvC	0.5		2.03	2.11	3.94	PASS

			Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL					
CO, ppmvC	0.4		1036	1024	-1.19	PASS
CH ₄ , ppmvC	0.3		1068	1018	-4.66	PASS
CO ₂ , ppmvC	0.4		1022	1019	-0.33	PASS
C ₂ H ₄ , ppmvC	0.4		NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4		1044	1009	-3.38	PASS
NM/NEOC, ppmvC	0.5		1024	990	-3.25	PASS

DATE ANALYZED	11/02/2016
REFERENCE NO:	16QM2AD QM2-101-129

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

WO #: 1630127

DIST
 INV
 LAP
LABOR



TO: SCAQMD LAB: OTHER: _____
 SOURCE NAME: Southern California Gas Co. I.D. No. _____
 Source Address: 12801 Tampa Ave City: Porter Ranch
 Mailing Address: _____ City: _____ Zip: 91326
 Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 11/01/16
 Approved by: Jason Low Office: _____ Budget #: 44716
 REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill
 Suspected Violation Rule(s) _____ Other _____

Sample Collected by: Bo Vongphachanh Date: 10/28/16 Time: 11:35am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Reseda Station	54717	10/27/16 / 00:00 / 24 hours	-30"	14"

Relinquished by	Received by	Firm/Agency	Date	Time
		SCAQMD Lab	11/1/2016	9:58

Remarks: 1:3 scheduled samples from station at Reseda.
 Reseda Station – 18328 Gault St, Los Angeles, CA 91335
 GPS (34.199225, -118.532743). Right sampler sn 6113