

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2

TO Jason Low, Ph.D., ADEO Monitoring & Analysis	LABORATORY NO. <u>1700726</u>
SAMPLE DESCRIBED AS 24 hour sample Canister 54588	DATE RECEIVED <u>01/13/2017</u>
SAMPLING LOCATION Reseda Station 18328 Gault St Los Angeles, CA 91335	FACILITY ID NO. <u>NA</u>
	REQUESTED BY <u>Sumner Wilson</u>
	PROJECT/ST NO. <u>NA</u>
	REPORT GENERATED <u>1/17/2017</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54588</u>
Pressure (Torr)	1307
CO, ppm	< 3
CH ₄ , ppm	3
CO ₂ , ppm	485
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Reviewed: 1-25-17

Reviewed By: Joan Niertit For JFN
Joan Niertit, Principle A.Q. Chemist
Laboratory Services

Date Approved: 2-7-17

Approved By: Aaron Katzenstein
Aaron Katzenstein, Ph.D., Sr. Manager
Laboratory Services
909-396-2219

AM
1-25-17

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO.	<u>1700726</u>
24 hour sample Canister 54588	REQUESTED BY	<u>Sumner Wilson</u>
	DATE ANALYZED	<u>1/13/2017</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.13	10.59	PASS
CH ₄ , ppmvC	0.3	2.02	2.11	4.16	PASS
CO ₂ , ppmvC	0.4	1.57	2.02	28.76	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	2.06	1.24	PASS
NM/NEOC, ppmvC	0.5	2.03	2.22	9.29	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1051	1.44	PASS
CH ₄ , ppmvC	0.3	1068	1049	-1.79	PASS
CO ₂ , ppmvC	0.4	1022	1033	1.04	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1032	-1.14	PASS
NM/NEOC, ppmvC	0.5	1024	1003	-2.01	PASS

REFERENCE NO: 17QM2AA
 QM2-101-135

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DI
 IN
 LA
LABC

WO #: 1700726



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 01/13/17

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Bo Vongphachanh Date: 01/12/17 Time: 13:15pm

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Reseda Station	54588	01/07/17 / 00:00 / 24 hours	-26"	13"

Relinquished by	Received by	Firm/Agency	Date	Time
		SCAQMD Lab	1/13/2017	9:45

Remarks: 1:3 scheduled samples from station at Reseda.
Reseda Station - 18328 Gault St, Los Angeles, CA 91335
GPS (34.199225, -118.532743). Right sampler sn 6113