



Date Tested: March 24, 2021
Date Issued: April 13, 2021
Revision Number: --

**COMPLIANCE SOURCE TEST REPORT:
ONE (1) HOT OIL HEATER
ALL AMERICAN ASPHALT – IRVINE FACILITY
SCAQMD A/N 622276
ID #082207**

Source Location:
ALL AMERICAN ASPHALT
10600 Jeffrey Road
Irvine, California 92602

STID: PR21045

STA Received:
4/23/2021

Submitted to:
ALL AMERICAN ASPHALT
P.O. Box 2229
Corona, California 91719

Attention: John Gardner


For Submittal to:
South Coast Air Quality Management District
21854 Copley Drive
Diamond Bar, California 91765-4178

Prepared By:
AIRx Testing Services, Inc.
2472 Eastman Avenue Unit #34
Ventura, CA 93003

Job Number
1064

Laboratory Report Number
221-028

Test Team Leader
Ken Kennepohl


Tom Porter, Vice President of Testing Services


Ken Kennepohl, Project Engineer

AIR TESTING SERVICES, INC.

TABLE OF CONTENTS

<u>Section</u>	<u>Page</u>
1.0 SUMMARY	1-1
2.0 INTRODUCTION	2-1
3.0 SOURCE PROCESS AND EQUIPMENT DESCRIPTION.	3-1
4.0 TEST PROCEDURES	4-1
5.0 TEST RESULTS AND DISCUSSION	5-1
6.0 QUALITY ASSURANCE	6-1
Appendix A Method 100.1 Data and Calculation Sheets	A-1
Appendix B Flowrate & Fuel Data Sheets	B-1
Appendix C Flowrate Methods 1- 4 Data Sheets	C-1



1.0 SUMMARY

Source Tested:	All American Asphalt Hot Oil Heater - Crumb Rubber Plant
Test Location:	All American Asphalt 10600 Jeffrey Road Irvine, California
Test Requested by:	SCAQMD
Test Objectives	Determine for compliance reporting: Emissions of NOx & CO for SCAQMD Rule 1146 & Rule 441
Test Performed by:	AIRx Testing Services, Inc.
Personnel:	Ken Kennepohl & Ferodie Torres
Test Methodology:	O2, NOx, CO2 & CO: SCAQMD Method 100.1
Test Observed by:	--
Plant Contact:	John Gardner
Facility ID Number:	082207
SCAQMD Application Number:	622276
SCAQMD Permit Number:	N/A



2.0 INTRODUCTION

AIRx Testing Services, Inc. was contracted by All American Asphalt to conduct source testing on one (1) oil heater located at their facility located at 10600 Jeffrey Road, Irvine, California.

Source testing was conducted by AIRx Testing Services, Inc. personnel on March 24, 2021 to measure carbon monoxide, oxides of nitrogen and oxygen emissions from a natural gas fired hot oil heater. The testing was performed at a low load of 32.6% operating load & maximum operating load of 98%. The subject testing was performed to show compliance with SCAQMD Rule 1146 & Rule 441. The test program was conducted by Ken Kennepohl & Ferodie Torres with AIRx Testing Services, Inc. Arrangements for the testing were made through John Gardner with All American Asphalt.



3.0 SOURCE PROCESS AND EQUIPMENT DESCRIPTION

The device is identified as follows:

Modification of the current 7,600,000 BTU/Hr Process Oil Heater by installation of a 4,500,000 BTU/HR CEI Enterprises, Model HDI-400, Natural Gas Fired Low NO_x Burner, Power Flame, Model NP2R-G-520, with Flue Gas Recirculation, and a 10-Hp blower.

The subject unit is equipped with a 20 inch round exhaust stack with a 1 inch sample port located 42 inches upstream of any disturbance and 20 inches downstream from the stack exit.



4.0 TEST PROCEDURES

The source testing was conducted according to applicable SCAQMD test methodology. SCAQMD Method 100.1 was followed for carbon monoxide, oxides of nitrogen, carbon dioxide and oxygen determinations.

SAMPLING AND ANALYTICAL PROCEDURES

STACK GAS ANALYSIS: Samples of the stack gas were taken through a heated sample line from the exhaust stack to an iced water dropout and then to a stack gas conditioner yielding a conditioned gas to a point of <37F and analyzed for oxides of nitrogen, oxygen and carbon monoxide using SCAQMD 100.1. The oxygen was determined by a Servomex Model 1400 paramagnetic analyzer. The carbon monoxide was analyzed with an API Model 300EM infrared analyzer with a gas correlation filter system. The oxides of nitrogen were analyzed with an API Model 200EH chemiluminescent analyzer. The Carbon Dioxide concentrations was determined by using a Servomex Model 1400 NDIR analyzer. Data was recorded with a data acquisition system and also on a 10" strip chart. Data was reduced to show one (1) minute averages using appropriate spreadsheets. A 30 minute run was performed on the hot oil heater exhaust stack for the determination of emissions concentrations at each of two load conditions of maximum and less than 35% as per Rule 1146.

Three (3) point calibration error checks using two (2) span gases and zero gas were performed before and after each emissions concentration run. Additionally, bias check (probe tip) calibrations were performed at the beginning and end of each test and used for correcting analyzer measurements. For details of the sampling and measurement systems please refer to the attachments.

EXHAUST FLOW RATE: The stack exhaust flow rate was calculated using the fuel expansion factor of natural gas (EPA 8710 dscf/MMBtu), the measured fuel consumption rate and the measured effluent O₂ concentrations. All values were calculated at SCAQMD standard conditions (60°F & 29.92 in. Hg).

SCAQMD Methods 1.1-4.1 were utilized as well to determine exhaust flow rate and moisture content. The reference method data was used in the mass calculations.

The fuel usage was recorded by AIRx personnel from the existing dedicated fuel meter. The "F" Factor for the lb/MMBTU calculation was taken from EPA Method 19 - Table 19.1. The fuel meter readings were assumed to be corrected to standard conditions of 29.92 "Hg and 60°F.



5.0 TEST RESULTS AND DISCUSSION

The testing for carbon monoxide, oxides of nitrogen, carbon dioxide and oxygen were conducted while the unit was operating at two operating conditions: low load of less than 35% and maximum load.

The results of the emissions tests are presented on pages 5-2 & 5-3.

AIR TESTING SERVICES, INC.

SUMMARY OF SOURCE TEST RESULTS

All American Asphalt
Irvine
3/24/2021

Hot Oil Heater Low

CONSTITUENT	RESULTS	ALLOWABLE
Oxides of Nitrogen (NOx)		
ppmv	2.6	-
ppmv @ 3% O ₂	4.1	9
lb/hr	0.0064	-
lb/MMBtu	0.0051	-
Carbon Monoxide (CO) (<i>Actual Observed</i>)		
ppmv	0.1	-
ppmv @ 3% O ₂	0.1	400
lb/hr	0.000076	-
lb/MMBtu	0.000060	-
Carbon Monoxide (CO) (<i>20% Full Scale</i>)		
ppmv	< 2.0	-
ppmv @ 3% O ₂	< 3.2	400
lb/hr	< 0.0030	-
lb/MMBtu	< 0.0024	-
Oxygen, percent	9.5	-
Exhaust Flowrate, dscfm	337	-
Fuel Usage, dscfm	23.3	-
Heat Input during Test	32.6%	< 35%



SUMMARY OF SOURCE TEST RESULTS

All American Asphalt

Irvine

3/24/2021

Maximum

Hot Oil Heater max

CONSTITUENT	RESULTS	ALLOWABLE
Oxides of Nitrogen (NOx)		
ppmv	2.1	-
ppmv @ 3% O ₂	3.0	9
lb/hr	0.015	-
lb/MMBtu	0.0037	-
Carbon Monoxide (CO) (<i>Actual Observed</i>)		
ppmv	0.1	-
ppmv @ 3% O ₂	0.2	400
lb/hr	0.00051	-
lb/MMBtu	0.00012	-
Carbon Monoxide (CO) (<i>20% Full Scale</i>)		
ppmv	< 2.0	-
ppmv @ 3% O ₂	< 2.9	400
lb/hr	< 0.0088	-
lb/MMBtu	< 0.0022	-
Oxygen, percent	8.5	-
Exhaust Flowrate, dscfm	998	-
Fuel Usage, dscfm	70.0	-
Heat Input during Test	98%	-



6.0 QUALITYASSURANCE

Quality control procedures used during the continuous emissions portion of the test included the use of nonreactive stainless steel and Teflon sample lines and fittings, a sample system that operates under positive rather than negative pressure, pre-test and post test leak checks, zero and span calibrations, multi-range linearity checks, system calibration and interference checks. Zero and Span drift errors were corrected by using the zero and span values obtained during each sample run from the initial and final calibrations. During the test runs, the total sample flow rates and pressures from the sample extraction and conditioning system and flows to the analyzers were normal, indicating good equipment operation. The sample gas conditioning unit was maintained at <37°F. The field data sheets in Appendix A summarize the performance of the CEMS during this source test.

The analyzers used for the continuous emissions monitoring of CO, Nox, CO2 and O2 have been approved by the California Air Resources Board. The analyzers were calibrated before and after each test run to ensure the validity of the test and a probe tip calibration through the entire sampling system (systems check) was also done to determine if any interference was caused by the sampling system. No interferences were detected. The zero and calibration gases were prepared to SCAQMD specifications for Method 100.1. Certification of analysis for the calibration gases used during this test is provided in Appendix A.



CONFFLICT OF INTEREST NEGATIVE DECLARATION

AIRx Testing is an independent emissions testing contractor.

AIRx Testing maintains that no conflict of interest exists between the partners and employees of AIRx Testing, and the partners, employees or interests involved in the facility detailed in this report.

INDEPENDENT CONTRACTOR

A handwritten signature in black ink, appearing to read "Tom Porter".

Signature

Tom Porter – Vice President

4/13/2021

Date of Signature



APPENDIX A METHOD 100.1

CONTINUOUS EMISSIONS MONITORING - CARB METHOD 100

Client : All American Asphalt
Site : Irvine
Unit : Hot Oil Heater Low

Date : 3/24/2021
Job# : 7019
Lab# : 213-039

FIELD DATA

Test Length 60 mins. Points -

Standard Temperature, T std: 60 °F

Drift Corrected Emissions Data

<i>20% FS</i> <i>Actual</i>	NOx	<u>2.61</u>	ppmv
	CO	<u>< 2.0</u>	ppmv
	CO	<u>0.05</u>	ppmv
	O2	<u>9.54</u>	%

Flow Data

Q std 337 dscfm

Equations used:

$$\text{NOx or CO @ 3% O}_2 = [\text{ppmv}] * (17.9/(20.9-\text{O}_2\%))$$

$$\text{T Factor} = (10^{-6} * (29.92 / (21.85 * (460 + \text{Tstd}))) * 60$$

$$\text{lb/hr} = [\text{ppmv}] * \text{Q std} * \text{MW} * (\text{T Factor})$$

$$\text{lb/MMBtu} = \text{F-Factor} * \text{lb/hr} / (60 * \text{Qstd}) * 20.9 / (20.9 - \text{O}_2)$$

$$\text{MW NOx} = 46; \text{ MW CO} = 28$$

CALCULATED EMISSIONS

<i>20% FS</i> <i>Actual</i>	NOx	<u>2.61</u>	ppmv
		<u>4.11</u>	ppmv @ 3% O ₂
		<u>0.0064</u>	lb/hr
		<u>0.0051</u>	lb/MMBtu
	CO	<u>< 0.05</u>	ppmv
		<u>< 0.08</u>	ppmv @ 3% O ₂
		<u>< 0.000076</u>	lb/hr
		<u>< 0.000060</u>	lb/MMBtu
	CO	<u>< 2.0</u>	ppmv
		<u>< 3.2</u>	ppmv @ 3% O ₂
		<u>< 0.0030</u>	lb/hr
		<u>< 0.0024</u>	lb/MMBtu

CONTINUOUS EMISSIONS MONITORING - CARB METHOD 100

Client : All American Asphalt
 Site : Irvine
 Unit : Hot Oil Heater max

Date : 3/24/2021
 Job# : 7019
 Lab# : 213-039

FIELD DATA

Test Length 60 mins.

Points -

Standard Temperature, T std: 60 °F

Drift Corrected Emissions Data

<i>20% FS</i> <i>Actual</i>	NOx	2.09	ppmv
	CO	< 2.0	ppmv
	CO	0.12	ppmv
	O2	8.49	%

Flow Data

Q std 998 dscfm

Equations used:

$$\text{NOx or CO @ 3% O}_2 = [\text{ppmv}] * (17.9/(20.9-\text{O}_2\%))$$

$$\text{T Factor} = (10^{-6} * (29.92 / (21.85 * (460 + \text{Tstd}))) * 60$$

$$\text{lb/hr} = [\text{ppmv}] * \text{Q std} * \text{MW} * (\text{T Factor})$$

$$\text{lb/MMBtu} = \text{F-Factor} * \text{lb/hr} / (60 * \text{Qstd}) * 20.9 / (20.9 - \text{O}_2)$$

$$\text{MW NOx} = 46; \text{ MW CO} = 28$$

CALCULATED EMISSIONS

<i>Actual</i>	NOx	2.09	ppmv
		3.02	ppmv @ 3% O ₂
		0.015	lb/hr
		0.0037	lb/MMBtu
<i>Actual</i>	CO	0.1	ppmv
		0.2	ppmv @ 3% O ₂
		0.00051	lb/hr
		0.00012	lb/MMBtu
<i>20% FS</i>	CO	< 2.0	ppmv
		< 2.9	ppmv @ 3% O ₂
		< 0.0088	lb/hr
		< 0.0022	lb/MMBtu

SCAQMD Method 100.1 Linearity

All American Asphalt

Source: Hot Oil Heater

Lab No.: 221-028

Test Date: 3-24-2021

PRETEST

LEAK CHECK

** LINEARITY CHECK **

RANGE :	25 O2	10 CO	5 NOx	10 CO2
ZERO				
Instrument	0.0	0.0	0.0	0.0
Cylinder	0.0	0.0	0.0	0.0
Difference (%)	0.0	-0.1	0.0	-0.1
LOW LEVEL				
Instrument	0.0	0.0	0.0	0.0
Cylinder				
Difference (%)				
MID LEVEL				
Instrument	12.0	3.9	2.1	4.0
Cylinder	12.0	3.89	2.10	4.01
Difference (%)	-0.2	-0.3	0.2	-0.2
HIGH LEVEL				
Instrument	20.1	8.1	4.0	7.9
Cylinder	20.1	8.12	4.04	7.93
Difference (%)	0.0	0.3	0.2	0.0

POST TEST

LEAK CHECK

	25 O2	10 CO	5 NOx	10 CO2
ZERO				
Instrument	0.0	0.0	0.0	0.0
Cylinder	0.0	0.0	0.0	0.0
Difference (%)	-0.1	-0.3	-0.8	-0.4
LOW LEVEL				
Instrument	0.0	0.0	0.0	0.0
Cylinder				
Difference (%)				
MID LEVEL				
Instrument	12.0	3.9	2.1	4.0
Cylinder	12.0	3.89	2.10	4.01
Difference (%)	0.0	-0.3	0.0	0.2
HIGH LEVEL				
Instrument	20.1	8.1	4.0	7.9
Cylinder	20.1	8.12	4.04	7.93
Difference (%)	0.0	0.3	0.4	0.2

SYSTEM BIAS	SAMPLE CONC.
PreTest	

POINT	NOX	O2	TRAVERSED
c			
A-1	2.1		
c	2.10		
A-2	0.2		
c	pass		
A-3			
c			
A-4			
c			
B-1			
c			
B-2			
c			
B-3			
c			
B-4			
c			
AVERAGE			0 #DIV/0!
SAMP. PT.			

Post test

System Response Time (seconds)	#1	#2	#3
Upscale			
NOx	2.1		
CO	2.09		
O2	-0.6		
CO2	pass		
Downscale			
NOx			
CO			
O2			
CO2			

NO2 to NO Converter Efficiency (%)	cylinder	instr.	efficienc
ppm	19.0	18.1	95.0

EMISSIONS TEST - SCAQMD Method 100.1

Date: 3/24/2021

Page : 1 of 2

** Test Information **

** Personnel **

Client: All American Asphalt
 County: SCAQMD
 Site(s): Irvine
 Unit(s): Hot Oil Heater Low

AIRx: FT/KK
 Client: JG
 APCD: None

Run Length:	60	Inlet () S.T.(X)	Outlet(X) E.I.()	Fuel () Data()	HC's () H2S ()	NH3 ()
-------------	----	----------------------	----------------------	---------------------	---------------------	---------

** Instrument Information **

Instrument	"On"	Unit#	Make/Model
Outlet NOx:	1	1	API 200EH
O2:	1	12	Servomex 1400
CO:	1	3	API 300EM
CO2:	1	9	Servomex 1440

** Calibration Information **

	Units	Zero	Span	Range	Gas Cyl.#	Gas Flow
NOx:	ppmv	0	2.10	5	DT0027640	1
NOx:	ppmv	0	4.04	5	DT0035144	1
NOx:	ppmv	0	20.2	25	CC724456	1
O2:	%	0	12.0	25	CC134279	0.6
O2:	%	0	20.1	25	CC144381	0.6
CO:	ppmv	0	3.89	10	CC28151	1
CO:	ppmv	0	8.12	10	CC196750	1
NO2	ppmv	0	19.0	25	CC3240	1
CO2:	%	0	4.01	10	CC134279	0.6
CO2:	%	0	7.93	10	CC144381	0.6

** Recorder Information **

Recorder Type:	Soltec 6 pin			10	cm/hr
	Chanl.	Pen Type	Color		
Outlet NOx:	ppmv	1	Cont.	Black	
O2:	%	2	Cont.	Green	
CO:	ppmv	3	Cont.	Purple	
CO2:	%	4	Cont.	Blue	

EMISSIONS TEST - CARB Method 100

Date: 3/24/2021
 Client: All American Asphalt
 Site: Irvine

Page : 2 of 2
 Unit : Hot Oil Heater Low
 Run Number : 1

** Measured Emissions Components and Calibration Gas Concentrations **

Source:	Outlet NOx	Outlet O2	Outlet CO	Outlet CO2	Start Run @: 8:40
Component:	ppmv	%	ppmv	%	End Run @: 9:10
Span Gas Conc:	2.10	12.0	3.89	4.01	
Analyzer Range:	5	25	10	10	

** Initial Calibration Values **

Bias Zero:	0.0	0.0	0.0	0.0	NOx ppmv	CO ppmv
Bias Span:	2.1	12.0	3.8	4.0	@ 3% O2	@ 3% O2

** Raw Emission Data **

8:40	2.48	9.58	0.01	6.18	3.9	0.0
8:41	2.54	9.58	0.01	6.18	4.0	0.0
8:42	2.53	9.58	0.00	6.18	4.0	0.0
8:43	2.51	9.58	0.01	6.18	4.0	0.0
8:44	2.51	9.58	0.01	6.18	4.0	0.0
8:45	2.57	9.58	0.01	6.18	4.1	0.0
8:46	2.47	9.58	0.01	6.18	3.9	0.0
8:47	2.56	9.58	0.01	6.18	4.0	0.0
8:48	2.65	9.58	0.01	6.17	4.2	0.0
8:49	2.60	9.57	0.01	6.17	4.1	0.0
8:50	2.50	9.58	0.01	6.18	4.0	0.0
8:51	2.61	9.59	0.01	6.18	4.1	0.0
8:52	2.63	9.60	0.01	6.19	4.2	0.0
8:53	2.67	9.66	0.01	6.22	4.3	0.0
8:54	2.73	9.67	0.01	6.23	4.4	0.0
8:55	2.61	10.43	0.01	6.23	4.7	0.0
8:56	2.73	10.19	0.01	6.24	4.6	0.0
8:57	2.65	9.68	0.03	6.25	4.2	0.0
8:58	2.69	9.65	0.08	6.25	4.3	0.1
8:59	2.72	9.65	0.01	6.25	4.3	0.0
9:00	2.57	9.57	0.03	6.21	4.1	0.0
9:01	2.63	9.44	0.06	6.13	4.1	0.1
9:02	2.67	9.36	0.11	6.08	4.1	0.2
9:03	2.50	9.31	0.13	6.04	3.9	0.2
9:04	2.62	9.28	0.15	6.02	4.0	0.2
9:05	2.68	9.25	0.11	6.01	4.1	0.2
9:06	2.57	9.24	0.15	6.00	3.9	0.2
9:07	2.52	9.24	0.10	6.00	3.9	0.2
9:08	2.56	9.24	0.12	6.00	3.9	0.2
9:09	2.59	9.24	0.18	6.00	4.0	0.3
9:10	2.52	9.24	0.15	6.00	3.9	0.2
Raw Avg:	2.59	9.55	0.05	6.14	4.1	0.1

** Final Calibration Values **

Bias Zero:	0.0	0.0	0.0	0.0
Bias Span:	2.1	12.0	3.9	4.0
zero drift, %:	-0.2	-0.1	0.0	0.0
span drift, %:	0.2	0.1	0.5	0.0

** System Bias / Drift Corrected Emissions **

Average:	2.61	9.54	0.05	6.12
----------	------	------	------	------

Avg = (RawAvg-Avg Zero Response)*SpanConc/(Avg Span Response-Avg Zero Response)

Final Internal Locomotive

Set 1
0202160

Cor 11
0211170

Set 2
0202151

Set 1
0202160

Mon 10/13/15
N.Y.C. 2014
N.Y.C. 2014
N.Y.C. 2014
N.Y.C. 2014
N.Y.C. 2014
N.Y.C. 2014
N.Y.C. 2014

No 2 Due
No 2 Due

Conus on

6



4

三

Skarp
run
30%
d. 16

Con_{0.6}
 NO_x
 L-1
 Treated Gas Cet
 All
 Burner
 Sights
 221-028
 3-24-21
 10 vol.
 10% NO_x
 0.5% O₂
 0.1% CO
 0.1% CO₂
 3% H₂O
 27% O₂

AIR TESTING SERVICES Inc.

C.E.M. TEMPERATURE DATA

Facility: All Hawaiian Asphalt

Date: 3-24-21

Job No: 221-028

Run #: 1 30%

Source: Oil Heater

Probe Temp. Settings: 250 deg. F

Heated Line Temp. Settings: 250 deg. F

TIME	TEMPERATURES Deg. F		
	Condenser Outlet	Probe	Teflon Line
1 0	34	252	255
2 5	34	254	253
3 10	35	253	256
4 15	35	250	256
5 20	35	251	254
6 25	34	254	254
7 30	35	252	255
8			
9			
10			
11			
12			
13			
14			
15			

EMISSIONS TEST - SCAQMD Method 100.1

Date: 3/24/2021

Page : 1 of 2

**** Test Information ****

Client: All American Asphalt
 County: SCAQMD
 Site(s): Irvine
 Unit(s): Hot Oil Heater max

**** Personnel ****

AIRx: FT/KK
 Client: JG
 APCD: None

Run Length:	60	Inlet () S.T.(X)	Outlet(X) E.I.()	Fuel () Data()	HC's() H2S()	NH3()
-------------	----	----------------------	----------------------	---------------------	-------------------	--------

**** Instrument Information ****

Instrument	"On"	Unit#	Make/Model
Outlet NOx:	1	1	API 200EH
O2:	1	12	Servomex 1400
CO:	1	3	API 300EM
CO2:	1	9	Servomex 1440

**** Calibration Information ****

	Units	Zero	Span	Range	Gas Cyl.#	Gas Flow
NOx:	ppmv	0	2.10	5	DT0027640	1
NOx:	ppmv	0	4.04	5	DT0035144	1
NOx:	ppmv	0	20.2	25	CC724456	1
O2:	%	0	12.0	25	CC134279	0.6
O2:	%	0	20.1	25	CC144381	0.6
CO:	ppmv	0	3.89	10	CC28151	1
CO:	ppmv	0	8.12	10	CC196750	1
NO2	ppmv	0	19.0	25	CC3240	1
CO2:	%	0	4.01	10	CC134279	0.6
CO2:	%	0	7.93	10	CC144381	0.6

**** Recorder Information ****

Recorder Type:	Soltec 6 pin			10	cm/hr
Outlet NOx:	ppmv	Chanl.	Pen Type	Color	
O2:	%	1	Cont.	Black	
CO:	ppmv	2	Cont.	Green	
CO2:	%	3	Cont.	Purple	
		4	Cont.	Blue	

EMISSIONS TEST - CARB Method 100

Date: 3/24/2021
 Client: All American Asphalt
 Site: Irvine

Page : 2 of 2
 Unit : Hot Oil Heater max
 Run Number : 1

** Measured Emissions Components and Calibration Gas Concentrations **

Source:	Outlet NOx	Outlet O2	Outlet CO	Outlet CO2	Start Run @: 9:40
Component:	ppmv	%	ppmv	%	End Run @: 10:10
Span Gas Conc:	2.10	12.0	3.89	4.01	
Analyzer Range:	5	25	10	10	

** Initial Calibration Values **

Bias Zero:	0.0	0.0	0.0	0.0	NOx ppmv	CO ppmv
Bias Span:	2.1	12.0	3.9	4.0	@ 3% O2	@ 3% O2

** Raw Emission Data **

9:40	2.06	8.40	0.18	6.70	3.0	0.3
9:41	2.02	8.40	0.13	6.70	2.9	0.2
9:42	2.06	8.40	0.07	6.70	2.9	0.1
9:43	2.06	8.40	0.15	6.70	3.0	0.2
9:44	2.01	8.40	0.14	6.70	2.9	0.2
9:45	2.13	8.40	0.12	6.70	3.1	0.2
9:46	2.13	8.40	0.11	6.70	3.1	0.2
9:47	2.08	8.40	0.11	6.70	3.0	0.2
9:48	2.06	8.40	0.13	6.70	3.0	0.2
9:49	2.06	8.40	0.13	6.70	3.0	0.2
9:50	2.13	8.40	0.12	6.70	3.1	0.2
9:51	2.14	8.40	0.12	6.70	3.1	0.2
9:52	2.03	8.40	0.12	6.70	2.9	0.2
9:53	2.04	8.40	0.11	6.70	2.9	0.2
9:54	2.13	8.40	0.13	6.70	3.1	0.2
9:55	1.99	8.40	0.08	6.70	2.9	0.1
9:56	2.10	8.97	0.14	6.70	3.3	0.2
9:57	1.99	10.75	0.07	6.71	4.4	0.2
9:58	2.02	8.40	0.06	6.70	2.9	0.1
9:59	2.05	8.39	0.07	6.70	2.9	0.1
10:00	2.16	8.39	0.06	6.70	3.1	0.1
10:01	2.06	8.39	0.12	6.70	3.0	0.2
10:02	2.14	8.39	0.15	6.70	3.1	0.2
10:03	2.15	8.39	0.16	6.70	3.1	0.2
10:04	2.12	8.39	0.08	6.70	3.0	0.1
10:05	2.08	8.39	0.10	6.70	3.0	0.2
10:06	2.15	8.39	0.13	6.70	3.1	0.2
10:07	2.23	8.39	0.11	6.70	3.2	0.2
10:08	2.23	8.39	0.14	6.70	3.2	0.2
10:09	2.17	8.39	0.13	6.70	3.1	0.2
10:10	2.23	8.39	0.12	6.70	3.2	0.2
Raw Avg:	2.10	8.49	0.12	6.70	3.1	0.2

** Final Calibration Values **

Bias Zero:	0.0	0.0	0.0	0.0
Bias Span:	2.1	12.0	3.9	4.0
zero drift, %:	0.0	0.0	0.0	0.0
span drift, %:	0.6	0.0	0.2	0.0

** System Bias / Drift Corrected Emissions **

Average:	2.09	8.49	0.12	6.67
----------	------	------	------	------

Avg= (RawAvg-Avg Zero Response)*SpanConc/(Avg Span Response-Avg Zero Response)

82

80

80

CO₂ 384CO₂
4.0Not
inCO₂
384

Internal con

CO₂
384
Carb
Not
in
Bias adStark
K-MAT 4.0
55.0 1.0

CO_{2,84}CO_{2,84}CO_{2,84}CO_{2,84}CO_{2,84}CO_{2,84}NOT_{2,84} Post External Cleaning calCO_{2,84}CO_{2,84}NOT_{2,84}

Post Bias Cal


MARK TESTING SERVICES INC.
C.E.M. TEMPERATURE DATA

 Facility: All Hemwin Asphalt

 Date: 3-24-21

 Job No: 221-028

 Run #: 2 MAX

 Source: O.T. Henton

 Probe Temp. Settings: 250 deg. F

 Heated Line Temp. Settings: 250 deg. F

		TEMPERATURES Deg. F		
	TIME	Condenser Outlet	Probe	Teflon Line
1	0	35	252	255
2	5	35	250	256
3	10	35	253	257
4	15	36	251	254
5	20	36	250	255
6	25	34	250	254
7	30	35	252	255
8				
9				
10				
11				
12				
13				
14				
15				

Job Number 221-028
 Run Number 1
 Client All American Asphalt
 Plant Irvine
 Unit Heater
 Operator Ken Kennepohl
 Start Timestamp 3/24/2021 7:32

Raw Data

Timestamp (s)	NOX (ppm)	CO (ppm)	O2 (%)	CO2 (%)	
7:33	0.00	0.01	0.01	0.01	Zero;
7:34	0.01	0.01	0.03	0.01	
7:35	0.01	0.01	0.03	0.01	
7:36	2.23	0.01	0.52	0.01	
7:37	16.65	0.00	0.22	0.01	
7:38	19.13	0.01	0.06	0.01	
7:39	20.18	0.01	0.05	0.01	Nox 20.2;
7:40	12.52	0.01	0.04	0.01	
7:41	0.09	0.01	0.04	0.01	
7:42	0.12	0.01	0.05	0.04	
7:43	4.51	0.01	1.95	0.04	
7:44	8.81	0.01	2.13	0.04	
7:45	17.64	0.01	2.13	0.04	
7:46	18.05	0.01	0.80	0.04	NO2 18.05;
7:47	10.30	0.01	0.06	0.04	
7:48	2.02	0.01	0.00	0.01	
7:49	0.61	0.01	0.00	0.01	
7:50	0.01	0.01	0.00	0.01	All Zero;
7:51	0.06	0.01	0.00	0.01	
7:52	1.57	0.04	0.72	0.01	
7:53	2.51	0.02	0.06	0.01	
7:54	3.17	0.01	0.01	0.01	
7:55	4.03	0.01	0.04	0.01	High Nox;
7:56	4.01	0.01	0.12	0.01	
7:57	2.34	0.00	0.01	0.01	
7:58	2.11	0.00	0.00	0.04	Low Nox;
7:59	1.80	0.01	13.04	4.71	
8:00	0.92	0.01	20.14	7.93	High O2 Co2;
8:01	0.36	0.01	20.02	7.93	
8:02	0.11	0.01	15.34	5.80	
8:03	0.08	0.01	12.03	4.11	
8:04	0.04	0.01	12.01	4.03	Low O2 Co2;
8:05	0.05	2.29	4.46	1.54	
8:06	0.04	5.23	1.37	0.64	
8:07	0.04	7.77	0.24	0.08	
8:08	0.02	8.02	0.10	0.05	
8:09	0.01	8.09	0.17	0.03	High CO;
8:10	0.01	6.62	0.10	0.03	
8:11	0.01	4.45	0.08	0.01	
8:12	0.01	4.36	0.01	0.01	
8:13	0.01	4.13	0.01	0.01	

8:14	0.01	4.00	0.01	0.01		
8:15	0.01	3.92	0.01	0.01	Low CO;	
8:16	0.01	3.33	0.01	0.01		
8:17	0.01	0.74	0.01	0.01		
8:18	0.01	0.37	0.01	0.01		
8:19	0.01	0.00	4.36	0.01		
8:20	0.01	0.04	2.57	0.01		
8:21	0.01	0.00	0.03	0.01	Bias Zero;	
8:22	0.01	0.00	0.01	0.01		
8:23	0.04	0.00	0.01	0.01		
8:24	0.01	0.03	0.05	0.01		
8:25	1.02	0.00	0.31	0.08		
8:26	2.09	0.00	0.32	0.08	Nox Bias;	
8:27	1.99	0.00	1.36	0.34		
8:28	1.59	0.00	11.51	3.75		
8:29	0.28	0.00	11.97	4.03	O2 Co2 Bias;	
8:30	0.10	0.00	11.91	4.02		
8:31	0.15	1.79	1.33	0.55		
8:32	0.06	3.84	0.00	0.01	CO Bias;	
8:33	0.06	3.43	0.00	0.00		
8:34	0.17	0.62	0.01	0.00		
8:35	0.18	0.02	0.07	0.02	NOX (Corr) (ppm)	CO (Corr) (ppm)
8:36	0.91	0.00	3.36	0.08	1.03	0.00
8:37	2.12	0.01	0.88	0.07	1.90	0.01
8:38	2.12	0.01	0.08	0.08	1.82	0.01
8:39	2.12	0.01	0.08	0.08	1.82	0.01
8:40	2.48	0.01	9.58	6.18	3.92	0.01
8:41	2.54	0.01	9.58	6.18	4.02	0.01
8:42	2.53	0.00	9.58	6.18	4.00	0.01
8:43	2.51	0.01	9.58	6.18	3.97	0.01
8:44	2.51	0.01	9.58	6.18	3.97	0.01
8:45	2.57	0.01	9.58	6.18	4.07	0.01
8:46	2.47	0.01	9.58	6.18	3.90	0.01
8:47	2.56	0.01	9.58	6.18	4.04	0.01
8:48	2.65	0.01	9.58	6.17	4.18	0.01
8:49	2.60	0.01	9.57	6.17	4.11	0.01
8:50	2.50	0.01	9.58	6.18	3.96	0.01
8:51	2.61	0.01	9.59	6.18	4.12	0.01
8:52	2.63	0.01	9.60	6.19	4.17	0.01
8:53	2.67	0.01	9.66	6.22	4.25	0.01
8:54	2.73	0.01	9.67	6.23	4.36	0.01
8:55	2.61	0.01	10.43	6.23	4.67	0.01
8:56	2.73	0.01	10.19	6.24	4.59	0.01
8:57	2.65	0.03	9.68	6.25	4.23	0.04
8:58	2.69	0.08	9.65	6.25	4.28	0.13
8:59	2.72	0.01	9.65	6.25	4.32	0.01
9:00	2.57	0.03	9.57	6.21	4.06	0.04
9:01	2.63	0.06	9.44	6.13	4.10	0.10
9:02	2.67	0.11	9.36	6.08	4.14	0.18
9:03	2.50	0.13	9.31	6.04	3.86	0.20
9:04	2.62	0.15	9.28	6.02	4.03	0.23

9:05	2.68	0.11	9.25	6.01	4.12	0.16
9:06	2.57	0.15	9.24	6.00	3.94	0.24
9:07	2.52	0.10	9.24	6.00	3.87	0.16
9:08	2.56	0.12	9.24	6.00	3.93	0.19
9:09	2.59	0.18	9.24	6.00	3.98	0.27
9:10	2.52	0.15	9.24	6.00	3.87	0.23
9:11	2.51	0.15	9.24	6.00		
9:12	0.83	0.11	9.25	6.00		
9:13	0.29	0.01	0.03	0.04		
9:14	0.00	0.00	0.01	0.01	Bias Zero;	
9:15	0.00	0.00	0.02	0.02		
9:16	0.01	0.03	0.05	0.03		
9:17	1.02	0.00	0.31	0.08		
9:18	2.09	0.00	0.32	0.08	Nox Bias;	
9:19	1.99	0.00	1.36	0.34		
9:20	1.59	0.00	11.51	3.75		
9:21	0.28	0.00	11.97	4.03	O2 Co2 Bias;	
9:22	0.10	0.00	11.91	4.02		
9:23	0.15	1.79	1.33	0.55		
9:24	0.06	3.89	0.00	0.01	CO Bais;	
9:25	0.01	3.48	0.00	0.00		
9:26	0.01	3.43	0.00	0.00		
9:27	0.01	0.00	0.00	0.01	Zero In;	
9:28	0.26	0.00	0.00	0.01		
9:29	1.57	0.00	0.01	0.00		
9:30	2.08	0.01	0.01	0.00	Nox In;	
9:31	2.02	0.01	0.41	0.12		
9:32	1.39	0.01	10.37	3.63		
9:33	0.39	0.01	11.97	4.03	O2 CO2 In;	
9:34	0.16	0.23	12.00	3.99		
9:35	0.01	0.72	5.42	1.92		
9:36	0.00	3.63	0.06	0.03		
9:37	0.00	3.91	0.02	0.00	CO In;	
9:38	0.00	2.33	0.01	0.00		
9:39	1.60	0.76	6.83	5.96	NOX (Corr) (ppm)	CO (Corr) (ppm)
9:40	2.06	0.18	8.40	6.70	2.95	0.26
9:41	2.02	0.13	8.40	6.70	2.89	0.19
9:42	2.06	0.07	8.40	6.70	2.94	0.11
9:43	2.06	0.15	8.40	6.70	2.95	0.22
9:44	2.01	0.14	8.40	6.70	2.88	0.19
9:45	2.13	0.12	8.40	6.70	3.05	0.18
9:46	2.13	0.11	8.40	6.70	3.05	0.15
9:47	2.08	0.11	8.40	6.70	2.98	0.16
9:48	2.06	0.13	8.40	6.70	2.95	0.18
9:49	2.06	0.13	8.40	6.70	2.95	0.18
9:50	2.13	0.12	8.40	6.70	3.05	0.17
9:51	2.14	0.12	8.40	6.70	3.07	0.17
9:52	2.03	0.12	8.40	6.70	2.91	0.17
9:53	2.04	0.11	8.40	6.70	2.92	0.16
9:54	2.13	0.13	8.40	6.70	3.05	0.18
9:55	1.99	0.08	8.40	6.70	2.85	0.11

9:56	2.10	0.14	8.97	6.70	3.32	0.22	PC;
9:57	1.99	0.07	10.75	6.71	4.42	0.19	
9:58	2.02	0.06	8.40	6.70	2.89	0.09	
9:59	2.05	0.07	8.39	6.70	2.93	0.10	
10:00	2.16	0.06	8.39	6.70	3.09	0.09	
10:01	2.06	0.12	8.39	6.70	2.95	0.16	
10:02	2.14	0.15	8.39	6.70	3.07	0.21	
10:03	2.15	0.16	8.39	6.70	3.07	0.23	
10:04	2.12	0.08	8.39	6.70	3.04	0.12	
10:05	2.08	0.10	8.39	6.70	2.98	0.15	
10:06	2.15	0.13	8.39	6.70	3.08	0.18	
10:07	2.23	0.11	8.39	6.70	3.19	0.15	
10:08	2.23	0.14	8.39	6.70	3.19	0.19	
10:09	2.17	0.13	8.39	6.70	3.11	0.19	
10:10	2.23	0.12	8.39	6.70	3.20	0.17	
10:11	2.15	0.12	8.39	6.70	3.07	0.18	
10:12	0.04	0.00	0.05	0.01			
10:13	0.00	0.00	0.00	0.01	Bias Zero;		
10:14	0.26	0.00	0.00	0.01			
10:15	1.57	0.01	0.10	0.03			
10:16	2.12	0.01	0.25	0.09	Bias Nox;		
10:17	1.12	0.01	0.41	0.12			
10:18	1.39	0.01	10.37	3.63			
10:19	0.39	0.01	11.97	4.03	O2 CO2 Bias;		
10:20	0.16	0.01	12.00	3.99			
10:21	0.04	0.72	5.42	1.92			
10:22	0.02	3.63	0.08	0.03			
10:23	0.01	3.91	0.01	0.01	CO Bias;		
10:24	0.00	2.33	0.01	0.01			
10:25	0.00	0.48	0.01	0.01			
10:26	0.00	0.00	0.00	0.01	All Zero;		
10:27	0.07	0.01	0.00	0.01			
10:28	1.29	0.01	0.01	0.01			
10:29	2.10	0.01	0.21	0.09	Low Nox;		
10:30	1.95	0.01	0.18	0.07			
10:31	2.12	0.01	0.19	0.07			
10:32	3.10	0.01	0.07	0.02			
10:33	4.02	0.01	0.01	0.01	High Nox;		
10:34	3.87	0.00	0.00	0.01			
10:35	3.65	0.00	7.14	2.25			
10:36	1.14	0.00	11.89	3.90			
10:37	0.23	0.00	11.98	3.99	Low O2 Co2;		
10:38	0.10	0.00	12.38	4.08			
10:39	0.01	0.00	19.79	7.37			
10:40	0.01	0.02	20.13	7.91	High O2 Co2;		
10:41	0.00	1.03	4.16	2.15			
10:42	0.00	3.57	0.06	0.20			
10:43	0.00	3.92	0.02	0.16	Low CO;		
10:44	0.00	3.94	0.04	0.16			
10:45	0.01	5.91	0.06	0.15			
10:46	0.01	7.89	0.03	0.15			

10:47

0.01

8.09

0.03

0.15


High CO;

AIRX TESTING SERVICES Inc.

STACK DIAGRAM

Client: ALL AMERICAN
 Site: IRVINE
 Unit: OIL HEATER


Date: 03.23.21
 Client #: 1064
 Lab #: 221.028



"A" Side: 20 (in)=Dia: 1.0
 "B" Side: 42 (in)=Dia: 2.0
 Diameter: 20 inches
 Port Size: 1 inches
 Port Offset: 2 inches
 Stack Area: _____ sq feet
 Port Type: M

Total Traverse Points: _____

Points per Port	Distance inches	w/Offset inches
1	0.64	2.64
2	2.10	4.10
3	3.80	5.80
4	6.46	8.46
5	13.54	15.54
6	16.12	18.12
7	17.90	19.90
8	19.36	21.36



AIRx Testing NOx - API 200EH 1/5/2021
Semi- Annual Linearity @ 2.5 & 5 ppm Ranges

Range = 2.5

Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
1.1	1.10	0.0 %	DT0013694
2.1	2.10	0.0 %	DT0027640

Range = 5

Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
1.1	1.1	0.0 %	DT0013694
2.1	2.095	-0.1 %	DT0027640
4.02	3.985	-0.7 %	DT009817

40.0 NOX DT0028765

20.2 NOX EC724456

12.0 NOX EC106804

8.12 NOX CC122225

NOX 50 scale

20.2 NOX CC724456

12.0 NOX CC106804

8.12 NOX CC122225

4.02 NOX DT001817

NOX 25 scale

8.12 NOX
CC122225

4.02 NOX DT001817

2.1 NOX DT0027640

2.1 NOX
DT0023694

NOX 10 scale

4.02 NOX DT001817

2.1 NOX DT0027640

1.1 NOX
DT0013694

NOX 5 scale

2.1 NOX
DT0027640

1.1 NOX DT0013694

NOX 2.5 scale

AIRx Testing O2 - Servomex 1400 1/5/2021
Semi- Annual Linearity @ 10 & 25 % Ranges

Range =

10

Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
3.97	3.96	0.1 %	CC87396
7.83	7.83	0.0 %	CC77079

Range =

25

Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
3.97	3.95	-0.1 %	CC87396
7.83	7.83	0.0 %	CC77079
11.97	11.98	0.0 %	CC134279
20.04	20.05	0.0 %	CC150711

20.04 02 CC150711

11.97 02 CC134279

7.93 02 CC77079

3.97 02 CC87396

02 25 SCALE

3.97 02 CC87396

7.93 02
CC77079

02 10 SCALE

AIRx Testing CO - API 300EM 1/5/2021
 Semi- Annual Linearity @ 10,25,50,100 & 200 ppm Ranges

Range =	10		
Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
3.89	3.89	0.00 %	CC28151
8.02	8.02	0.00 %	DT0020113

Range =	25		
Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
3.89	3.93	0.14 %	CC28151
8.02	8.10	0.32 %	DT0020113
12.11	12.15	0.16 %	CC7493
20.5	20.48	-0.10 %	DT0012296

Range =	50		
Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
8.02	8.15	0.26 %	DT0020113
12.11	12.15	0.08 %	CC7493
20.5	20.55	0.10 %	DT0012296
40.2	40.15	-0.10 %	CC272461

Range =	100		
Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
12.11	12.20	0.09 %	CC7493
20.5	20.6	0.10 %	DT0012296
40.2	40.3	0.10 %	CC272461
80.3	80.3	0.00 %	SA18996

Range =	200		
Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
20.5	20.60	0.05 %	DT0012296
40.2	40.20	0.00 %	CC272461
80.3	80.80	0.25 %	SA18996
181	181	0.00 %	CC243390

181.00
CC243390

80.3 00 SA189910

40.2 00 CC272461

20.5 00 DT0012296

CO 200 scale

80.3 00 SA189910

40.2 00 CC272461

20.5 00 DT0012296

12.11 00
CC74993

CO 100 SCALE

40.2 00 CC272461

20.5 00 DT0012296

12.11 00 CC7493

8.02 00
DT0020113

CO 50 scale

20.5 00 DT0012296

12.11 00 CC7493

8.02 00 DT0020113

3.89 00 CC28151

CO 25 scale

8.02 00 DT0020113

3.89 00 CC28151

CO 10 scale

AIRx Testing CO2 - Servomex 1440 1/5/2021
Semi- Annual Linearity @ 10 & 20 % Ranges

Range =

10

Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
4.01	4.00	0.1 %	CC134279
7.99	8.00	0.1 %	CC87396

Range =

20

Actual Concentration	Avg. Analyzer Response	Error % of Range	Cylinder #
4.01	4.04	0.2 %	CC134279
7.99	8	0.0 %	CC87396
17.51	17.48	-0.2 %	CC77079

10 SCALE CO₂

4.01 CO₂ CC134279

7.99 CO₂ CC87396

20 SCALE CO₂

7.99 CO₂
CC87396

17.51 CO₂

ACT11079



Praxair Distribution, Inc.
5700 S. Alameda Street
Los Angeles, CA 90058
Tel: 323-585-2154
Fax: 714-542-6689

Customer & Order Information:

PRAXAIR PKG OXNARD CA HPS
455 E WOOLEY RD,
OXNARD, CA 93030-7224
Praxair Order Number: 71263547
Customer PO Number: 79252704

Certificate Issuance Date: 3/4/2020

Certification Date: 4/4/2020
Lot Number: 70086005802
Part Number: NI NO2.2MM-AS
DocNumber: 196523
Expiration Date: 4/3/2021

CERTIFICATE OF ANALYSIS
Certified Master

Component	Requested Concentration (Molar)	Certified Concentration (Molar)	Analytical Reference	Analytical Uncertainty
Nitric oxide	2.2 ppm	2.1 ppm	1	± 0.2 ppm Abs.
Nitrogen	Balance	Balance		

Cylinder Style: AS
Cylinder Pressure @ 70 F: 2000 psig
Cylinder Volume: 143 ft³
Valve Outlet Connection: CGA 660
Cylinder Number(s): DT0027640
For Reference Only: NOx 2.1 ppm

Analyst: Henry Koung

Approved Signer: Leeanna Flores

Key to Analytical Techniques:

Reference	Analytical Instrument - Analytical Principle	Reference Standard
1	Thermo Electron 42i-LS S/N 1030645077 - Chemiluminescence	NO/N2 4.95 ppm# DT0026850, Expiration Date: 09/26/2021, Traceable to PRM#APEX1236603

The gas calibration cylinder standard prepared by Praxair Distribution, Inc. is considered a certified standard. It is prepared by gravimetric, volumetric, or partial pressure techniques. The calibration standard provided is certified against Praxair Distribution, Inc. Reference Materials which are traceable to the International System of Units (SI) through either weights traceable to the National Institute of Standards and Technology (NIST) or Measurement Canada, or through NIST Standard Reference Materials or equivalent where available.

Note: All expressions for concentration (e.g., % or ppm) are for gas phase, by volume (e.g., ppmv) unless otherwise noted. Analytical uncertainty is expressed as a Relative % unless otherwise noted.

IMPORTANT
The information contained herein has been prepared at your request by personnel within Praxair Distribution, Inc. While we believe the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall liability of Praxair Distribution, Inc. arising out of the use of the information contained herein exceed the fee established for providing such information.



Making our planet more productive

DocNumber: 297346



Praxair Distribution, Inc.
5700 S. Alameda Street
Los Angeles CA 90058
Tel: 323-585-2154
Fax: 714-542-6589
PGVP ID: F22020

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

Customer & Order Information

PRAXAIR PKG OXNARD CA HPS
455 E WOOLEY RD
OXNARD CA 93030-7224

Certificate Issuance Date: 01/18/2020
Praxair Order Number: 71200500
Part Number: NI NO4ME-AS
Customer PO Number: 78188373

Fill Date: 01/03/2020
Lot Number: 70086000305
Cylinder Style & Outlet: AS CGA 660
Cylinder Pressure and Volume: 2000 psig 140 ft³

Certified Concentration

Expiration Date:	01/18/2023	NIST Traceable
Cylinder Number:	DT0035144	Expanded Uncertainty
4.01 ppm	Nitric oxide	± 0.6 %
Balance	Nitrogen	

ProSpec EZ Cert



For Reference Only: NOx 4.04 ppm

Certification Information: Certification Date: 01/18/2020 Term: 36 Months Expiration Date: 01/18/2023

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.
Do Not Use this Standard If Pressure is less than 100 PSIG.

Analytical Data:

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: Nitric oxide

Requested Concentration: 4 ppm
Certified Concentration: 4.01 ppm
Instrument Used: Thermo Electron 42i-LS S/N 1030645077
Analytical Method: Chemiluminescence
Last Multipoint Calibration: 01/13/2020

Reference Standard: Type / Cylinder #: GMIS / CC499176
Concentration / Uncertainty: 5.06 ppm ± 0.51%
Expiration Date: 09/26/2021
Traceable to: SRM # / Sample # / Cylinder #: PRM/RAPEX1235603 / n/a /
SRM Concentration / Uncertainty: 10.00 ppm / ± 0.05 ppm
SRM Expiration Date: 09/03/2020

First Analysis Data:				Date
Z: 0	R: 5.06	C: 4.01	Cone: 4.01	01/11/2020
R: 5.05	Z: 0	C: 4.02	Cone: 4.02	
Z: 0	C: 4.02	R: 5.05	Cone: 4.02	
UOM: ppm				Mean Test Assay: 4.02 ppm

Second Analysis Data:				Date
Z: 0	R: 5.05	C: 4	Cone: 4	01/18/2020
R: 5.06	Z: 0	C: 4	Cone: 4	
Z: 0	C: 4	R: 5.05	Cone: 4	
UOM: ppm				Mean Test Assay: 4 ppm

Analyzed By

Henry Koung

Certified By

Leeanna Flores

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use



Making our planet more productive

DocNumber: 253680



Praxair Distribution, Inc.
5700 S. Alameda Street
Los Angeles CA 90058
Tel: 323-585-2154
Fax: 714-542-6689
PGVP ID: F22019

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

Customer & Order Information

PRAXAIR PKG OXNARD CA HPS
455 E WOOLEY RD
OXNARD CA 93030

Certificate Issuance Date: 06/27/2019

Praxair Order Number: 70998286

Part Number: NI NO20ME-AS

Customer PO Number: 78981976

Fill Date: 06/07/2019

Lot Number: 70086915805

Cylinder Style & Outlet: AS CGA 660

Cylinder Pressure and Volume: 2000 psig 140 ft³

Certified Concentration		
Expiration Date:	06/27/2022	NIST Traceable
Cylinder Number:	CC724456	Expanded Uncertainty
20.1 ppm	Nitric oxide	± 0.5 %
Balance	Nitrogen	

ProSpec EZ Cert



For Reference Only: NOx 20.2 ppm

Certification Information: Certification Date: 06/27/2019 Term: 36 Months Expiration Date: 06/27/2022

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.
Do Not Use this Standard if Pressure is less than 100 PSIG.

Analytical Data: (R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component:	Nitric oxide	Reference Standard:	Type / Cylinder #: GMIS / CC2919
Requested Concentration:	20 ppm	Concentration / Uncertainty:	19.80 ppm ± 0.504%
Certified Concentration:	20.1 ppm	Expiration Date:	03/26/2022
Instrument Used:	Thermo Electron 42i-LS S/N 1030845077	Traceable to:	SRM # / Sample # / Cylinder #: APEX1161149 / NA / APEX1161149
Analytical Method:	Chemiluminescence	SRM Concentration / Uncertainty:	20.03 ppm / ±0.10 ppm
Last Multipoint Calibration:	06/10/2019	SRM Expiration Date:	01/27/2020
First Analyze Data:	Date 06/20/2019	Second Analyze Data:	Date 06/27/2019
Z: 0 R: 19.8 C: 20 Conc: 20	Z: 0 R: 19.8 C: 20.1 Conc: 20.1	Z: 0 R: 19.89 C: 20.1 Conc: 20.1	Z: 0 R: 19.89 C: 20.2 Conc: 20.2
R: 19.88 Z: 0 C: 20.1 Conc: 20.1	Z: 0 R: 19.91 C: 20.1 Conc: 20.1	Z: 0 R: 19.89 C: 20.1 Conc: 20.1	Z: 0 R: 19.89 C: 20.2 Conc: 20.2
Z: 0 C: 20.1 R: 19.91 Conc: 20.1	Mean Test Assay: 20.1 ppm	UOM: ppm	Mean Test Assay: 20.1 ppm

Analyzed By

Henry Koung

Certified By

Leeanna Flores

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc. arising out of the use of the information contained herein exceed the fee established for providing such information.



Praxair Distribution, Inc.
5700 S. Alameda Street
Los Angeles, CA 90058
Tel: 323-585-2154
Fax: 714-542-6689

Customer & Order Information:

PRAIR PKG OXNARD CA HPS
455 E WOOLEY RD,
OXNARD, CA 93030-7224
Praxair Order Number: 71378203
Customer PO Number: 79369772

Certificate Issuance Date: 7/10/2020

Certification Date: 7/10/2020
Lot Number: 70086018201
Part Number: NI NX20MP-AS
DocNumber: 232651
Expiration Date: 7/10/2021

CERTIFICATE OF ANALYSIS
Primary Standard

Component	Requested Concentration (Molar)	Certified Concentration (Molar)	Analytical Reference	Analytical Uncertainty
Nitrogen dioxide (as NOx)	20 ppm	19.01 ppm	1	± 1%
Nitrogen	Balance	Balance		

Cylinder Style: AS
Cylinder Pressure @ 70 F: 2000 psig
Cylinder Volume: 142 ft³
Valve Outlet Connection: CGA 660
Cylinder Number(s): CC3240

Comments: This mixture contains 2% Oxygen.

Analyst: Leeanna Flores

Approved Signer: Henry Koung

Key to Analytical Techniques:

Reference Analytical Instrument - Analytical Principle
1 MKS MG2031 - FTIR

The gas calibration cylinder standard prepared by Praxair Distribution, Inc. is considered a certified standard. It is prepared by gravimetric, volumetric, or partial pressure techniques. The calibration standard provided is certified against Praxair Distribution, Inc. Reference Materials which are traceable to the International System of Units (SI) through either weights traceable to the National Institute of Standards and Technology (NIST) or Measurement Canada, or through NIST Standard Reference Materials or equivalent where available.

Note: All expressions for concentration (e.g., % or ppm) are for gas phase, by volume (e.g., ppmv) unless otherwise noted. Analytical uncertainty is expressed as a Relative % unless otherwise noted.

IMPORTANT

The information contained herein has been prepared at your request by personnel within Praxair Distribution, Inc.. While we believe the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any particular purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall liability of Praxair Distribution, Inc. arise out of the use of the information contained herein exceed the fee established for providing such information.



Making our planet more productive

DocNumber: 229037



Praxair Distribution, Inc.
5700 S. Alameda Street
Los Angeles CA 90058
Tel: 323-585-2154
Fax: 714-542-6689
PGVP ID: F22018

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

Customer & Order Information

PRAXAIR PKG OXNARD CA HPS
455 E WOOLEY RD
OXNARD CA 93030

Certificate Modification Date: 11/15/2018
Praxair Order Number: 70784180
Part Number: EV NICDOXE90-AS

Full Date: 11/09/2018
Lot Number: 70086831305
Cylinder Style & Outlet: AS CGA 580
Cylinder Pressure and Volume: 2000 psig 140 ft³

Certified Concentration		
Expiration Date:	11/15/2026	NIST Traceable
Cylinder Number:	CC134279	Expanded Uncertainty
4.01 %	Carbon dioxide	± 0.4 %
11.97 %	Oxygen	± 0.1 %
Balance	Nitrogen	

ProSpec EZ Cert



Certification Information:

Certification Date: 11/15/2018

Term: 96 Months

Expiration Date: 11/15/2026

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.
Do Not Use this Standard if Pressure is less than 100 PSIG.

CO2 responses have been corrected for Oxygen IR Broadening effect. O2 responses have been corrected for CO2 interference.

Analytical Data:

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: Carbon dioxide

Requested Concentration: 4 %
Certified Concentration: 4.01 %
Instrument Used: Horiba VIA-510 S/N 20C194WK
Analytical Method: NDIR
Last Multipoint Calibration: 10/23/2018

Reference Standard: Type / Cylinder #: GMIS / CC243762
Concentration / Uncertainty: 6.88 % ± 0.208%
Expiration Date: 06/07/2023
Traceable to: SRM # / Sample # / Cylinder #: SRM 1674b / 7-H-07 / FF10631
SRM Concentration / Uncertainty: 6.944% / ± 0.013%
SRM Expiration Date: 06/17/2019

First Analysis Data:				Date
Z: 0	R: 6.98	C: 4	Conc: 4	11/15/2018
R: 6.98	Z: 0	C: 4.01	Conc: 4.01	
Z: 0	C: 4.01	R: 6.98	Conc: 4.01	
UOM: %			Mean Test Assay: 4.01 %	

Second Analysis Data:				Date
Z: 0	R: 0	C: 0	Conc: 0	
R: 0	Z: 0	C: 0	Conc: 0	
Z: 0	C: 0	R: 0	Conc: 0	
UOM: %			Mean Test Assay: %	

2. Component: Oxygen

Requested Concentration: 12 %
Certified Concentration: 11.97 %
Instrument Used: OXYMAT 5E
Analytical Method: Paramagnetic
Last Multipoint Calibration: 11/05/2018

Reference Standard: Type / Cylinder #: GMIS / CC75674
Concentration / Uncertainty: 20.88 % ± 0.111%
Expiration Date: 11/07/2025
Traceable to: SRM # / Sample # / Cylinder #: SRM 2659a / 71-E-19 / FF2231
SRM Concentration / Uncertainty: 20.863% / ± 0.021%
SRM Expiration Date: 08/23/2021

First Analysis Data:				Date
Z: 0	R: 20.88	C: 11.88	Conc: 11.88	11/15/2018
R: 20.88	Z: 0	C: 11.97	Conc: 11.97	
Z: 0	C: 11.97	R: 20.88	Conc: 11.97	
UOM: %			Mean Test Assay: 11.97 %	

Second Analysis Data:				Date
Z: 0	R: 0	C: 0	Conc: 0	
R: 0	Z: 0	C: 0	Conc: 0	
Z: 0	C: 0	R: 0	Conc: 0	
UOM: %			Mean Test Assay: %	

Analyzed By

Danielle Burns

Certified By

Jose Vasquez

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.



DocNumber: 000081223

Praxair
5700 South Alameda Street
Los Angeles, CA 90058
Tel: (323) 585-2154 Fax: (714) 542-6689
PGVID: F22015

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS**Customer & Order Information:**

PRAXAIR WHSE OXNARD CA
455 E WOOLEY RD
OXNARD CA 930300

Praxair Order Number: 31577381
Customer P. O. Number: 05587085
Customer Reference Number:

Fill Date: 6/18/2016
Part Number: EV NICDOXE98-AS
Lot Number: 109516809
Cylinder Style & Outlet: AS CGA 690
Cylinder Pressure & Volume: 2000 psig 140 cu. ft.

Certified Concentration:

Expiration Date:	6/23/2023 <th>NIST Traceable</th>	NIST Traceable
Cylinder Number:	CC144381	Analytical Uncertainty:
7.93	%	CARBON DIOXIDE
20.13	%	OXYGEN
Balance		NITROGEN

Certification Information: Certification Date: 6/23/2015 Term: 96 Months Expiration Date: 6/23/2023

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

O2 responses have been corrected for CO2 interference.

Analytical Data: (R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: CARBON DIOXIDE

Requested Concentration: 8 %
Certified Concentration: 7.93 %
Instrument Used: Horiba VIA-510 S/N 2807014
Analytical Method: NDIR
Last Multipoint Calibration: 6/15/2015

Reference Standard Type: GMIS
Ref. Std. Cylinder #: SA17G95
Ref. Std. Conc: 8.87%
Ref. Std. Traceable to SRM #: 1674b
SRM Sample #: 7-H-07
SRM Cylinder #: FF10631

First Analyze Date:	Date:
Z: 0 R: 8.87	C: 7.93 Conc: 7.927
R: 8.87 Z: 0	C: 7.93 Conc: 7.927
Z: 0 C: 7.93 R: 8.88	Conc: 7.927
UOM: %	Mean Test Assay: 7.927 %

Second Analyze Date:	Date:
Z: 0 R: 0	C: 0 Conc: 0
R: 0 Z: 0	C: 0 Conc: 0
Z: 0 C: 0 R: 0	Conc: 0
UOM: %	Mean Test Assay: 0 %

2. Component: OXYGEN

Requested Concentration: 20 %
Certified Concentration: 20.13 %
Instrument Used: OXYMAT SE
Analytical Method: PARAMAGNETIC
Last Multipoint Calibration: 5/26/2015

Reference Standard Type: GMIS
Ref. Std. Cylinder #: SA16022
Ref. Std. Conc: 19.90%
Ref. Std. Traceable to SRM #: 2559a
SRM Sample #: 71-E-18
SRM Cylinder #: FF22331

First Analyze Date:	Date:
Z: 0 R: 19.88	C: 20.12 Conc: 20.127
R: 19.88 Z: 0	C: 20.12 Conc: 20.127
Z: 0 C: 20.12 R: 19.88	Conc: 20.127
UOM: %	Mean Test Assay: 20.127 %

Second Analyze Date:	Date:
Z: 0 R: 0	C: 0 Conc: 0
R: 0 Z: 0	C: 0 Conc: 0
Z: 0 C: 0 R: 0	Conc: 0
UOM: %	Mean Test Assay: 0 %

Analyzed by:

Certified by:

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc. arising out of the use of the information contained herein exceed the fee established for providing such information.



Making our planet more productive

DocNumber: 253974



Praxair Distribution, Inc.
5700 S. Alameda Street
Los Angeles CA 90058
Tel: 323-585-2154
Fax: 714-542-6689
PGVP ID: F22019

CERTIFICATE OF ANALYSIS/EPA PROTOCOL GAS

Customer & Order Information

PRAXAIR PKG OXNARD CA HPS
455 E WOOLEY RD
OXNARD CA 93030

Certificate Issuance Date: 06/27/2019

Praxair Order Number: 70998288

Part Number: NI CO4ME-AS

Customer PO Number: 78981978

Fill Date: 06/12/2019

Lot Number: 70086916303

Cylinder Style & Outlet: AS

CGA 350

Cylinder Pressure and Volume: 2000 psig 140 ft³

Certified Concentration

Expiration Date:	06/21/2027	NIST Traceable
Cylinder Number:	CC28151	Expanded Uncertainty
3.89 ppm	Carbon monoxide	± 1.0 %
Balance	Nitrogen	

ProSpec EZ Cert



Certification Information:

Certification Date: 06/21/2019

Term: 96 Months

Expiration Date: 06/21/2027

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1.
Do Not Use this Standard if Pressure is less than 100 PSIG.

Analytical Data:

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: Carbon monoxide

Requested Concentration: 4 ppm

Certified Concentration: 3.89 ppm

Instrument Used: MKS Multigas 2031 FTIR

Analytical Method: FTIR

Last Multipoint Calibration: 06/18/2019

Reference Standard: Type / Cylinder #: GMIS / SA13564

Concentration / Uncertainty: 5.055 ppm ± 0.553%

Expiration Date: 09/05/2026

Traceable to: SRM # / Sample # / Cylinder #: SRM 1677c / 5-J-42 / CAL015337

SRM Concentration / Uncertainty: 9.825 PPM / ± 0.047 PPM

SRM Expiration Date: 06/24/2024

First Analysis Data:				Date
Z: 0	R: 6.056	C: 3.889	Conc: 3.88	06/21/2019
R: 6.053	Z: 0	C: 3.911	Conc: 3.9	
Z: 0	C: 3.893	R: 5.087	Conc: 3.89	
UOM: ppm		Mean Test Assay:	3.89 ppm	

Second Analysis Data:				Date
Z: 0	R: 0	C: 0	Conc: 0	
R: 0	Z: 0	C: 0	Conc: 0	
Z: 0	C: 0	R: 0	Conc: 0	
UOM: ppm		Mean Test Assay:	ppm	

Analyzed By

Brauldon Aguilar

Certified By

José Vasquez

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

DocNumber: 000087119

Praxair
 5700 South Alameda Street
 Los Angeles, CA 90058
 Tel: (323) 585-2154 Fax: (714) 542-6689
 PGVP ID: F22015

CERTIFICATE OF ANALYSIS/EPA PROTOCOL GAS**Customer & Order Information:**

PRAXAIR WHSE OXNARD CA
 455 E WOOLEY RD
 OXNARD CA 93030

Praxair Order Number: 33101951
 Customer P. O. Number: 05830311
 Customer Reference Number:

FHI Date: 11/21/2015
 Part Number: NI CÓ8ME-AS
 Lot Number: 109532500
 Cylinder Style & Outlet: AS CGA 350
 Cylinder Pressure & Volume: 2000 psig 140 cu. ft.

Certified Concentration:

Expiration Date:	11/30/2023	NIST Traceable
Cylinder Number:	CC196750	Analytical Uncertainty:
8.12 ppm CARBON MONOXIDE		± 0.8 %
Balance NITROGEN		

Certification Information: Certification Date: 11/30/2015 Term: 96 Months Expiration Date: 11/30/2023

This cylinder was certified according to the 2012 EPA Traceability Protocol, Document #EPA-600/R-12/531, using Procedure G1. Do Not Use this Standard if Pressure is less than 100 PSIG.

Analytical Data: (R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: CARBON MONOXIDE

Requested Concentration: 8 ppm
 Certified Concentration: 8.12 ppm
 Instrument Used: Horiba VIA-510 S/N 576876015
 Analytical Method: NDIR
 Last Multipoint Calibration: 11/13/2015

Reference Standard Type: GMIS
 Ref. Std. Cylinder #: CC130322
 Ref. Std. Conc: 10.15 ppm
 Ref. Std. Traceable to SRM #: 1677c
 SRM Sample #: 5-J-42
 SRM Cylinder #: CAL015337

First Analysis Data:				Date:
Z: 0	R: 84	C: 67.4	Conc: 8.134	
R: 84.2	Z: 0	C: 67.2	Conc: 8.11	
Z: 0	C: 67.2	R: 84.1	Conc: 8.11	
UOM: ppm		Mean Test Assay:	8.118 ppm	

Second Analysis Data:				Date:
Z: 0	R: 0	C: 0	Conc: 0	0
R: 0	Z: 0	C: 0	Conc: 0	0
Z: 0	C: 0	R: 0	Conc: 0	0
UOM: ppm		Mean Test Assay:	0 ppm	

Analyzed by:

Certified by:

Information contained herein has been prepared at your request by qualified experts within Praxair Distribution, Inc. While we believe that the information is accurate within the limits of the analytical methods employed and is complete to the extent of the specific analyses performed, we make no warranty or representation as to the suitability of the use of the information for any purpose. The information is offered with the understanding that any use of the information is at the sole discretion and risk of the user. In no event shall the liability of Praxair Distribution, Inc., arising out of the use of the information contained herein exceed the fee established for providing such information.

SOURCE EMISSION INSTRUMENTATION LIST

CARBON DIOXIDE

Unit No. - 3: Manufacturer: ACS (Fugi) Model No.: 3300 Serial No.: N8M6611T Method: NDIR Range (%) 0-5 & 20	Unit No. - 8: Manufacturer: Servomex Model No.: 1400 Serial No.: X1415/B202 Method: NDIR Range (%) 0-10
Unit No. - 4: Manufacturer: ACS (Fugi) Model No.: 3300 Serial No.: N9C5479T Method: NDIR Range (%) 0-5 & 20	
Unit No. - 5: Manufacturer: Milton-Roy (Fugi) Model No.: 3300 Serial No.: N2EO363T Method: NDIR Range (%) 10/20/25	
Unit No. - 6: Manufacturer: Horiba Model No.: 3400 Serial No.: N1P7019T Method: NDIR Range (%) 0- 20	
Unit No. - 7: Manufacturer: Servomex Model No.: 1400 Serial No.: 01415/B7-103 Method: NDIR Range (%) 0-10	



SOURCE EMISSION INSTRUMENTATION LIST

CARBON MONOXIDE

<u>Unit No. - 1:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	API 300 EM 239 NDIR/GFC 0-5, 10, 25, 50, 100, 250, 500, 1000, 2500, 5000	<u>Unit No. - 2:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	API 300 EM 240 NDIR/GFC 0-5, 10, 25, 50, 100, 250, 500, 1000, 2500, 5000
<u>Unit No. - 3:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	API 300 EM 104 NDIR/GFC 0-5, 10, 25, 50, 100, 250, 500, 1000, 2500, 5000	<u>Unit No. - 5:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	Thermo Environmental (TECO) 48H 25184-219 NDIR/GFC 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000
<u>Unit No. - 6:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	Thermo Environmental (TECO) 48H 29031-233 NDIR/GFC 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000	<u>Unit No. - 7:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	Siemens Ultramat 21p AO4-254 NDIR 0-300
<u>Unit No. - 10:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	Thermo Environmental (TECO) 48H 38391-257 NDIR/GFC 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000	<u>Unit No. - 11:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	Thermo Environmental (TECO) 48H 35226-249 NDIR/GFC 0-50, 100, 200, 500, 1000, 2000, 5000, 10000, 20000, 50000
<u>Unit No. - 13:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	Thermo Environmental (TECO) 48 48-15970-159 NDIR/GFC 0-1, 2, 5, 10, 20, 50, 100, 200, 500, 1000	<u>Unit No. - 14:</u> Manufacturer: Model No.: Serial No.: Method: Range (ppmv)	Thermo Environmental (TECO) 48 48-23925-213 NDIR/GFC 0-1, 2, 5, 10, 20, 50, 100, 200, 500, 1000

SOURCE EMISSION INSTRUMENTATION LIST

OXIDES OF NITROGEN

<u>Unit No. - 1:</u> Manufacturer: API Model No.: 200 EH Serial No.: 233 Method: Chemiluminescence Range (ppmv) 0-5000	<u>Unit No. - 6:</u> Manufacturer: API Model No.: 200 A Serial No.: 1013 Method: Chemiluminescence Range (ppmv) 0-50
<u>Unit No. - 2:</u> Manufacturer: API Model No.: 200 EH Serial No.: 234 Method: Chemiluminescence Range (ppmv) 0-5000	<u>Unit No. - 7:</u> Manufacturer: Thermo Environmental (TECO) Model No.: 10AR Serial No.: 25559-221 Method: Chemiluminescence Range (ppmv) 0-2.5, 10, 25, 100, 250, 1000, 2500, 10000
<u>Unit No. - 3:</u> Manufacturer: API Model No.: 200 EH Serial No.: 109 Method: Chemiluminescence Range (ppmv) 0-5000	<u>Unit No. - 8:</u> Manufacturer: Thermo Environmental (TECO) Model No.: 10AR Serial No.: 38586-258 Method: Chemiluminescence Range (ppmv) 0-2.5, 10, 25, 100, 250, 1000, 2500, 10000
<u>Unit No. - 4:</u> Manufacturer: API Model No.: 200 EH Serial No.: 442 Method: Chemiluminescence Range (ppmv) 0-5000	
<u>Unit No. - 5:</u> Manufacturer: API Model No.: 200 EH Serial No.: 441 Method: Chemiluminescence Range (ppmv) 0-5000	



SOURCE EMISSION INSTRUMENTATION LIST

OXYGEN

<u>Unit No. - 5:</u> Manufacturer: Teledyne Model No.: 320-AX Serial No.: 108743 Method: Electrochemical Range (%) 0-5, 10, 25	<u>Unit No. - 13:</u> Manufacturer: Servomex Model No.: 1400 Serial No.: X1420/B707 Method: Paramagnetic Range (%) 0-25
<u>Unit No. - 7:</u> Manufacturer: Teledyne Model No.: 320-AX Serial No.: 108742 Method: Electrochemical Range (%) 0-5, 10, 25	
<u>Unit No. - 9:</u> Manufacturer: Servomex Model No.: 1400 Serial No.: 01420/B701/730 Method: Paramagnetic Range (%) 0-25, 100	
<u>Unit No. - 10:</u> Manufacturer: Servomex Model No.: 1400 Serial No.: 01420/B308 Method: Paramagnetic Range (%) 0-25	
<u>Unit No. - 11:</u> Manufacturer: Teledyne Model No.: 320-A Serial No.: 111211 Method: Electrochemical Range (%) 0-5, 10, 25	
<u>Unit No. - 12:</u> Manufacturer: Servomex Model No.: 1400 Serial No.: 01420/B7103 Method: Paramagnetic Range (%) 0-25, 100	

AIR TESTING


Method 100 – Sample Train Assembly

Probe:

1. AIRx Testing Services, Inc. uses a stainless steel sampling probe.
2. The probe is connected to the sample conditioner using a heated (if necessary) Teflon sampling line.

Sample Conditioner (Condenser System):

1. The sample conditioner consists of a moisture knock-out bottle immersed in an ice bath.
2. All parts of the conditioner exposed to the sample are either glass, stainless steel, or Teflon.
3. The sample conditioner is setup so that the sample gas is not bubbled through the condensate.
4. A temperature gauge is used to determine the temperature of the condenser outlet.



Manifold Panel & Strip Chart Recorder:

1. The sample gas exiting the condenser is transported to the instrument manifold panel.
2. The sample manifold directs the sample gas through to the various analyzers.
3. All components of the system that contact the sample are either stainless steel, Teflon, or glass.
4. The data is continuously recorded on a Soltec 10" strip chart recorder with a 1% resolution.

Calibration System:


1. The system is composed of the Analyzer Calibration and the System BIAS Check.
2. Certified Calibration gases are supplied to the analyzers for calibration purposes.
3. The selection of zero or span calibration gases or sample gas is accomplished by operation of selector knobs located on the main flow control panel.

Nitrogen Oxides – Chemiluminescent Analyzer: Based on the chemiluminescent reaction of NO and ozone to form NO₂ in an excited state. Light emission is monitored through an optical filter by a high sensitivity photomultiplier tube, the output of which is electronically processed so it is linearly proportional to the NO concentration. The output is in units of ppmv.

Carbon Monoxide – Non-Dispersive Infrared (NDIR) Analyzer: Radiation from an infrared source is passed through a narrow band-pass filter and a multiple optical pass sample cell where absorption by the sample gas occurs. The infrared radiation exits the sample cell and falls on a solid state infrared detector. The output is in units of ppmv.

Oxygen – Electrochemical Analyzer: Oxygen in the flue gas sample diffuses through a Teflon membrane and is reduced on the surface of the cathode. A corresponding oxidation occurs at the anode and an electric current proportional to the concentration of oxygen is produced. The output is in units of percent O₂ by volume.

Carbon Dioxide – Non-Dispersive Infrared (NDIR) Analyzer: The instrument measures the differential in infrared energy absorbed from energy beams passed through a reference cell (containing a gas selected to have minimal absorption of infrared energy in the wavelength absorbed by CO₂) and a sample cell through which the sample gas flows continuously. The output is in units of percent CO₂ by volume.



SCAQMD Method 100.1

Procedures For Continuous Gaseous Emission Sampling

Principle: A sample of an exhaust gas stream is continuously extracted, conditioned, and conveyed to instrumental analyzers for the determination of:

SO₂ concentration using ultraviolet analyzer
NO_x concentration using chemiluminescent analyzer
O₂ concentration using electrochemical (fuel cell) type analyzer
CO concentration using non-dispersive infrared analyzer
CO₂ concentration using non-dispersive infrared analyzer

Applicability:

Stationary Source Gas Streams flowing in ducts, stacks, and flues

Alternative to SCAQMD reference methods 3.1, 4.1, 7.1, 10.1

Range:

The analytical range is selected so that the sample gas concentration for each run is between 20 and 95% of the range, for 95% of the test period.

Sensitivity:

The minimum detectable limit shall be less than 2% of the range (i.e. range is 20000 ppm, MDL shall be less than 400 ppm)

Measurement system: Sample interface, Gas analyzer, Data acquisition

Probe, Sample pump, Probe calibration system, Barometer

Sample line: Teflon (to moisture removal system, heated to prevent condensation)

Sample conditioning: reduce moisture content to below a dewpoint of 35°F.

Sample transport lines: Teflon (from moisture removal system to sample gas manifold).

Particulate filter: to prevent accumulation of particulate in the measurement system

Sample flow rate control: control valve + rotameter

Sample gas manifold: Divert a portion of the gas stream to the analyzer and the remainder to the bypass vent

Gas analyzer: SO₂, NO_x, O₂, CO₂, CO. An NO₂ to NO converter is not necessary if data is presented to demonstrate that the NO₂ portion of the exhaust gas is less than 5% of the total NO_x concentration.

Data recorder: strip chart recorder, analog computer, digital recorder. Resolution or readability should be 0.5% of range. Sampling measurements should be obtained at a minimum of 1-minute intervals.

Interference response sampling system: Introduce an interference test gas to the analyzer. The analyzer zero should be given a positive offset prior to the test to allow measurement of a negative interference.

Pitot tubes: Method 1.1 and 1.2

Differential Pressure gauges: Method 1.1 and 1.2

Sample gas moisture content equipment: Method 4.1

Vacuum gauge: leak checking

Thermocouple: temperature stack gas

Range: Upper limit of the gas concentration measurement range displayed on the data recorder

Calibration gas (CalGas): A gas of known concentration in an inert diluent gas

- High-Range: 80 to 100% of the range
- Mid-Range: 40 to 60% of the range
- Zero gas: impurity concentration < 0.25% of the range → Nitrogen

Analyzer calibration error: The difference between the known concentration of the CalGas and the gas concentration exhibited by the gas analyzer when the CalGas is introduced directly to the analyzer.
Performance Spec: less than 2% of the range for the zero, mid-range, and high-range CalGas.

Sampling system BIAS: The difference between the gas concentrations exhibited by the measurement system when CalGas is introduced at the sampling probe tip filter and when the same CalGas is introduced directly to the analyzer.
Performance Spec: less than 5% of the range for the zero, mid-range, and high-range CalGas.

Zero Drift: The difference in the measurement system responses at a zero concentration level during the initial calibration, and final calibration check after a test.
No adjustment to the measurement system is allowed at that point.
Performance Spec: less than 3% of the range over the period of each run.

Calibration Drift: The difference in the measurement system responses at a mid-range concentration level during the initial calibration and final calibration check after a test.
No adjustment to the measurement system is allowed at that point.
Performance Spec: less than 3% of the range over the period of each run.

Response time: the time required for the system to display 95% of a step change in gas concentration on the data recorder.

Interference response: The output response of the measurement system to a component in the sample gas, other than the gas component being measured.

Performance Spec: For example, an SO₂ analyzer should respond no more than 30 ppm when the NO₂ concentration in the sample gas is 2000 ppm.

- SO₂ analyzer: NO₂ → 2000/30, NH₃ → 1000000/0
- CO₂ analyzer: H₂O → 10000/1, CO → 15000/1, CH₄ → 20000/1
- CO analyzer: H₂O → 200000/1, CO₂ → 500000/1, SO₂, NO, NO₂, HC, H₂O → 1000000/1

Calibration Curve: A graph establishing the relationship between the analyzer response and the actual CalGas concentration introduced to the analyzer.

Linearity: Maximum deviation as a percent of range, between a mid range calibration reading and the reading predicted by a straight line drawn between high-range and zero gas calibration points.

Performance Spec: less than 1% of the range for the pretest and post test values.

Measurement System Performance Test Procedures:

- **Cleaning of sample train:** Flush probe, lines and conditioner with DW, then acetone. Dry with filtered dry air.
- **Allow Continuous Analyzers to warm up**
- **Sampling system preparation:** assemble sample train as shown in Fig 100.1-1, 100.1-2 and 100.1-3. Leak check the vacuum side of the assembly to a minimum of 20" of Hg (gauge). The sampling system should hold 20" of Hg vacuum for 5 minutes with less than 1" Hg loss. Check the pressure side of the system with liquid soap solution and correct any leaks.
- **Allowable modifications:** probe heating element can be eliminated if stack is at or below ambient temperature and condensation is not observed. Pitot tube can be eliminated if flow measurements are not required.
- **Calibrate analyzers and data recorders:** introduce CalGases directly to the instruments and make all necessary adjustments to calibrate the analyzer and data recorder. Adjust system components to achieve manufacturer's recommended sampling rates.

- **Analyzer calibration error check:** at the beginning and end of each test run
 1. Introduce zero, mid-range, high range CalGas
 2. Make no adjustments to the system except those necessary to achieve the correct flow rate.
 3. Record the analyzer responses to each CalGas on a form similar to 100.1-4
 4. If invalid calibration is exhibited (> 2% of the range), take corrective action and repeat check.
- **Instrument response time:** establish during semi-annual certification
- **Sampling system BIAS check: Mandatory**
 1. Backflush gas through the probe as necessary to prevent particulate buildup
 2. Zero, and either mid-range or high-range (whichever is closest to effluent concentration)
 3. Introduce upscale CalGas and record concentration on a form similar to 100.1-5
 4. Introduce zero CalGas and record concentration
 5. Make no adjustments to the system except those necessary to achieve the correct flow rate.
 6. If invalid calibration is exhibited (> 5% of the range), take corrective action and repeat check.
 7. If adjustments are required, repeat the analyzer calibration error check, then the sampling system BIAS check.
- **NO₂ to NO conversion check:** EPA Method 20 or gas mixture of NO₂ in air

Emission Test Procedure:

- Traverse stack to determine presence of stratification
- Single-point gas sampling is acceptable if gas composition is homogenous (<10% variation)
- Determine moisture content and velocity pressures or Mass flow rate may be obtained by stoichiometric and gas composition relations
- **Chart recorder label:** turn on strip chart recorder and label the chart as to pollutant, source, range, calibration cylinder ID number, certified expiration date, zero and upper range calibration settings, chart speeds, date, time and operator.
- **Sample probe traverse and minimum sampling time:**
 1. Insert probe in stack
 2. determine if single point sampling is acceptable
 3. If traverse is required, leave the probe at each traverse point for at least the system response time + 1 minute.
 4. Minimum sampling time of 60 minute is recommended. See District Rules and Regulations and permit conditions for special requirements.
 5. When test duration exceeds one hour, conduct zero and span checks every 2 hours. Adjust settings as necessary, mark strip charts and record in log books.
- **Zero and Calibration DRIFT Tests:**
 1. Immediately preceding and following each run, or if adjustments are necessary during the run, repeat the sampling system BIAS procedure. Make no adjustments to the system until after the DRIFT checks are completed. Record the information on a form similar to Figure 100.1-5.
 2. If run is invalid (sampling system BIAS specs exceeded), repeat entire procedure before repeating run.
 3. If both the zero and upscale calibration values are within the sampling system BIAS specs, then use the average of the initial and final BIAS check values to calculate the gas concentration for the run.
 4. If the zero or upscale calibration DRIFT exceeds the DRIFT limits, repeat entire procedure before conducting additional runs.
- **Post Run Leak Check:**

APPENDIX B FUEL DATA SHEETS

"F" FACTOR EXHAUST GAS FLOWRATE CALCULATION

Client : All American Asphalt
Site : Irvine
Unit : Hot Oil Heater Low

Date : 3/24/2021
Job# : 7019
Lab# : 213-039

"F" Factor, Q(std): dscf/MMBTU

Standard Temperature, T std: deg. F

Run #1

Oxygen, %	9.54
Fuel Usage, dscfm (avg)	23.3
MMBTU/min	0.024
Stack Gas Flowrate, dscfm	392

formulae:

$$\text{MMBTU/min} = (\text{Fuel Usage} * 1050 \text{ Btu/ft}^3)/10^6$$

$$\text{Flowrate ("F" Factor)} = \text{"F" Factor} * \text{MMBTU/min} * (20.9/(20.9-\text{O}_2))$$

"F" FACTOR EXHAUST GAS FLOWRATE CALCULATION

Client : All American Asphalt
Site : Irvine
Unit : Hot Oil Heater max

Date : 3/24/2021
Job# : 7019
Lab# : 213-039

"F" Factor, Q(std): dscf/MMBTU

Standard Temperature, T std: deg. F

Run #1

Oxygen, %
Fuel Usage, dscfm (avg)
MMBTU/min
Stack Gas Flowrate, dscfm

8.49
70.0
0.074
1078

formulae:

$$\text{MMBTU/min} = (\text{Fuel Usage} * 1050 \text{ Btu/ft}^3)/10^6$$

$$\text{Flowrate ("F" Factor)} = \text{"F" Factor} * \text{MMBTU/min} * (20.9/(20.9-\text{O}_2))$$

Asphalt Plant - Field Record Sheet

Client: All American Asphalt
Location: Fairfax

Test Date: 3-24-21
Lab #: 221-028
Plant Operator: -

Meter #1				
Serial #:	<u> </u>			
Manufacturer:	<u>Honeywell</u>			
Units:	<input checked="" type="radio"/> ccf	<input type="radio"/> mcf	<input type="radio"/> mmcf	other: _____
Description:	<u>Totalizer</u>			
Temp. Cor.:	<input checked="" type="radio"/> Yes	No		
Pres. Cor.:	<input checked="" type="radio"/> Yes	No		

Meter #2				
Serial #:	—			
Manufacturer:	<u>Eaton</u>			
Units:	<input checked="" type="checkbox"/>	mcf	mmcf	other: _____
Description:	<u>Analog</u>			
Temp. Cor.:	Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
Pres. Cor.:	Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>

Total Fuel Consumption:

Run #1

Run #2

Run #2

High
Low

$$\frac{7 \times 100 = 700}{30} = 23.3 \text{ cm}$$

28

10/2000 (Zn = 70)

Total Production During:

第1頁

Page 42

Page 40

High
Low

33.01%

98%

APPENDIX C METHODS 1- 4 DATA SHEETS

CARB Method 1-4 Flow Rate Determination Calculations

Client : All American Asphalt
 Site : Irvine
 Unit : Hot Oil Heater Low
 Run # : 1

Date : 3/24/2021
 Job # : 1064
 Lab # : 221-028
 Temp (Tstd): 60

***** SOURCE FIELD DATA*****

Pbar	Barometer	29.96
Y	Meter Calibration Fac.	1.0078
Cp	Pilot Calibration Fac.	0.84
Pg	Stack Static Pressure(in. H2O)	-0.01
dcO2	Dry Concentration Oxygen	9.5
dcCO2	Dry Concentration Carbon Monoxide	6.1
tstd	Area Standard Temperature (deg F)	60
ts	Temperature of Stack Gas (deg.F)	239.9
tm	Temperature of Meter (deg.F)	68.3
Delta P	Delta P Average (in H2O)	0.240
sqrtDP	Average Square root Delta P	0.490
Vw	Total Volume of Condensable water (g)	3.0
Vm	Dry gas Volume Measured (dcf)	21.819
Ds	Stack Diameter (in.)	20
Time	Sample duration (min)	30

***** INTERMEDIATE CALCULATIONS*****

Ps	Absolute Stack Pressure (in.Hg)	29.96	Ps =Pbar+Pg/13.6
Tstd	Area Standard Temperature (deg R)	520	Tstd =tstd+460
Ts	Temperature of Stack Gas (deg.R)	699.9	Ts =ts+460
Tm	Temperature of Meter (deg.R)	528.3	Tm =tm+460
Vwstd	Volume of water vapor standard (scf)	0.14	Vwstd =Vlc*.04715
Vmstd	Sample gas volume (dcf)	2.5113	Vmstd =Vm*Y*(Tstd/Tm)*((Pbar+DH/13.6)/29.92)
Bws	Moisture Content Stack Gas	5.33%	Bws =Vwstd/(Vwstd+Vmstd)
dcN2	Dry Concentration Nitrogen	84.3	dcN2=100-((dcO2)+(dcCO2))
Md	Molecular Weight Stack Gas (dry)	29.361	
Ms	Molecular Weight Stack Gas (wet)	28.755	Ms =(Md*(1-Bws))+18*Bws
As	Area of Stack (Ft^2)	2.18	As =3.141592654*(Ds/12)^2/4

***** RESULTS*****

Vs	Stack Gas Velocity (ft/sec)	31.7	Vs =85.49*Cp*sqrtDp*(SQRT(Ts/(Ps*Ms)))
Qa	Stack Gas Flow Rate (Acfm)	4,150	Qa =Vs*60*As
Qstd	Stack Gas Flow Rate (Dscfm)	337	Qstd =60*(1-Bws)*Vs*As*(Tstd/Ts)*(Ps/29.92)

X AIR TESTING



Plant : Allergan Inc Aspato
Location : Tijuana
Unit : O/S Heater
Date : 3-24-21
Run # : 1 LOW
Cold Box : 3
Meter # : J
ter Factor : 1.0075

Amb. Temp : 34°
 Pbar : 29.96
 Pitot : HT
 Pyro : HT
 Mag Δ P : 114.6400
 Mag Δ H : 114.6400
 % H₂O : —
 Box Heat : —

Nozzle : 3/8
Prob Heat : -
Wind Vel. : ~~0-10~~ 0-10
Static Press. : -101
O2 : 9.5
CO2 : 6.1
Engineer : KK
Technician : ST

Stack Dia. : 20
 "A" : 20
 "B" : 20
 Port Size : 1
 Offset : 2
 M / F : F

Stack Sample Port Location



Imp.	Gross	Tare	Total
1	689.0	687.6	
2	615.4	615.3	
3	581.9	581.5	
4	698.3	697.2	

START TIME: 8:40 END TIME: 9:10

"K" FACTOR: _____

Filter 1 :

Filter 2:

Average: 30.25.5 6.446 21.849 1.75 68.3 3

Leak Checks : Pitots

Sample Train Leak Check

Pre	Top	Bottom
ΔP	5140	01-36

Post	Top	Bottom
AP	2/4/2	2/3/2

CFM : 100 In. HG : 12
CFM : 1200 In. HG : 6

CARB Method 1-4 Flow Rate Determination Calculations

Client : All American Asphalt	Date : 3/24/2021
Site : Irvine	Job # : 1064
Unit : Hot Oil Heater max	Lab # : 221-028
Run # : 1	Temp (Tstd): 60

*** SOURCE FIELD DATA***

Pbar	Barometer	29.96
Y	Meter Calibration Fac.	1.0078
Cp	Pitot Calibration Fac.	0.84
Pg	Stack Static Pressure(in. H2O)	-0.02
dcO2	Dry Concentration Oxygen	8.5
dcCO2	Dry Concentration Carbon Monoxide	6.7
tsd	Area Standard Temperature (deg F)	60
ts	Temperature of Stack Gas (deg.F)	400.2
tm	Temperature of Meter (deg.F)	73.1
Delta P	Delta P Average (in H2O)	2.554
sqrtDP	Average Square root Delta P	1.598
Vw	Total Volume of Condensable water (g)	2.6
Vm	Dry gas Volume Measured (dcf)	22.115
Ds	Stack Diameter (in.)	20
Time	Sample duration (min)	30

*** INTERMEDIATE CALCULATIONS***

Ps	Absolute Stack Pressure (in.Hg)	29.96	Ps =Pbar+Pg/13.6
Tstd	Area Standard Temperature (deg R)	520	Tstd =tsd+460
Ts	Temperature of Stack Gas (deg.R)	860.2	Ts =ts+460
Tm	Temperature of Meter (deg.R)	533.1	Tm =tm+460
Vwstd	Volume of water vapor standard (scf)	0.12	Vwstd =Vlc*.04715
Vmstd	Sample gas volume (dscf)	2.5227	Vmstd =Vm*Y*(Tstd/Tm)*((Pbar+DH/13.6)/29.92)
Bws	Moisture Content Stack Gas	4.63%	Bws =Vwstd/(Vwstd+Vmstd)
dcN2	Dry Concentration Nitrogen	84.8	dcN2=100-(dcO2)+(dcCO2))
Md	Molecular Weight Stack Gas (dry)	29.407	Md =(dcCO2*0.44)+(dcO2*0.32)+(dcN2*0.28)
Ms	Molecular Weight Stack Gas (wet)	28.879	Ms =(Md*(1-Bws))+18*Bws
As	Area of Stack (Ft^2)	2.18	As =3.141592654*(Ds/12)^2/4

*** RESULTS***

Vs	Stack Gas Velocity (ft/sec)	114.4	vs =85.49*Cp*sqrtDp*(SQRT(Ts/(Ps*Ms)))
Qa	Stack Gas Flow Rate (Acfm)	14,976	Qa =Vs*60*As
Qstd	Stack Gas Flow Rate (Dscfm)	998	Qstd =60*(1-Bws)*Vs*As*(Tstd/Ts)*(Ps/29.92)

X AIR TESTING



Plant : All American Asphalt
Location : Turin
Unit : Oil Heater
Date : 3-24-21
Run # : Z MAX
Cold Box : Z
Meter # : 5
Meter Factor : 1.0078

Amb. Temp : >5
 Pbar : 24.95
 Pitot : AT
 Pyro : AT
 Mag Δ P : 1111111111111111
 Mag Δ H : 1111111111111111
 % H2O : -
 Box Heat : -

Nozzle : 38
Prob Heat :
Wind Vel. : 0-10
Static Press. : -102
O2 : 8.5
CO2 : 6.7
Engineer : KK
Technician : RET

Stack Dia.: 20
"A": 20
"B": 42
Port Size: 1
Offset: 2
M / F: 1

Stack Sample Port Location



START TIME: 9:39 END TIME: 10:01

"K" FACTOR:

Imp.	Gross	Tare	Total
1	646.9	646.8	
2	629.9	629.7	
3	538.1	538.1	
4	794.2	792.1	

Filter 1 : _____

2.6

Filter 2 :

Average: 30 400.2 1,588 3,215 1,75 23.1 3.0

Leak Checks : Pitots

Pre	Top	Bottom
ΔP	0/4.1	0/-3.8

Post	Top	Bottom
ΔP	0/4/3	07-40

Sample Train Leak Check

CFM : 0.00 In. HG : 15
CFM : 0.00 In. HG : 5

Method 1: Sample and Velocity Traverses for Stationary Sources

Principle:

To aid in the representative measurements of pollutant emissions and/or total volumetric flow rate from a stationary source, a measurement site where the effluent stream is flowing in a known direction is selected and the cross-section of the stack is divided into a number of equal areas. A traverse point is then located within each of these equal areas.

Not applicable when:

- 1) The flow is cyclonic,
- 2) The stack is smaller than 0.30 meter (12 in.) or 0.071 m^2 (113 in. 2)
- 3) The measurement site is < 2 diameter downstream or < $\frac{1}{2}$ diameter upstream from a flow disturbance

Rectangular stack: → Equivalent diameter: $D_e = 2LW / (L+W)$

Cyclonic flow: → May exist after such devices as cyclones, demisters and venturi scrubbers.

- If the pitot reading is not zero at the 0° reference, rotate the pitot tube until a null reading is obtained.
- Record the value of the rotation angle (α) to the nearest degree.
- Repeat for each traverse point.
- If the average value of α is $< 20^\circ$, flow is not cyclonic (EPA, CARB)
- If the average value of α is $< 10^\circ$, flow is not cyclonic (SCAQMD)

Note

- For stacks > 24 in., no traverse point shall be located within 1 in. of the stack wall
- For stacks < 24 in., no traverse point shall be located within $\frac{1}{2}$ in. of the stack wall

EPA Method 2, CARB 2, SCAQMD Method 2.1

Determination of Stack gas Velocity and Volumetric Flow Rate

Principle: The average gas velocity in a stack is determined from the gas density and from measurement of the average velocity head with a Type S or standard pitot tube.

Not applicable when:

- 1) The flow is cyclonic
- 2) The stack is smaller than 0.30 meter (12 in.) or 0.071 m^2 (113 in.^2);
- 3) The measurement site is < 2 diameter downstream or $\frac{1}{2}$ diameter upstream from a flow disturbance.

What to do when confronted with one of these three situations:


- install straightening vanes;
- calculate volumetric flow rate stoichiometrically;
- move to another measurement site where flow is acceptable.

Apparatus

- Type S Pitot Tube → shall have known coefficient and identification number engraved on it
- Differential Pressure Gauge → inclined manometer
- Temperature Gauge → Thermocouple
- Barometer
- Gas density determination equipment → Method 3 (gas molecular weight), Method 4 or 5 (moisture)


Procedure

- Set up as shown in the figure;
- Conduct leak check:
 1. blow thru pitot impact opening until at least 3 inches H₂O velocity pressure registers
 2. Close off impact opening; The pressure shall remain stable for 15 seconds
 3. Do the same for the static pressure side, except using suction to obtain the minimum 3 inches
- Level and zero the manometer
- Measure the velocity head and temperature at the traverse points; conduct a post-test leak check
- Measure the static pressure in the stack (one reading is enough)
- Determine the atmospheric pressure
- Determine stack gas molecular weight:
 1. For combustion processes → perform independent measurements for O₂, CO₂ and H₂O concentration
 2. For processes emitting essentially air → use 29 g/mole



The stack gas velocity is determined from the measured average velocity head, the measured dry concentrations of O₂ and CO₂ and the measured concentration of H₂O.

EPA Method 4, SCAQMD Method 4.1, CARB Method 4



DRY GAS METER CALIBRATION

Standard Pressure
 Standard Temperature
 Ambient pressure
 Ambient temperature

<u>29.92</u>	in. hg.
<u>60</u>	F
<u>29.96</u>	in. hg.
<u>65</u>	F

Unit Number J
 Date: 1/4/2021
 Leak Check: .004 @ 20"

ΔH in. H ₂ O	TIME	VOL. cf	WET GAS		DRY GAS				Temperature				*Y	†ΔH@ in. H ₂ O
			VOL. in/out cf	W.G. AVG F	D.G. IN F	D.G. OUT F	D.G. AVG. F							
0.75	9.94	5.000	818.537	62.0	61.0			61.8	1.0034	1.6419				
			823.518	61.0	63.0	61.0								
0.75	9.94	5.000	823.518	61.0	61.0			61.8	1.0030	1.6419				
			828.501	61.0	63.0	62.0								
0.75	9.96	5.000	828.501	62.0	62.0			62.5	1.0126	1.6462				
			833.444	61.0	64.0	62.0								
1.50	7.30	5.000	833.512	63.0	62.0			62.5	1.0042	1.7686				
			838.487	61.0	63.0	62.0								
1.50	7.32	5.000	838.487	63.0	63.0			63.0	1.0054	1.7766				
			843.461	61.0	64.0	62.0								
1.50	7.32	5.000	843.461	63.0	63.0			63.3	1.0051	1.7792				
			848.434	61.5	64.0	63.0								
2.25	5.92	5.000	848.513	64.0	63.0			63.5	1.0078	1.7447				
			853.466	61.5	64.0	63.0								
2.25	5.92	5.000	853.466	64.0	63.0			64.0	1.0078	1.7464				
			858.419	62.0	65.0	64.0								
2.25	5.95	5.000	858.419	65.0	64.0			64.5	1.0083	1.7625				
			863.374	62.0	65.0	64.0								
3.00	5.18	5.000	863.444	65.0	64.0			64.8	1.0085	1.7871				
			868.382	63.0	66.0	64.0								
3.00	5.18	5.000	868.382	65.0	64.0			64.8	1.0099	1.7871				
			873.313	63.0	66.0	64.0								
3.00	5.15	5.000	873.313	66.0	64.0			64.8	1.0091	1.7664				
			878.248	63.0	66.0	64.0								
3.75	4.63	5.000	878.275	66.0	64.0			65.0	1.0119	1.7838				
			883.190	63.0	66.0	64.0								
3.75	4.62	5.000	888.098	64.0	65.0	64.0		64.5	1.0104	1.7846				
			888.098	65.0	65.0	64.0								
3.75	4.62	5.000	893.008	64.0	65.0	64.0		64.5	1.0100	1.7846				
			AVERAGE											

Validity checks:

Meter Factor: 1.0078

* Y(max - min) ≤ .02 ?

† | ΔH@ - ΔH@ avg. | ≤ .20 in. H₂O ?

ΔH@ : 1.7468

Calibration by: FT

Reviewed by: KK

EQUATIONS USED:

$$Y = (\text{VWG} * \text{PBAR} * (\text{TDGavg} + 460)) / ((\text{VDG} * (\text{PBAR} + (\Delta H / 13.6))) * (\text{TWGavg} + 460))$$

$$\Delta H@ = ((0.0319 * \Delta H) / (\text{PBAR} * (\text{TDGavg} + 460))) * (((\text{TWG} + 460) * T) / \text{VWG})^2$$

TYPE S PITOT TUBE INSPECTION SHEET

CAL DATE: 1/5/2021

NEXT DUE DATE: JULY 2021

PITOT ID: PT-HT

Parameter	Values	Allowable Range
Level and Perpendicular?	Yes OR No	Yes
Obstruction?	Yes OR No	No
Damaged?	Yes OR No	No
α_1	2	$-10^\circ \leq \alpha_1 \leq +10^\circ$
α_2	2	$-10^\circ \leq \alpha_2 \leq +10^\circ$
β_1	3	$-5^\circ \leq \beta_1 \leq +5^\circ$
β_2	2	$-5^\circ \leq \beta_2 \leq +5^\circ$
γ	1	NA
θ	3	NA
$Z = A (\tan \gamma)$	0.015	≤ 0.125 in.
$W = A (\tan \theta)$	0.045	≤ 0.031 in.
D_t	0.375	$0.188 \leq D_t \leq 0.375$
A	0.85	NA
$A/2/(D_t)$	1.13	$1.05 \leq PA/D_t \leq 1.5$

Certification:

I certify that this pitot tube meets or exceeds all specifications, criteria and/or applicable design features and is hereby assigned a pitot tube calibration factor C_p of 0.84.

Certified By: FT Date: 1/5/2021

PYROMETER CALIBRATION

Date: 1/5/2021

Unit: HT-1

Point	* Standard Temperature <i>Tstd</i>		Pyrometer Temperature <i>Tpyr</i>		Error %
		deg. F		deg. F	
1 Ambient		65		65	0.10%
2 Ice		32		32	0.02%
3 Boil		212		210	0.30%

Std. Corr. Factor

Calibration by: FT

*Standard ID: T-1

Reviewed by: KK



Measurement Control Systems
1331 South Lyon Street
Santa Ana, CA 92705
800.826.1682
mcsometers.com

Serial Number	77003573
Test Date	2/3/2021
Record No	4304

Page 1 of 1

Gas Meter Accuracy Certification


Customer Meter	ALL AMERICAN ASPHALT American Meter 3.5M RABO 2" Rotary	Customer P/O No. Job No. RGA No.	11525 S/C
Meter Location	RUBBER PLANT-IRVINE	Technician - License No.	RUDY LUCERO 30278
Options	MINI MAX-PT		
Totalizer Readings		Test Status As Found/As Left	Next Test Date 2/3/2022
Before Test	31032 x 100	Units Cubic Feet	Capacity (cfh) 3500
After Test	31041 x 100	ODO Multiplier	Turbine Meter Spin (seconds) 100 N/A

Dresser Model 5m Prover				Customer Meter			
Test No	Flow Rate	Test Amount	Prover Amt %	Meter Reading	Meter Error	Meter Accuracy	Correction Factor
1	3497	100.00	100.05	99.950	-0.05%	99.95	1.0005
2	3514	100.00	99.87	100.130	0.13%	100.13	0.9987
3	3503	100.00	100.00	100.000	0.00%	100.00	1.0000
4	1926	100.00	99.96	100.040	0.04%	100.04	0.9996
5	1920	100.00	100.08	99.920	-0.08%	99.92	1.0008
6	1920	100.00	99.99	100.010	0.01%	100.01	0.9999
7	349	20.00	99.95	20.010	0.05%	100.05	0.9995
8	349	20.00	99.91	20.018	0.09%	100.09	0.9991
9	349	20.00	99.87	20.026	0.13%	100.13	0.9987

Overall Average Accuracy is 100.036%
Overall Correction Factor is 0.9996%

Test Report - Pass

Test Medium - Atmospheric Air



Additional Information	Approved By:
------------------------	--------------

Approved By:

60