

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS

Page 1 of 2


TO	Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO	<u>1623626</u>
SAMPLE DESCRIBED AS	Grab sample Canister 54124	DATE RECEIVED	<u>08/23/2016</u>
SAMPLING LOCATION	Firmin Drill Site 323 Firmin Street Los Angeles	FACILITY ID NO	<u>NA</u>
		REQUESTED BY	<u>Sumner Wilson</u>
		ST NO / PROJECT	<u>NA</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54124</u>
Pressure (Torr)	760
CO, ppm	6
CH ₄ , ppm	2
CO ₂ , ppm	553
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 9/2/16

Approved By: _____


Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

Alan 8/23/16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1623626</u>
Grab sample Canister 54124	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC106783	MDL				
CO, ppmvC	0.4	1.92	2.29	19.19	PASS
CH ₄ , ppmvC	0.3	2.02	1.86	-8.20	PASS
CO ₂ , ppmvC	0.4	1.57	9.78	523.01	FAIL
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.78	-12.42	PASS
NM/NEOC, ppmvC	0.5	2.03	1.77	-12.54	PASS

		Theoretical	Measured	Percent Difference	QC Limit ±5% or ± 1
CC12628	MDL				
CO, ppmvC	0.4	1036	1024	-1.11	PASS
CH ₄ , ppmvC	0.3	1068	1021	-4.41	PASS
CO ₂ , ppmvC	0.4	1022	1018	-0.39	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1010	-3.30	PASS
NM/NEOC, ppmvC	0.5	1024	990	-3.24	PASS

DATE ANALYZED	08/24/2016
REFERENCE NO:	16QM2AD
	QM2-101-120

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**

DISTRICT
 INVOICE
 LAP AUD
LABORATC

1623626



-01

TO: SCAQMD LAB: OTHER:

SOURCE NAME: Firmin Drill site I.D. No. _____

Source Address: 323 Firmin Street City: Loa Angeles

Mailing Address: _____ City: _____ Zip: 90026

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 08/23/2016

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Robert Wimmer Date: 08/22/2016 Time: 15:20-17:32

REQUESTED ANALYSIS: VOC - PAMS

-01
-02
-03

City/Location	Can#	Start day / duration	Start vac	End vac
323 Firmin St (Porch of residence) 34.063904, -118.254530	54124	08-22-16 1520 3 min	-30"	0
(Near water return for drill) 34.064048, -118.254369	54502	08-22-16 15:30 3 min	-30"	0
425 Firmin St. (background site) 34.064868, -118.253879	53489	08-22-16 17:32 3 min	-30"	-1

Relinquished by	Received by	Firm/Agency	Date	Time
R Wimmer		SCAQMD Lab	8/23/2016	15:02

Remarks: Samples were taken during drilling operations in addition to one taken at a background site in the absence of drilling activity