

BFB

Data Path : C:\msdchem\1\DATA\
Data File :
Acq On :
Operator :
Sample : IOM
Misc : (Sig #1); (Sig #2)
ALS Vial : 99 Sample Multiplier: 1

Integration File signal 1:
Integration File signal 2:

Method :
Title :
Last Update :

Spectrum Information: Scan 2724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result
50	95	8	40	11.7	5268	PASS
75	95	30	66	41.3	18656	PASS
95	95	100	100	100.0	45192	PASS
96	95	5	9	7.6	3427	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	109.2	49344	PASS
175	174	4	9	7.1	3526	PASS
176	174	93	101	93.1	45960	PASS
177	176	5	9	6.4	2924	PASS

11D0831E THF.M Tue Jan