



AQ-SPEC

Air Quality Sensor Performance Evaluation Center

Field Evaluation Report for

Bettair Static Node MK2 Series

Report ID: F20251101.0

Published on:

11/12/2025

Published by:

South Coast Air Quality Management District
21865 Copley Drive
Diamond Bar, CA 91765

Citation:

South Coast Air Quality Management District (2025). Air Quality Sensor Performance Evaluation Center (AQ-SPEC). Field Evaluation Report for Bettair Static Node MK2 Series [Revision 0]. Available at <http://www.aqmd.gov/aq-spec>

Performance Snapshot

Bettair Static Node MK2 Series




Tested: 02/28/2025 to 04/24/2025 (CO, NO, NO₂, O₃)
07/10/2025 to 09/03/2025 (PM₁, PM_{2.5}, PM₁₀)

✓ CO	○ SO ₂
○ CO ₂	○ VOC
○ CH ₄	✓ PM ₁
○ H ₂ S	✓ PM _{2.5}
○ NH ₃	○ PM ₄
✓ NO	✓ PM ₁₀
✓ NO ₂	○ UFP
✓ O ₃	○ BC

✓ Tested ○ Not tested ○ Not available

Does it hit the target?



The closer the sensor lands near the center, the more accurate its readings were compared to the pollution levels.

How well does it track?




A longer bar means the sensor did a better job of tracking the real changes in air pollution levels — going up when the concentration levels went up, and down when they dropped.



Cost



Web portal



Weight
3.3 lbs



Display



Battery



Solar



Weatherproof



Wi-Fi



Cellular



Bluetooth



Internal memory



Serial




USB




Ethernet

Performance Snapshot Guide




Pollutant List: This list shows the pollutants that the sensor is capable of measuring. Pollutants highlighted in blue with a check mark were tested for performance, those in gray were not tested, and those in white are not measured by the sensor.

◆ **CH₄** methane ◆ **CO** carbon monoxide ◆ **CO₂** carbon dioxide ◆ **H₂S** hydrogen sulfide ◆ **NO** nitric oxide ◆ **NO₂** nitrogen dioxide ◆ **NO_x** nitrogen oxides ◆ **O₃** ozone ◆ **SO₂** sulfur dioxide ◆ **VOC** volatile organic compounds ◆ **BC** black carbon ◆ **PM₁** mass of particles smaller than 1 micrometer ◆ **PM_{2.5}** mass of particles smaller than 2.5 micrometers ◆ **PM₄** mass of particles smaller than 4 micrometers ◆ **PM₁₀** mass of particles smaller than 10 micrometers ◆ **UFP** ultrafine particles, smaller than 0.1 micrometers



Target Graphic: The closer the sensor “hits” the center, the closer the sensor’s readings were to the actual concentrations. If the sensor hit inside the center circle, its readings were within 20% of the actual concentration. Each ring going outward is another 20% further from the actual concentration. If the sensor falls off the target entirely, its readings were either zero, or more than twice the actual concentration!

More technically, the distance from the center is calculated from sensor-reference relative absolute errors, averaged across all 1-hour means, averaged across the number of sensor units tested. These distances are precise and not binned in 20% intervals.



Bar Graphic: The longer the bar, the better the sensor followed the ups-and-downs of the actual concentrations. A long bar doesn’t always mean that the sensor exactly “matched” the actual concentration, but it does mean the sensor was responding when the air was clean or dirty. A long bar also means it’s possible to adjust the sensor’s readings to match the actual concentrations if you can gather data side-by-side with a reference monitor to make a formula to correct the readings!

More technically, the bar length ranges between 0 to 1 and is calculated from sensor-reference coefficients of determination (R^2 ; square of the Pearson correlation coefficient), with 1-hour means, averaged across the number of sensor units tested.



Feature Symbols: Some sensors can be configured with extra features. The price we list in the reports was the price for the product version we tested. Your price may vary from ours. If a symbol has the word **option** in it, it means the manufacturer offers that option at no extra cost. If a symbol has a small \$ sign in it, that means it is a paid option.

The number of \$ signs used for sensor “cost” is based on the 2022 average cell phone price of \$735 (<https://www.wsj.com/business/telecom/how-much-is-too-much-for-a-smartphone-3a300905>), adjusted for inflation for the year we tested the sensor. One \$ sign means the sensor cost less than an average cell phone; two \$\$ signs means it cost less than twice an average cell phone; three \$\$\$ signs means it cost more than twice an average cell phone. For other options, only one \$ sign is used for simplicity as it is too complicated to describe the variety of add-on costs through symbols.

Revision History

Version	Date	Note
0	11/12/2025	Original issued report

Disclaimer: All documents, reports, data, and other information provided are for informational and/or educational use only.

Some sensors evaluated by AQ-SPEC were field-tested inside a custom-made aluminum enclosure to protect the sensors from windblown rain, harsh sunlight, and animals. The field evaluation reports contain data collected at an air monitoring station during a specific 30- to 60-day period and cannot be duplicated at a different location, season, or time period. As sensor performance may be affected by time- and location-specific environmental conditions at the test site, replication and/or duplication of results may not be possible to achieve. The sensor assembly, installation, and use can also impact the performance of products evaluated by AQ-SPEC. No sensor calibration was performed by South Coast AQMD staff for this evaluation. Laboratory chamber testing may be necessary to fully evaluate the performance of these sensors under controlled temperature, humidity, pollutant, and interferent concentrations.

South Coast AQMD makes no claim, warranty, or guarantee that these devices will or will not work when operated by other users for their specific applications.

South Coast AQMD's AQ-SPEC aims at providing information to and for the benefit of the public to make informed purchasing decisions on air quality sensors. In accordance with this mission, the general policy of the Governing Board of the Agency is to exclude all commercial advertising and promotional material, including links which provide exclusive private or financial benefit to commercial, non-public enterprises and which do not promote or enhance a public benefit to the general public. As a Government Agency, the South Coast AQMD neither endorses nor supports individual private commercial enterprises through testing of products by AQ-SPEC or through providing links to the sites of such commercial enterprises.

Report Role	Name	Date
Tested by	Jenna Drewitz, Victor Rocha Jr., Leslie Garcia, and Randy Lam	09/03/2025
Analysis by	Namrata Shanmukh Panji, Ph.D.	10/24/2025
Quality Control Review by	Michelle Kuang, Ph.D.	10/30/2025
Approved by	Wilton Mui, Ph.D.	11/07/2025
Revision by	-----	-----

Table of Contents

Section	Topic	Page Number
1	Background	7
2	Manufacturer Specs	7
3	CO	11
	3.1 Data Overview	11
	3.2 Data Recovery	12
	3.3 Intra-model variability	12
	3.4 Linearity (R^2)	13
4	NO	15
	4.1 Data Overview	15
	4.2 Data Recovery	16
	4.3 Intra-model variability	16
	4.4 Linearity (R^2)	17
5	NO₂	19
	5.1 Data Overview	19
	5.2 Data Recovery	20
	5.3 Intra-model variability	20
	5.4 Linearity (R^2)	21
6	O₃	23
	6.1 Data Overview	23
	6.2 Data Recovery	24
	6.3 Intra-model variability	24
	6.4 Linearity (R^2)	25
7	PM₁	27
	7.1 Data Overview	27
	7.2 Data Recovery	28
	7.3 Intra-model variability	28
	7.4 Linearity (R^2)	29

Table of Contents

Section	Topic	Page Number
8	PM_{2.5}	31
	8.1 Data Overview	31
	8.2 Data Recovery	32
	8.3 Intra-model variability	32
	8.4 Linearity (R²)	33
9	PM₁₀	35
	8.1 Data Overview	35
	8.2 Data Recovery	36
	8.3 Intra-model variability	36
	8.4 Linearity (R²)	37
10	Summary Metrics	39

Section 1: Background

Three Bettair Static Node MK2 Series (hereinafter Bettair Node) units (IDs: 79, 80, 82) were deployed at the South Coast AQMD stationary ambient monitoring site in Rubidoux, CA from 02/28/2025 to 04/24/2025 for gases (CO, NO, NO₂, and O₃) and 07/10/2025 to 09/03/2025 for particulates (PM₁, PM_{2.5}, and PM₁₀). The evaluation period lasted 8 weeks. The sensor units were co-located with reference grade instruments as described below.



Bettair Static Node MK2 Series



Test site at Rubidoux, CA

Section 2: Manufacturer Specs

Parameter	Sensor: Bettair Static Node MK2 Series (raw sensor is Alphasense)	Reference Instrument: Horiba APMA 370
Pollutant	CO	CO (FRM)
Cost	\$6,330 (as-tested)	~\$10,000
Weight	3.3 pounds	35 pounds
Dimensions (LxWxH)	8.27 x 7.01 x 4.49 inches	9 x 17 x 21 inches
Power	9-36 VDC; Power over Ethernet option	100-240 VAC
Battery	Yes (3-5 days)	No
Data transmission	Ethernet, Cellular, LoRaWAN, Narrowband-IoT	Serial
Internal memory	Yes (< 1 week)	No
Operating temperature range	14-104 degrees F	32-104 degrees F
Operating RH range	15-85%	Not specified
Product website	https://bettaircities.com/bettair-node/	https://www.horiba.com/int/process-and-environmental/products/detail/action/show/Product/apma-370-453/
Operating principle	Electrochemical	Non-dispersive infrared (NDIR) absorption
Time resolution	300-seconds (as-tested)	1 minute (as-configured)
Concentration range	0-500 ppm	0-100 ppm

Section 2: Manufacturer Specs

Parameter	Sensor: Bettair Static Node MK2 Series (raw sensor is Alphasense)	Reference Instrument: Teledyne T200
Pollutant	NO, NO ₂	NO, NO ₂ (FRM), NO _x
Cost	\$6,330 (as-tested)	~\$11,000
Weight	3.3 pounds	40 lbs
Dimensions (LxWxH)	8.27 x 7.01 x 4.49 inches	7 x 17 x 23.5 inches
Power	9-36 VDC; Power over Ethernet option	100-240 VAC
Battery	Yes (3-5 days)	No
Data transmission	Ethernet, Cellular, LoRaWAN, Narrowband-IoT	Serial, Ethernet, USB
Internal memory	Yes (< 1 week)	Yes
Operating temperature range	14-104 degrees F	41-104 degrees F
Operating RH range	15-85%	Not specified
Product website	https://bettaircities.com/bettair-node/	https://www.teledyne-api.com/en-us/products/t200
Operating principle	Electrochemical	Chemiluminescence
Time resolution	300-seconds (as-tested)	1 minute (as-configured)
Concentration range	0-20,000 ppb	0-20,000 ppb

Section 2: Manufacturer Specs

Parameter	Sensor: Bettair Static Node MK2 Series (raw sensor is Alphasense)	Reference Instrument: Teledyne T400
Pollutant	O ₃	O ₃ (FEM)
Cost	\$6,330 (as-tested)	~\$7,000
Weight	3.3 pounds	31 lbs
Dimensions (LxWxH)	8.27 x 7.01 x 4.49 inches	7 x 17 x 23.5 inches
Power	9-36 VDC; Power over Ethernet option	100-240 VAC
Battery	Yes (3-5 days)	No
Data transmission	Ethernet, Cellular, LoRaWAN, Narrowband-IoT	Serial, Ethernet, USB
Internal memory	Yes (< 1 week)	Yes
Operating temperature range	14-104 degrees F	41-104 degrees F
Operating RH range	15-85%	Not specified
Product website	https://bettaircities.com/bettair-node/	https://www.teledyne-api.com/en-us/products/t400
Operating principle	Electrochemical	UV absorption
Time resolution	300-seconds (as-tested)	1 minute (as-configured)
Concentration range	0-20,000ppb	0-10,000 ppb


Section 2: Manufacturer Specs

Parameter	Sensor: Bettair Static Node MK2 Series (raw sensor is Sensirion)	Reference Instrument: Teledyne API T640
Pollutant	PM ₁ , PM _{2.5} , PM ₁₀	PM ₁ , PM _{2.5} (FEM), PM ₁₀
Cost	\$6,330 (as-tested)	~\$21,000
Weight	3.3 pounds	19 pounds
Dimensions (LxWxH)	8.27 x 7.01 x 4.49 inches	7 x 17 x 14 inches
Power	9-36 VDC; Power over Ethernet option	100-240 VAC
Battery	Yes (3-5 days)	No
Data transmission	Ethernet, Cellular, LoRaWAN, Narrowband-IoT	Serial, Ethernet, USB
Internal memory	Yes (< 1 week)	Yes (with USB flash drive)
Operating temperature range	14-104 degrees F	32-122 degrees F
Operating RH range	15-85%	0%-100%
Product website	https://bettaircities.com/bettair-node/	https://www.teledyne-api.com/en-us/products/t640
Operating principle	Optical light scattering	Optical light scattering
Time resolution	300-seconds (as-tested)	1 minute (as-configured)
Concentration range	0-1000 µg/m ³	0.1-10,000 µg/m ³


Section 3: CO

Section 3.1: Data Overview


Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Section 3: CO

Section 3.2: Data Recovery


Basic QA/QC procedures such as removal of duplicate records was performed. Nulls, negatives, out of instrument bounds as specified by the manufacturer, and values flagged as invalid by the sensor were considered invalid. Data recovery was calculated as the percent of valid readings through the entire evaluation.

Parameter	79	80	82
CO	99.1%	99.2%	96.1%

Section 3.3: Intra-model Variability

Absolute intra-model variability was calculated as the standard deviation of the mean values of the sensors. Relative intra-model variability was calculated as the absolute intra-model variability divided by the sensor grand mean. Calculations were performed using data resampled to 5-minute averages.

Parameter	Absolute intra-model variability (ppm)	Relative intra-model variability (%)
CO	0.0	3.5



Interpretation: Two out of the three Bettair Node units had similar pollutant distributions.

Section 3: CO



Section 3.4: Linearity (R^2)

Basic QA/QC procedures were used to validate the collected data (i.e., obvious outliers, negative values, readings flagged by the sensor, and invalid data-points were eliminated from the data-set).

A summary of the mean R^2 between the sensor and FRM APMA370 across all units tested.

Parameter	Time Resolution	FRM APMA370 (mean $\pm$ SD)
CO	5-minute	0.90 $\pm$ 0.00
	1-hour	0.92 $\pm$ 0.00
	24-hour	0.91 $\pm$ 0.01

Timeseries of a 1-week subset of the 8-week evaluation




Interpretation: The Bettair Node units showed very strong correlation with the corresponding FRM APMA370 data ($R^2 = 0.90$) for 5-minute averaging.


Section 3: CO

Section 3.4: Linearity (R^2)

Timeseries of a 1-week subset of the 8-week evaluation



Interpretation: The Bettair Node units showed very strong correlation with the corresponding FRM APMA370 data ($0.91 < R^2 < 0.92$) for 1-hour averaging.




Interpretation: The Bettair Node units showed very strong correlation with the corresponding FRM APMA370 data ($0.90 < R^2 < 0.91$) for 24-hour averaging.


Section 4: NO

Section 4.1: Data Overview


Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Note: The sensors were not evaluated at 24-hour averaging due to the NO reference instrument (T200) not observing at least 75% of records above that reference instrument's lower limit of detection for NO.

Section 4: NO

Section 4.2: Data Recovery


Basic QA/QC procedures such as removal of duplicate records was performed. Nulls, negatives, out of instrument bounds as specified by the manufacturer, and values flagged as invalid by the sensor were considered invalid. Data recovery was calculated as the percent of valid readings through the entire evaluation.

Parameter	79	80	82
NO	99.1%	99.2%	96.1%

Section 4.3: Intra-model Variability

Absolute intra-model variability was calculated as the standard deviation of the mean values of the sensors. Relative intra-model variability was calculated as the absolute intra-model variability divided by the sensor grand mean. Calculations were performed using data resampled to 5-minute averages.

Parameter	Absolute intra-model variability (ppb)	Relative intra-model variability (%)
NO	0.6	10.9



Interpretation: The Bettair Node units had dissimilar pollutant distributions. Note that the means are greater than the 75th percentiles due to the observed NO concentrations being very positively skewed.

Section 4: NO

Section 4.4: Linearity (R^2)



Basic QA/QC procedures were used to validate the collected data (i.e., obvious outliers, negative values, readings flagged by the sensor, and invalid data-points were eliminated from the data-set).

A summary of the mean R^2 between the sensor and T200 across all units tested.

Parameter	Time Resolution	T200 (mean $\pm$ SD)
NO	5-minute	0.92 ± 0.01
	1-hour	0.94 ± 0.01
	24-hour	N/A*

*Note: The sensors were not evaluated at 24-hour averaging due to the NO reference instrument (T200) not observing at least 75% of records above that reference instrument's lower limit of detection for NO.

Timeseries of a 1-week subset of the 8-week evaluation




Interpretation: The Bettair Node units showed very strong correlation with the corresponding T200 data ($0.90 < R^2 < 0.93$) for 5-minute averaging.

Section 4: NO

Section 4.4: Linearity (R^2)

Timeseries of a 1-week subset of the 8-week evaluation




Interpretation: The Bettair Node units showed very strong correlation with the corresponding T200 data ($0.92 < R^2 < 0.95$) for 1-hour averaging.

Note: The sensors were not evaluated at 24-hour averaging due to the NO reference instrument (T200) not observing at least 75% of records above that reference instrument's lower limit of detection for NO.


Section 5: NO₂

Section 5.1: Data Overview


Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Section 5: NO₂

Section 5.2: Data Recovery


Basic QA/QC procedures such as removal of duplicate records was performed. Nulls, negatives, out of instrument bounds as specified by the manufacturer, and values flagged as invalid by the sensor were considered invalid. Data recovery was calculated as the percent of valid readings through the entire evaluation.

Parameter	79	80	82
NO ₂	99.1%	99.2%	96.1%

Section 5.3: Intra-model Variability

Absolute intra-model variability was calculated as the standard deviation of the mean values of the sensors. Relative intra-model variability was calculated as the absolute intra-model variability divided by the sensor grand mean. Calculations were performed using data resampled to 5-minute averages.

Parameter	Absolute intra-model variability (ppb)	Relative intra-model variability (%)
NO ₂	0.9	7.7



Interpretation: Two out of the three Bettair Node units had similar pollutant distributions.

Section 5: NO₂



Section 5.4: Linearity (R²)

Basic QA/QC procedures were used to validate the collected data (i.e., obvious outliers, negative values, readings flagged by the sensor, and invalid data-points were eliminated from the data-set).

A summary of the mean R² between the sensor and FRM T200 across all units tested.

Parameter	Time Resolution	FRM T200 (mean ± SD)
NO ₂	5-minute	0.86 ± 0.02
	1-hour	0.88 ± 0.02
	24-hour	0.85 ± 0.01

Timeseries of a 1-week subset of the 8-week evaluation




Interpretation: The Bettair Node units showed strong correlation with the corresponding FRM T200 data ($0.84 < R^2 < 0.88$) for 5-minute averaging.


Section 5: NO₂

Section 5.4: Linearity (R²)

Timeseries of a 1-week subset of the 8-week evaluation



Interpretation: The Bettair Node units showed strong correlation with the corresponding FRM T200 data ($0.86 < R^2 < 0.89$) for 1-hour averaging.




Interpretation: The Bettair Node units showed strong correlation with the corresponding FRM T200 data ($0.84 < R^2 < 0.86$) for 24-hour averaging.


Section 6: O₃

Section 6.1: Data Overview


Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Section 6: O₃

Section 6.2: Data Recovery


Basic QA/QC procedures such as removal of duplicate records was performed. Nulls, negatives, out of instrument bounds as specified by the manufacturer, and values flagged as invalid by the sensor were considered invalid. Data recovery was calculated as the percent of valid readings through the entire evaluation.

Parameter	79	80	82
O ₃	99.1%	99.2%	96.1%

Section 6.3: Intra-model Variability

Absolute intra-model variability was calculated as the standard deviation of the mean values of the sensors. Relative intra-model variability was calculated as the absolute intra-model variability divided by the sensor grand mean. Calculations were performed using data resampled to 5-minute averages.

Parameter	Absolute intra-model variability (ppb)	Relative intra-model variability (%)
O ₃	1.5	4.9



Interpretation: Two out of the three Bettair Node units had similar pollutant distributions.

Section 6: O₃


Section 6.4: Linearity (R²)

Basic QA/QC procedures were used to validate the collected data (i.e., obvious outliers, negative values, readings flagged by the sensor, and invalid data-points were eliminated from the data-set).

A summary of the mean R² between the sensor and FEM T400 across all units tested.

Parameter	Time Resolution	FRM T200 (mean ± SD)
O ₃	5-minute	0.91 ± 0.01
	1-hour	0.92 ± 0.01
	8-hour	0.91 ± 0.01

Timeseries of a 1-week subset of the 8-week evaluation





Interpretation: The Bettair Node units showed very strong correlation with the corresponding FEM T400 data ($0.90 < R^2 < 0.92$) for 5-minute averaging.



Section 6: O₃

Section 6.4: Linearity (R²)

Timeseries of a 1-week subset of the 8-week evaluation



Interpretation: The Bettair Node units showed very strong correlation with the corresponding FEM T400 data ($0.91 < R^2 < 0.93$) for 1-hour averaging.




Interpretation: The Bettair Node units showed very strong correlation with the corresponding FEM T400 data ($0.9 < R^2 < 0.92$) for 8-hour averaging.


Section 7: PM₁

Section 7.1: Data Overview


Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Section 7: PM₁

Section 7.2: Data Recovery


Basic QA/QC procedures such as removal of duplicate records was performed. Nulls, negatives, out of instrument bounds as specified by the manufacturer, and values flagged as invalid by the sensor were considered invalid. Data recovery was calculated as the percent of valid readings through the entire evaluation.

Parameter	79	80	82
PM ₁	99.8%	99.8%	99.6%

Section 7.3: Intra-model Variability

Absolute intra-model variability was calculated as the standard deviation of the mean values of the sensors. Relative intra-model variability was calculated as the absolute intra-model variability divided by the sensor grand mean. Calculations were performed using data resampled to 5-minute averages.

Parameter	Absolute intra-model variability (µg/m³)	Relative intra-model variability (%)
PM ₁	0.4	4.1



Interpretation: The Bettair Node units had dissimilar pollutant distributions.

Section 7: PM₁


Section 7.4: Linearity (R²)

Basic QA/QC procedures were used to validate the collected data (i.e., obvious outliers, negative values, readings flagged by the sensor, and invalid data-points were eliminated from the data-set).

A summary of the mean R² between the sensor and T640 across all units tested.

Parameter	Time Resolution	T640 (mean ± SD)
PM ₁	5-minute	0.63 ± 0.03
	1-hour	0.63 ± 0.03
	24-hour	0.60 ± 0.05

Timeseries of a 1-week subset of the 8-week evaluation




Interpretation: The Bettair Node units showed moderate correlation with the corresponding T640 data ($0.59 < R^2 < 0.65$) for 5-minute averaging.


Section 7: PM₁

Section 7.4: Linearity (R^2)

Timeseries of a 1-week subset of the 8-week evaluation



Interpretation: The Bettair Node units showed moderate correlation with the corresponding T640 data ($0.59 < R^2 < 0.65$) for 1-hour averaging.




Interpretation: The Bettair Node units showed moderate correlation with the corresponding T640 data ($0.53 < R^2 < 0.64$) for 24-hour averaging.


Section 8: PM_{2.5}

Section 8.1: Data Overview


Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Section 8: PM_{2.5}

Section 8.2: Data Recovery


Basic QA/QC procedures such as removal of duplicate records was performed. Nulls, negatives, out of instrument bounds as specified by the manufacturer, and values flagged as invalid by the sensor were considered invalid. Data recovery was calculated as the percent of valid readings through the entire evaluation.

Parameter	79	80	82
PM _{2.5}	99.8%	99.8%	99.6%

Section 8.3: Intra-model Variability

Absolute intra-model variability was calculated as the standard deviation of the mean values of the sensors. Relative intra-model variability was calculated as the absolute intra-model variability divided by the sensor grand mean. Calculations were performed using data resampled to 5-minute averages.

Parameter	Absolute intra-model variability (µg/m ³)	Relative intra-model variability (%)
PM _{2.5}	0.4	3.9



Interpretation: The Bettair Node units had similar pollutant distributions.

Section 8: PM_{2.5}


Section 8.4: Linearity (R²)

Basic QA/QC procedures were used to validate the collected data (i.e., obvious outliers, negative values, readings flagged by the sensor, and invalid data-points were eliminated from the data-set).

A summary of the mean R² between the sensor and FEM T640 across all units tested.

Parameter	Time Resolution	FEM T640 (mean ± SD)
PM _{2.5}	5-minute	0.51 ± 0.03
	1-hour	0.52 ± 0.03
	24-hour	0.43 ± 0.05

Timeseries of a 1-week subset of the 8-week evaluation




Interpretation: The Bettair Node units showed weak to moderate correlation with the corresponding FEM T640 data ($0.47 < R^2 < 0.54$) for 5-minute averaging.


Section 8: PM_{2.5}

Section 8.4: Linearity (R²)

Timeseries of a 1-week subset of the 8-week evaluation



Interpretation: The Bettair Node units showed weak to moderate correlation with the corresponding FEM T640 data ($0.48 < R^2 < 0.54$) for 1-hour averaging.




Interpretation: The Bettair Node units showed weak correlation with the corresponding FEM T640 data ($0.37 < R^2 < 0.47$) for 24-hour averaging.


Section 9: PM₁₀

Section 9.1: Data Overview


Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Timeseries of the 8-week evaluation



Section 9: PM₁₀

Section 9.2: Data Recovery


Basic QA/QC procedures such as removal of duplicate records was performed. Nulls, negatives, out of instrument bounds as specified by the manufacturer, and values flagged as invalid by the sensor were considered invalid. Data recovery was calculated as the percent of valid readings through the entire evaluation.

Parameter	79	80	82
PM ₁₀	99.6%	99.7%	99.6%

Section 9.3: Intra-model Variability

Absolute intra-model variability was calculated as the standard deviation of the mean values of the sensors. Relative intra-model variability was calculated as the absolute intra-model variability divided by the sensor grand mean. Calculations were performed using data resampled to 5-minute averages.

Parameter	Absolute intra-model variability (µg/m³)	Relative intra-model variability (%)
PM ₁₀	1.1	3.9



Interpretation: Two out of three Bettair Node units had similar pollutant distributions.

Section 9: PM₁₀


Section 9.4: Linearity (R²)

Basic QA/QC procedures were used to validate the collected data (i.e., obvious outliers, negative values, readings flagged by the sensor, and invalid data-points were eliminated from the data-set).

A summary of the mean R² between the sensor and T640 across all units tested.

Parameter	Time Resolution	T640 (mean ± SD)
PM ₁₀	5-minute	0.14 ± 0.01
	1-hour	0.16 ± 0.02
	24-hour	0.07 ± 0.02

Timeseries of a 1-week subset of the 8-week evaluation




Interpretation: The Bettair Node units showed very weak correlation with the corresponding T640 data ($0.12 < R^2 < 0.15$) for 5-minute averaging.


Section 9: PM₁₀

Section 9.4: Linearity (R²)

Timeseries of a 1-week subset of the 8-week evaluation



Interpretation: The Bettair Node units showed very weak correlation with the corresponding T640 data ($0.13 < R^2 < 0.17$) for 1-hour averaging.



Interpretation: The Bettair Node units showed no correlation with the corresponding T640 data ($0.04 < R^2 < 0.08$) for 24-hour averaging.

Section 10: Summary Metrics

		CO		
		5-minute averages	1-hour averages	24-hour averages
Bettair Node	Average*	0.22	0.22	0.22
	SD*	0.25	0.24	0.15
	Range*	0.00 to 1.30	0.00 to 1.12	0.03 to 0.50
FRM APMA370	Average*	0.26	0.27	0.26
	SD*	0.16	0.16	0.09
	Range*	0.11 to 2.08	0.11 to 1.09	0.14 to 0.43
Bettair Node vs. FRM APMA370	R ²	0.90	0.91 to 0.92	0.90 to 0.91
	Slope	1.38 to 1.49	1.41 to 1.52	1.54 to 1.66
	Intercept*	-0.17 to -0.16	-0.18 to -0.17	-0.21 to -0.20
	MBE*	-0.06 to -0.04	-0.06 to -0.04	-0.06 to -0.04
	MAE*	0.10	0.10	0.07
	RMSE*	0.11 to 0.12	0.11	0.08 to 0.09

*Units in: ppm

Section 10: Summary Metrics

		NO		
		5-minute averages	1-hour averages	24-hour averages
Bettair Node	Average*	5.37	5.37	-
	SD*	10.89	10.50	-
	Range*	0.80 to 98.20	0.80 to 82.33	-
T200	Average*	15.21	17.45	-
	SD*	19.32	18.82	-
	Range*	0.21 to 98.37	0.45 to 77.54	-
Bettair Node vs. T200	R ²	0.90 to 0.93	0.92 to 0.95	-
	Slope	0.86 to 0.93	0.86 to 0.92	-
	Intercept*	0.43 to 2.61	0.15 to 2.64	-
	MBE*	-1.39 to 0.58	-2.03 to 0.21	-
	MAE*	3.51 to 4.30	3.27 to 4.34	-
	RMSE*	5.13 to 6.02	4.46 to 5.40	-

*Units in: ppb

Note: The sensors were not evaluated at 24-hour averaging due to the NO reference instrument (T200) not observing at least 75% of records above that reference instrument's lower limit of detection for NO.

Section 10: Summary Metrics

		NO ₂		
		5-minute averages	1-hour averages	24-hour averages
Bettair Node	Average*	11.35	11.36	11.36
	SD*	11.09	10.91	6.50
	Range*	0.50 to 67.60	0.50 to 48.10	0.60 to 23.74
FRM T200	Average*	10.79	11.08	10.73
	SD*	9.70	9.70	5.23
	Range*	0.81 to 47.14	1.25 to 44.10	3.01 to 20.23
Bettair Node vs. FRM T200	R ²	0.84 to 0.88	0.86 to 0.89	0.84 to 0.86
	Slope	1.02 to 1.08	1.03 to 1.09	1.10 to 1.16
	Intercept*	-1.05 to 0.44	-1.10 to 0.38	-1.52 to -0.22
	MBE*	-0.82 to 0.64	-0.71 to 0.78	-0.48 to 1.10
	MAE*	2.82 to 3.22	2.64 to 3.12	1.72 to 2.22
	RMSE*	3.82 to 4.40	3.55 to 4.20	2.37 to 2.96

*Units in: ppb

Section 10: Summary Metrics

		O ₃		
		5-minute averages	1-hour averages	8-hour averages
Bettair Node	Average*	31.10	31.09	31.08
	SD*	20.42	20.22	16.81
	Range*	0.50 to 108.60	0.50 to 101.03	0.50 to 73.68
FEM T400	Average*	35.77	34.82	34.80
	SD*	18.94	18.89	16.15
	Range*	0.58 to 102.17	1.03 to 95.44	1.92 to 68.43
Bettair Node vs. FEM T400	R ²	0.90 to 0.92	0.91 to 0.93	0.91 to 0.92
	Slope	0.95 to 1.07	0.95 to 1.07	0.93 to 1.05
	Intercept*	-5.97 to -3.22	-6.06 to -3.41	-4.89 to -2.72
	MBE*	-5.06 to -1.89	-5.14 to -2.11	-5.26 to -2.33
	MAE*	4.92 to 6.45	4.83 to 6.36	4.41 to 6.23
	RMSE*	6.35 to 7.93	6.16 to 7.75	5.42 to 7.24

*Units in: ppb

Section 10: Summary Metrics

		PM ₁		
		5-minute averages	1-hour averages	24-hour averages
Bettair Node	Average*	9.32	9.32	9.33
	SD*	6.53	6.46	4.78
	Range*	1.00 to 49.00	1.00 to 41.00	2.18 to 25.58
T640	Average*	9.67	9.67	9.69
	SD*	4.39	4.33	3.45
	Range*	2.02 to 46.38	2.29 to 29.40	3.49 to 21.68
Bettair Node vs. T640	R ²	0.59 to 0.65	0.59 to 0.65	0.53 to 0.64
	Slope	1.12 to 1.24	1.13 to 1.25	0.98 to 1.14
	Intercept*	-2.26 to -1.89	-2.35 to -1.97	-1.25 to -0.49
	MBE*	-0.71 to 0.05	-0.70 to 0.06	-0.67 to 0.09
	MAE*	2.94 to 3.30	2.89 to 3.25	2.32 to 2.76
	RMSE*	3.93 to 4.24	3.86 to 4.17	2.89 to 3.22

*Units in: µg/m³

Section 10: Summary Metrics

		PM _{2.5}		
		5-minute averages	1-hour averages	24-hour averages
Bettair Node	Average*	9.93	9.92	9.94
	SD*	6.95	6.87	5.06
	Range*	1.00 to 52.00	1.00 to 44.42	2.39 to 27.37
	PM _c Conc.*	16.24 to 17.49	16.24 to 17.49	16.24 to 17.51
	Fine Fraction	0.35	0.35	0.35
FEM T640	Average*	12.67	12.66	12.68
	SD*	5.21	5.12	3.98
	Range*	2.97 to 53.37	3.16 to 34.39	5.27 to 26.23
	PM _c Conc.*	25.51	25.47	25.46
	Fine Fraction	0.34	0.34	0.34
Bettair Node vs. FEM T640	R ²	0.47 to 0.54	0.48 to 0.54	0.37 to 0.47
	Slope	0.90 to 1.01	0.91 to 1.02	0.75 to 0.90
	Intercept*	-2.48 to -1.90	-2.62 to -2.02	-0.96 to 0.16
	MBE*	-3.12 to -2.33	-3.10 to -2.31	-3.06 to -2.28
	MAE*	4.54 to 4.99	4.49 to 4.95	3.75 to 4.28
	RMSE*	5.41 to 5.88	5.33 to 5.80	4.41 to 5.04

*Units in: µg/m³

Section 10: Summary Metrics

		PM ₁₀		
		5-minute averages	1-hour averages	24-hour averages
Bettair Node	Average*	26.96	26.96	26.98
	SD*	13.26	13.03	9.18
	Range*	5.00 to 105.00	6.08 to 90.25	9.97 to 56.93
	PM _c Conc.*	16.24 to 17.49	16.24 to 17.49	16.24 to 17.51
	Fine Fraction	0.35	0.35	0.35
T640	Average*	38.18	38.13	38.14
	SD*	16.58	15.07	9.80
	Range*	8.36 to 370.75	10.53 to 236.98	23.95 to 62.52
	PM _c Conc.*	25.51	25.47	25.46
	Fine Fraction	0.34	0.34	0.34
Bettair Node vs. T640	R ²	0.12 to 0.15	0.13 to 0.17	0.04 to 0.08
	Slope	0.26 to 0.32	0.30 to 0.36	0.17 to 0.28
	Intercept*	15.23 to 15.74	13.54 to 14.39	17.04 to 19.55
	MBE*	-12.35 to -10.31	-12.31 to -10.26	-12.20 to -10.16
	MAE*	14.01 to 15.12	13.73 to 14.89	11.51 to 13.31
	RMSE*	19.90 to 21.02	18.72 to 19.89	15.33 to 16.85

*Units in: µg/m³