# AQ-SPEC Air Quality Sensor Performance Evaluation Center

### Sensor Description

Manufacturer/Model: HabitatMap/AirBeam3

Pollutants: PM<sub>1.0</sub>, PM<sub>2.5</sub>, and PM<sub>10</sub> mass concentration

> Time Resolution: 1-min

#### Type: Optical



### Additional Information

#### Field evaluation report:

http://www.aqmd.gov/aqspec/evaluations/field

#### Lab evaluation report:

http://www.aqmd.gov/aqspec/evaluations/laboratory


AQ-SPEC website: http://www.aqmd.gov/aq-spec


### **Evaluation Summary**

- The accuracy of the AirBeam3 sensors for PM<sub>1.0</sub> was 74.9% to 98.9% and for PM<sub>2.5</sub> was 78.6% to 97.0% in the lab. Overall, the AirBeam3 sensors underestimated PM<sub>1.0</sub> and PM<sub>2.5</sub> measurements compared to the T640x in the lab.
- The AirBeam3 sensors exhibited high precision for all conc., T/RH combinations for PM<sub>1.0</sub> and PM<sub>2.5</sub>.
- The AirBeam3 sensors showed moderate intra-model variability for  $PM_{1.0}$  and  $PM_{2.5}$  in the lab, respectively.
- Data recovery was 99% and 100% from all units tested in the field and laboratory evaluations, respectively.
- AirBeam3 sensors showed very strong correlations for  $PM_{1.0}$ , strong to very strong correlations for  $PM_{2.5}$  and very weak correlations for  $PM_{10}$  with GRIMM and T640 from the field; and very strong correlations with the T640x in the laboratory studies ( $R^2 > 0.99$  for  $PM_{1.0}$  and  $PM_{2.5}$ ).
- The same AirBeam3 units were tested both in the field (1<sup>st</sup> stage of testing) and in the laboratory (2<sup>nd</sup> stage of testing) against reference PM instruments.

## Field Evaluation Highlights

- Deployment period 02/02/2022 04/03/2022: the AirBeam3 sensors showed very strong, strong to very strong and very weak correlations with the PM<sub>1.0</sub>, PM<sub>2.5</sub> and PM<sub>10</sub> mass concentration as recorded by GRIMM and T640, respectively.
- Data recovery from the units was ~99%.





Coefficient of Determination ( $R^2$ ) quantifies how the two sensors followed the  $PM_{1.0}$ ,  $PM_{2.5}$ , or  $PM_{10}$  concentration change by the reference instruments.

An R<sup>2</sup> approaching the value of 1 reflects a near perfect agreement, whereas a value of 0 indicates a complete lack of correlation.

## Laboratory Evaluation Highlights

### Accuracy (PM<sub>2.5</sub>)


A (%) = $100 - \frac{ \overline{X} - \overline{R} }{\overline{R}} * 100$			
Steady State #	Sensor Mean (µg/m³)	FEM T640x (µg/m³)	Accuracy (%)
1	7.7	9.8	78.6
2	54.9	50.7	91.8
3	109.6	102.4	93.0
4	193.3	199.3	97.0
5	274.1	294.4	93.1

Accuracy was evaluated by a concentration ramping experiment at 20 °C and 40% RH. The sensor's readings at each ramping steady state are compared to the reference instrument.

A negative % means sensor's overestimation by more than two fold. The higher the positive value (close to 100%), the higher the sensor's accuracy.



### Precision (PM<sub>2.5</sub>)




100% represents high precision.

Sensor's ability to generate precise measurements of  $PM_{2.5}$  concentration at low, medium, and high pollutant levels were evaluated under 9 combinations of T and RH, including extreme weather conditions like cold and dry (5 °C and 15% RH) cold and humid (5 °C and 65% RH), hot and humid (35 °C and 65% RH), or hot and dry (35 °C and 15% RH).

### Coefficient of Determination

i.



The AirBeam3 sensors showed very strong correlations with the corresponding FEM PM<sub>2.5</sub> data ( $R^2 > 0.99$ ) at 20 °C and 40% RH. For conc. ramping experiments of PM<sub>1.0</sub>, please see the lab report.

### Climate Susceptibility

From the laboratory studies, temperature and relative humidity had minimal effect on the AirBeam3 sensors' precision. Spiked concentrations were observed at the 65% RH change point.

### Observed Interferents N/A

All documents, reports, data, and other information provided in this document are for informational use only. Mention of trade names or commercial products does not constitute endorsement or recommendation. As a Government Agency, the South Coast AQMD and its AQ-SPEC program highly recommend interested entities to make use and purchase decisions based on the requirements of their study design, the technical aspects and features of their specific project applications.