



South Coast AQMD Site Survey Report for Central San Bernardino Mountains

Last updated: April 30, 2026



AQS ID	ARB Number	Site Start Date	Reporting Agency and Agency Code			
060710005	36181	10/1973	South Coast AQMD (0972)			
Site Address		County	Air Basin	Latitude	Longitude	Elevation
24171 Lake Drive Crestline, CA 92325		San Bernardino	South Coast	34.243100	-117.272350	1387m



Detailed Site Information

Local site name	Central San Bernardino Mountains			
AQS ID	060710005			
GPS coordinates (decimal degrees)	Latitude: 34.243100, Longitude: -117.272350			
Street Address	24171 Lake Drive, Crestline, CA 92325			
County	San Bernardino			
Distance to roadways (meters)	55			
Traffic count (AADT, year)	5114 / 2022			
Groundcover (e.g. asphalt, dirt, sand)	Dirt			
Representative statistical area name (i.e. MSA, CBSA, other)	40140-Riverside-San Bernardino-Ontario, CA MSA			
Pollutant, POC	Ozone, 1	PM10, 1	Continuous PM2.5, 3	
Primary / QA Collocated / Other	N/A	Primary	Other	
Parameter code	44201	81102	88101	
Basic monitoring objective(s)	NAAQS	NAAQS	General Public Info	
Site type(s)	Highest Concentration	Population Exposure	Population Exposure	
Monitor (type)	SLAMS	SLAMS	Other	
Network Affiliation	N/A	N/A	N/A	
Instrument manufacturer and model	Teledyne T400	Tisch SSI TE- PM10PLUS-BL	Met One BAM 1022	
Method code	087	141	209	
FRM/FEM/ARM/ other	FEM	FRM	Non-FEM	
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	South Coast AQMD	N/A	
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Spatial scale (e.g. micro, neighborhood)	Neighborhood	Neighborhood	Neighborhood	
Monitoring start date (MM/DD/YYYY)	10/01/1973	01/1985	07/24/2009	
Current sampling frequency (e.g.1:3, continuous)	Continuous	1:6	Continuous	
Calculated sampling frequency (e.g. 1:3/1:1)	N/A	1:6	N/A	
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	
Probe height (meters)	4.7	2.4	3.1	
Distance from supporting structure (meters)	1m	N/A	N/A	
Distance from obstructions on roof (meters)	N/A	N/A	N/A	

Distance from obstructions not on roof (meters)	N/A	N/A	N/A	
Distance from trees (meters)	7.5 m West 34.5 m Height 9.6 m East 36.4 m Height	4.1 m West 34.5 m Height 12.5 m East 36.4 m Height	1.8 m West 34.5 m Height 15.5 m East 36.4 m Height	
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	
Distance between collocated monitors (meters)	N/A	N/A	N/A	
Unrestricted airflow (degrees)	225°	225°	225°	
Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	Teflon	N/A	N/A	
Residence time for reactive gases (seconds)	16.8	N/A	N/A	
Will there be changes within the next 18 months? (Y/N)	No	No	No	
Is it suitable for comparison against the annual PM2.5? (Y/N)	N/A	N/A	N/A	
Frequency of flow rate verification for manual PM samplers	N/A	Monthly	N/A	
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	Monthly	
Frequency of one-point QC check for gaseous instruments	Nightly	N/A	N/A	
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	05/22/2025	N/A	N/A	
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	05/08/2025 10/17/2025	05/08/2025 10/17/2025	

Pollutant, POC	WS & D, 1/1	RH/T, 1/1	BP, 1	
Primary / QA Collocated / Other	N/A	N/A	N/A	
Parameter code	61101/61102	62201/62101	64101	
Basic monitoring objective(s)	Research	Research	Research	
Site type(s)	Meteorological	Meteorological	Meteorological	
Monitor (type)	SLAMS	SLAMS	SLAMS	
Network Affiliation	N/A	N/A	N/A	
Instrument manufacturer and model	RM Young 05305VP	Rotronic HC2A-S3	Met One 091	
Method code	065/065	063/063	015	
FRM/FEM/ARM/ other	N/A	N/A	N/A	
Collecting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Analytical Lab (i.e., weigh lab, toxics lab, other)	N/A	N/A	N/A	
Reporting Agency	South Coast AQMD	South Coast AQMD	South Coast AQMD	
Spatial scale (e.g. micro, neighborhood)	Neighborhood	Neighborhood	Neighborhood	
Monitoring start date (MM/DD/YYYY)	10/1973	10/1973	10/1973	
Current sampling frequency (e.g.1:3, continuous)	Continuous	Continuous	Continuous	
Calculated sampling frequency (e.g. 1:3/1:1)	N/A	N/A	N/A	
Sampling season (MM/DD-MM/DD)	01/01-12/31	01/01-12/31	01/01-12/31	
Probe height (meters)	10.0	3.5	3.5	
Distance from supporting structure (meters)	N/A	N/A	N/A	
Distance from obstructions on roof (meters)	N/A	N/A	N/A	
Distance from obstructions not on roof (meters)	N/A	N/A	N/A	
Distance from trees (meters)	11.4 m West 34.5 m Height 5.7 m East 36.4 m Height	11.4 m West 34.5 m Height 5.7 m East 36.4 m Height	11.4 m West 34.5 m Height 5.7 m East 36.4 m Height	
Distance to furnace or incinerator flue (meters)	N/A	N/A	N/A	
Distance between collocated monitors (meters)	N/A	N/A	N/A	
Unrestricted airflow (degrees)	225°	225°	360°	

Probe material for reactive gases (e.g. Pyrex, stainless steel, Teflon)	N/A	N/A	N/A	
Residence time for reactive gases (seconds)	N/A	N/A	N/A	
Will there be changes within the next 18 months? (Y/N)	No	No	No	
Is it suitable for comparison against the annual PM2.5? (Y/N)	N/A	N/A	N/A	
Frequency of flow rate verification for manual PM samplers	N/A	N/A	N/A	
Frequency of flow rate verification for automated PM analyzers	N/A	N/A	N/A	
Frequency of one-point QC check for gaseous instruments	N/A	N/A	N/A	
Last Annual Performance Evaluation for gaseous parameters (MM/DD/YYYY)	N/A	N/A	N/A	
Last two semi-annual flow rate audits for PM monitors (MM/DD/YYYY, MM/DD/YYYY)	N/A	N/A	N/A	

**Central San Bernardino Mountains
Site Photos**



Looking North from the probe.



Looking East from the probe.



Looking South from the probe.



Looking West from the probe.

**Central San Bernardino Mountains
Site Photos (Cont.)**



Looking at the probe from the North.



Looking at the probe from the East.



Looking at the probe from the South



Looking at the probe from the West.