

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
REPORT OF LABORATORY ANALYSIS


Page 1 of 2

TO Jason Low Atmospheric Measurements Manager Science and Technology Advancement	LABORATORY NO 1619006 DATE RECEIVED 07/09/2016
SAMPLE DESCRIBED AS 24 hour sample Canister 54053	FACILITY ID NO NA REQUESTED BY Sumner Wilson
SAMPLING LOCATION Reseda Station 18328 Gault St Los Angeles, CA 91335	ST NO / PROJECT NA

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

Type	Canister
Number	<u>54053</u>
Pressure (Torr)	760
CO, ppm	< 3
CH ₄ , ppm	2
CO ₂ , ppm	422
Ethane, ppmvC	< 1
NM/NEOC, ppmvC	< 1

Date Approved: 7/12/16

Approved By: 

Solomon Teffera, Acting Senior Manager
Laboratory Services
909-396-2199

AKW 7/12/16

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
21865 Copley Drive, Diamond Bar CA 91765-1482

MONITORING & ANALYSIS
QUALITY CONTROL SUMMARY

Page 2 of 2

SAMPLE DESCRIBED AS	LABORATORY NO	<u>1619006</u>
24 hour sample Canister 54053	REQUESTED BY	<u>Sumner Wilson</u>

Carbon monoxide (CO), methane (CH₄), carbon dioxide (CO₂), ethane (C₂H₆), and non-methane non-ethane organic carbon (NM/NEOC) in ppmvC by SCAQMD Method 25.1 (TCA FID).

QUALITY CONTROL -- End of run control recovery

		Theoretical	Measured	Percent Difference	QC Limit
CC106783	MDL				±5% or ± 1
CO, ppmvC	0.4	1.92	2.32	20.36	PASS
CH ₄ , ppmvC	0.3	2.02	2.04	1.00	PASS
CO ₂ , ppmvC	0.4	1.57	2.60	65.77	FAIL
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	2.03	1.89	-6.84	PASS
NM/NEOC, ppmvC	0.5	2.03	1.89	-6.79	PASS

		Theoretical	Measured	Percent Difference	QC Limit
CC12628	MDL				±5% or ± 1
CO, ppmvC	0.4	1036	1055	1.83	PASS
CH ₄ , ppmvC	0.3	1068	1053	-1.41	PASS
CO ₂ , ppmvC	0.4	1022	1038	1.53	PASS
C ₂ H ₄ , ppmvC	0.4	NA	NA	NA	NA
C ₂ H ₆ , ppmvC	0.4	1044	1034	-0.98	PASS
NM/NEOC, ppmvC	0.5	1024	1017	-0.68	PASS

DATE ANALYZED	07/11/2016
REFERENCE NO:	16QM2AD
	QM2-101-115

WO #: 1619006

**SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT
SAMPLE ANALYSIS REQUEST**



TO: SCAQMD LAB: OTHER:

SOURCE NAME: Southern California Gas Co. I.D. No. _____

Source Address: 12801 Tampa Ave City: Porter Ranch

Mailing Address: _____ City: _____ Zip: 91326

Contact Person: _____ Title: _____ Tel: _____

Analysis Requested by: Sumner Wilson Date: 7/9/16

Approved by: Jason Low Office: _____ Budget #: 44716

REASON REQUESTED: Court/Hearing Board Permit Pending Hazardous/Toxic Spill

Suspected Violation Rule(s) _____ Other

Sample Collected by: Qian Zhou Date: 7/9/16 Time: 11:30am

REQUESTED ANALYSIS: PAMS analysis

City/Location	Can#	Start day / time/ duration	Start vac	End Press
Reseda Station	54053	7/8/16 / 00:00 / 24 hours	<-30"	+15

Relinquished by	Received by	Firm/Agency	Date	Time
<i>Zhouqian</i>	<i>[Signature]</i>	SCAQMD Lab	7/9/16	13:30

Remarks: 1:3 scheduled samples from station at Reseda.
Reseda Station - 18328 Gault St, Los Angeles, CA 91335
GPS (34.199225, -118.532743). Right sampler sn 6113