

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT

**PROTOCOL FOR THE MEASUREMENT OF
NITROGEN OXIDES, CARBON MONOXIDE, AND OXYGEN
FROM SOURCES SUBJECT TO
SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT RULE 1146**

SEPTEMBER 21, 2001

MONITORING & SOURCE TEST ENGINEERING BRANCH

MONITORING & ANALYSIS

SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT

**PROTOCOL FOR THE MEASUREMENT OF
NITROGEN OXIDES, CARBON MONOXIDE, AND OXYGEN
FROM SOURCES SUBJECT TO
SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT RULE 1146**

SEPTEMBER 21, 2001

MONITORING & SOURCE TEST ENGINEERING BRANCH

MONITORING & ANALYSIS

TABLE OF CONTENTS

	PAGE
TABLE OF CONTENTS.....	i
1.0 OVERVIEW AND APPLICABILITY	1
2.0 ENVIRONMENTAL TEST CONDITIONS	3
2.1 AMBIENT TEMPERATURE.....	3
2.2 AMBIENT RELATIVE HUMIDITY	3
2.3 AMBIENT BAROMETRIC PRESSURE.....	3
3.0 COMBUSTION EQUIPMENT AND PROCESS DESCRIPTION	4
3.1 COMBUSTION EQUIPMENT DESCRIPTION.....	4
3.2 PROCESS DESCRIPTION.....	4
4.0 ANALYTICAL REQUIREMENTS AND SAMPLING PROCEDURES USING STANDARD DISTRICT METHODS.....	6
4.1 APPLICABILITY	6
4.2 DISTRICT METHOD 100.1	6
4.2.1 SAMPLE CONDITIONING REQUIREMENTS.....	6
4.2.2 INSTRUMENTAL ANALYZER PROCEDURES	8
4.2.3 FIELD SAMPLING PROCEDURE	8
4.2.4 DATA ACQUISITION SYSTEM (DAS)	9
4.3 DISTRICT METHODS 7.1 AND 10.1	9
4.3.1 FIELD SAMPLING PROCEDURES.....	10
4.3.2 LOW CONCENTRATION CO	10
4.3.3 LOW CONCENTRATION NO _x	11
5.0 ANALYTICAL REQUIREMENTS USING PORTABLE ANALYZER METHOD.....	12
5.1 START UP	12
5.1.1 ANALYZERS.....	12
5.1.2 MOISTURE REMOVAL SYSTEM.....	12
5.2 ANALYZER CALIBRATION	12
5.3 DATA RECORDING.....	13
5.4 FIELD SAMPLING PROCEDURE.....	13

TABLE OF CONTENTS

	PAGE
5.4.1 SAMPLE POINT LOCATIONS	13
5.4.2 TRAVERSE SAMPLING	13
5.4.3 CELL TEMPERATURE MONITORING.....	13
5.4.4 INTERFERENCE VERIFICATION AND RE-ZEROING	14
6.0 TESTING UNDER NON-IDEAL CONDITIONS.....	15
7.0 CALCULATIONS.....	17
7.1 HEATING VALUE.....	17
7.2 F-FACTOR CALCULATION	17
7.3 EMISSION RATE OF NO _x AND CO.....	18
7.3.1 CONCENTRATION CORRECTIONS.....	18
7.3.2 CONCENTRATION ADJUSTMENT OF PORTABLE ANALYZER DATA ...	18
7.3.3 NO _x OR CO EMISSION (Concentration at 3% O ₂)	19
7.3.4 NO _x OR CO EMISSION (lb/MMBtu).....	19
8.0 HIGH OXYGEN CONCENTRATIONS.....	20
8.1 EMISSION SAMPLING.....	20
8.2 FLOW DETERMINATION.....	20
8.2.1 VELOCITY TRAVERSES.....	20
8.2.2 STACK FLOW RATE.....	21
8.2.3 EXHAUST GAS MOLECULAR WEIGHT DETERMINATION	22
8.2.4 MOISTURE DETERMINATION.....	22
8.3 HEAT INPUT.....	22
8.3.1 HEATING VALUE	22
8.3.2 FUEL USAGE	22
8.4 CALCULATIONS	23
8.4.1 POLLUTANT MASS EMISSIONS	23
8.4.2 HEAT INPUT CALCULATION.....	23
8.4.3 NO _x OR CO EMISSION CALCULATION	23
9.0 QUALITY ASSURANCE/ QUALITY CONTROL (QA/QC)	25

TABLE OF CONTENTS

	PAGE
9.1 PRESSURE MEASUREMENT.....	25
9.2 TEMPERATURE MEASUREMENTS.....	25
9.3 FUEL GAS FLOW.....	25
9.4 NON-GASEOUS FUEL USAGE MEASUREMENTS.....	26
9.5 TIME.....	26
10.0 REPORT.....	27
10.1 REPORT INFORMATION.....	27
10.2 REPORT FORMAT.....	28
ATTACHMENT I ADDITIONAL QA/QC FOR NON-LAP APPROVED SOURCE TEST LABORATORIES	
ATTACHMENT II DISTRICT METHOD 100.1 GENERAL REQUIREMENTS AND NO ₂ TO NO CONVERTER EFFICIENCY CHECK PROCEDURE	
ATTACHMENT III EPA CONDITIONAL TEST METHOD CTM-030	

**PROTOCOL FOR THE MEASUREMENT OF
NITROGEN OXIDES, CARBON MONOXIDE, AND OXYGEN
FROM SOURCES SUBJECT TO
SOUTH COAST AIR QUALITY MANAGEMENT DISTRICT RULE 1146**

1.0 OVERVIEW AND APPLICABILITY

On November 17, 2000, the South Coast Air Quality Management District (District) amended Rule 1146, which limits emissions of nitrogen oxides (NO_x) and carbon monoxide (CO) from industrial, institutional and commercial boilers, steam generators, and process heaters with rated input capacity greater than 5 million Btu per hour. Compliance with this rule is based on District standard reference methods (Method 100.1, or Methods 7.1 and 10.1), or equivalent and approved methods that allow the use of hand-held portable monitors. This protocol has been developed to ensure standardization of test procedures including the use of: specified test conditions, required test methods, specifications for test equipment, data collection/reporting, and quality assurance requirements.

An independent testing laboratory, which meets the requirements of South Coast Air Quality Management District's Rule 304, Subdivision (k), may conduct the test. The source testing contractor must meet the non-conflict of interest requisite specified per Rule 304(k) and provide a non-conflict of interest statement in the final report. The source testing contractor need not be approved under the SCAQMD Laboratory Approval Program (LAP). If not LAP approved, detailed information including, but not limited to, the contractor's QA/QC procedures, resumes of source testers, description of the contractor's sampling system including the sampling system for SCAQMD Method 100.1, description of instruments used, and calibration procedures for all instruments are to be included in the final report for review (see Attachment I). The laboratory shall, in conducting the test, prepare a report of findings, including all raw data sheets/charts, and

analytical data. Testing must demonstrate to the satisfaction of the Executive Officer that the test data was obtained using the protocol described herein.

When equipment at a facility subject to District Rule 1146 cannot be tested using this protocol, the protocol may be modified following written approval of the Executive Officer.

2.0 ENVIRONMENTAL TEST CONDITIONS

2.1 AMBIENT TEMPERATURE

The temperature shall be monitored and recorded at fifteen minute intervals during each test.

2.2 AMBIENT RELATIVE HUMIDITY

The relative humidity shall be monitored and recorded before and after each test.

2.3 AMBIENT BAROMETRIC PRESSURE

The barometric pressure shall be monitored and recorded before and after each test.

3.0 COMBUSTION EQUIPMENT AND PROCESS DESCRIPTION

3.1 COMBUSTION EQUIPMENT DESCRIPTION

The type of combustion equipment (i.e., industrial boiler, steam generator, etc.), and control equipment shall be noted. Additionally, the rated heat input capacity and fuel type shall be recorded. Tester shall record the equipment make, model, and serial number.

Information regarding the fuel flow meter(s) which includes the make, model, serial number, and the most recent calibration date, shall also be recorded. The fuel meter calibration data shall be included with the final test report.

3.2 PROCESS DESCRIPTION

A description of the process operation during sampling shall include, but not be limited to, the following information:

- a) Provide a stepwise description explaining the entire process operation (relating only to the combustion equipment). The description shall also include a schematic of this process;
- b) Indicate whether the operation was a continuous or batch process. If it was a batch process, state the length of each batch interval, and when (within this interval) the test performed;
- c) If the equipment modulated during sampling, indicate the extent of modulation with respect to firing rate;
- d) State whether the equipment had operated normally;
- e) Report the amount of product (i.e.- steam) supplied by the equipment;
- f) If the equipment is a process heater, identify the product being heated;

- g) Include miscellaneous data, such as steam pressure, fuel pressure and temperature, firing rate, and the percent of rated capacity at which the combustion equipment operated during testing.
- h) Provide information regarding the operating schedule of the equipment. Include information such as whether the equipment operates "on demand", the peak periods of operation, and the operational hours in hours per day, days per week, and weeks per year.
- i) Submit information on the pollution control equipment, including the type of control (low NO_x burners, flue gas recirculation, selective catalytic reduction, etc.), control parameters (flue gas recirculation rate, ammonia injection rate, temperature of inlet gas to selective catalytic reduction equipment, etc.), and manufacturer of the control equipment.

**4.0 ANALYTICAL REQUIREMENTS AND SAMPLING
PROCEDURES USING STANDARD DISTRICT METHODS**

4.1 APPLICABILITY

This section applies to the use of District Methods 100.1, or District Methods 7.1 and 10.1.

4.2 DISTRICT METHOD 100.1

Simultaneous and concurrent stack gas measurements of NO_x, CO, and oxygen shall be determined by following District Method 100.1, and using the following types of analyzers:

Gas Constituent	Analyzer Principle
NO _x	Chemiluminescent
CO	Non-Dispersive Infrared (NDIR)
Oxygen	Electrochemical

The sample concentration must fall within 20 to 95% of the analyzer range for at least 95% of the time. To determine the best analyzer range, a preliminary screening of the gas concentration may be necessary. Since the carbon dioxide and oxygen concentrations in combustion products are related stoichiometrically, it is strongly recommended that carbon dioxide also be measured during gas sampling to partially verify the data quality. Carbon dioxide sampling is also addressed in District Method 100.1.

4.2.1 SAMPLE CONDITIONING REQUIREMENTS

For Method 100.1 testing, proper sample conditioning is essential for representative sampling. Sample conditioning includes removal of particulate matter and moisture in

the sample gas stream without removal of the pollutants of interest. In particular, NO₂, due to its high solubility in water, is more susceptible to scrubbing than, for example, NO or CO.

The District recommends a gas sampling system which can be used universally. The set up includes a heated 1/4 inch stainless steel probe with 50-80 μ (micron) size sintered 316 stainless steel or ceramic filter at the tip and a short (not more than 6 feet) heated Teflon line to a sample conditioning system. The temperature of the probe and the Teflon line should be maintained at about 250 °F. The conditioning system consists of a pair of standard Greenburg-Smith impingers with the stems cut to about 1 inch length from the top, immersed in a bath containing water and dry ice pellets, and immediately followed by a thermo-electric cooler or permeation drier. The gas temperature at the outlet of the impinger shall be less than 60 °F, and the gas at the drier outlet shall be maintained at a dew point less than 37 °F. If the drier cannot be directly connected to the impinger outlet then a Teflon line heated to 10 °F above the impinger outlet gas temperature can be used for connection. Another particulate filter (~ 5 μ) should be in line immediately after the cooler/drier. **All sample conditioning temperatures should be measured and recorded preferably on the strip chart at least every 10 minutes.** If the moisture content of the exhaust gas is below 5% and the sample gas flow rate is less than 10 liter/minute, the impinger set-up need not be used as long as no moisture condensation occurs in the system and the conditioned sample is maintained at the required dew point.

Precaution: Never let the water in the impingers accumulate more than 1/4 of the impinger's height. Do not let the water bath freeze around the impingers; it may cause cracking of glass impingers. Make sure that the thermo-electric cooler/permeation drier has adequate design capacity. A good maintenance schedule should be followed for the cooler/drier.

An example of a non-universally applicable water removal system is a refrigerated cooling coil based system. A refrigerated cooling coil system scrubs out a high percentage of the highly water soluble pollutants due to comparatively high residence time and intimate contact between the sample gases and water droplets collected on the inside wall of the coil. This type of sample conditioner is not suitable in all cases.

Other systems may be used, upon District's approval, as long as they meet the requirement for water removal immediately after separation from the gas stream and are designed to minimize water contact with the gas stream.

4.2.2 INSTRUMENTAL ANALYZER PROCEDURES

A general guideline for conducting Method 100.1 is included in Attachment II. The guideline is provided to clarify and emphasize some of the key points of Method 100.1 requirements.

Leak checks must be performed before and after each concentration sampling period. Follow the leak check procedure as stated in Method 100.1.

Analyzer calibration error, system bias and linearity checks must be performed before and after each concentration sampling period. A system response time test must be performed before the concentration sampling. Tester shall survey possible interferences emitted by the equipment, and perform interference response tests, if necessary.

NO₂ to NO converter efficiency check must be performed at the beginning of each test day. The converter efficiency check procedure is included in Attachment II.

4.2.3 FIELD SAMPLING PROCEDURE

Field sampling shall follow the procedures in Section 2.6 of District Method 100.1. All QA/QC requirements of Method 100.1 shall be strictly followed. All field and laboratory

data, including calculations, annotated chart traces, and calibration results, shall be submitted with the test report.

Record the process variables listed in Section 3.2. For combustion equipment with a totalizing fuel meter, volumetric readings shall be recorded at the beginning and end of the test, and once every three minutes during sampling. Combustion equipment with non-totalizing fuel meters shall have readings recorded at the beginning and end of the test, and once every minute during sampling.

4.2.4 DATA ACQUISITION SYSTEM (DAS)

A DAS must be used to report the analyzer readings. The DAS must provide every one-minute and 15-minute averages in units of ppm for NO_x and CO, and in percent for O₂ (and CO₂, if desired). The strip charts must be included in the final report as well as the DAS output.

4.3 **DISTRICT METHODS 7.1 AND 10.1**

Simultaneous and concurrent stack gas bulb samples of NO_x, CO, and oxygen shall be analyzed by the following procedures:

Gas Constituent	Analytical Principle
NO _x	Colorimetric or Ion Chromatograph
CO	Gas Chromatograph/ Non-Dispersive Infrared (GC/NDIR)
Oxygen	Gas Chromatograph-Thermal Conductivity (GC/TCD)

It is noted that an Orsat analysis for oxygen is not acceptable. Since the carbon dioxide and oxygen concentrations in combustion products are related stoichiometrically, it is strongly recommended that carbon dioxide also be measured during gas sampling to

partially verify the data quality. Carbon dioxide sampling is addressed in District Method 10.1.

4.3.1 FIELD SAMPLING PROCEDURES

Separate Methods 7.1 and 10.1 bulb samples shall be collected from a purged heat-traced sample line at three minute intervals, using the locations described in Section 5.4.1 of this protocol. For each bulb sample, identify the traverse point and record the sample times per Section 5.4.2. Field sampling shall follow procedures described in Section 2 of District Methods 7.1 and 10.1. Record the process variables and equipment firing rate per Section 4.2.3 of this protocol.

All QA/QC requirements of District Methods 7.1 and 10.1 shall be strictly followed. All field and laboratory data, including calculations, analytical chart traces (with annotations), and calibration results, shall be submitted with the test report.

4.3.2 LOW CONCENTRATION CO

Additional QA requirements are necessary for Method 10.1 if the CO concentration is less than 10 ppm. These requirements are:

- 1) The bulbs or canisters shall be purged with zero gas (nitrogen or hydrocarbon-free air) before taking any samples or blanks. Analyze the purged zero gas for CO, and reject any container that has more than 0.5 ppm CO.
- 2) Two field blanks of zero gas containing less than 0.5 ppm of CO in clean bulbs or canisters shall accompany the samples to the test site. These blanks shall be analyzed along with the field samples.
- 3) The CO shall be analyzed by Total Combustion Analysis using Flame Ionization Detection (GC/FID) according to Section 5 in District Method 25.1.
- 4) Analyze all samples within 24 hours of sampling.

4.3.3 LOW CONCENTRATION NO_x

Additional preparation and QA requirements are necessary for Method 7.1 if the NO_x concentration is less than 20 ppm. These requirements are:

- 1) Low NO_x concentrations shall be analyzed by ion chromatography using the Alternative Laboratory Procedure described in Section 5 of District Method 7.1.
- 2) Sample bulbs shall be rinsed clean with 0.1 N H₂SO₄ and deionized water before sampling. Rinse, soak, and re-rinse any rubber tubing in deionized water before sampling.
- 3) The absorbing solution of 0.1 N H₂SO₄ with 3% H₂O₂ shall be prepared daily from freshly distilled or deionized water.
- 4) Establish and report the detection limit as nitrate for each set of analyses.
- 5) Adjust the calibration curve downwards to bracket all samples.
- 6) Do not store samples more than three days before analysis.
- 7) Analyze replicate samples and average the results.
- 8) Take at least one field blank sample per set to determine if contamination occurred. Prepare a bulb with absorbing solution, and fill the bulb in the laboratory with zero gas to ambient pressure. Take the bulb into the field during testing, and return the bulb to the laboratory for analysis.

5.0 ANALYTICAL REQUIREMENTS USING PORTABLE ANALYZER METHOD

Portable analyzer sampling and analysis relies on EPA Conditional Test Method CTM-030. A copy of this method is included in Attachment III.

5.1 START UP

5.1.1 ANALYZERS

Allow analyzers to warm up according to manufacturer's instructions, and Section 7.2 of the EPA document, CTM-030.

5.1.2 MOISTURE REMOVAL SYSTEM

Energize sample pump and sample line. Allow temperatures and flows to come to equilibrium. The dew point of the conditioned gas shall not be greater than 50°F. Testing shall not be performed if the ambient temperature falls below 55°F.

5.2 ANALYZER CALIBRATION

Use NIST traceable calibration gases which are certified to a minimum accuracy of $\pm 1\%$. A high span calibration gas shall be selected according to CTM-030 procedures. Calibrate the analyzers according to CTM-030 using zero gas, mid-level gas, and high-range gas. Calibrations of each analyzer must be performed prior to and following each test on the day of sampling. Linearity may be checked using a gas divider if it is calibrated according to EPA Method 205.

The analyzers shall meet the zero calibration error, span calibration error, interference response, linearity, and stability check response specifications of CTM-030, Section 4. A fresh air purge shall be performed per manufacturer's recommendations between each stability check. The pre- and post-test zero and span calibration error checks must be performed at the sampling location according to Section 7.3 and 7.6 of CTM-030.

5.3 DATA RECORDING

The data shall be recorded using procedures in Section 5.1.10 of CTM-030. The recorder must allow each data point to be read at least once every minute.

5.4 FIELD SAMPLING PROCEDURE

5.4.1 SAMPLE POINT LOCATIONS

Traverse sampling shall be performed at five points across perpendicular axes of the exhaust stack. These points are: 16.7 and 83.3 percent of the diameter along the first axis; 16.7 and 83.3 percent of the diameter along the second axis; and, the center point.

5.4.2 TRAVERSE SAMPLING

At the start of the test, record the time and fuel flow. Do not interrupt the flow to the portable analyzer. Follow the Sample Collection procedures of CTM-030 (Section 7.5) with the following exception: at each traverse location, gas concentrations shall be recorded at one minute intervals for three consecutive minutes. Note that although fifteen minutes of testing will be conducted, the actual sampling time will be longer, depending on the response time or stability time. Identify the traverse points being sampled along with the gas concentrations at each point. Determine the arithmetic mean of the gas concentrations measured during the fifteen minute test period. If required, collect a sample of fuel for composition analysis. Record the process variables and equipment firing rate per Section 4.2.3.

5.4.3 CELL TEMPERATURE MONITORING

Not required if sensors are temperature controlled. Sensor temperature shall not exceed 125 degrees F.

5.4.4 INTERFERENCE VERIFICATION AND RE-ZEROING

Refer to Sections 7.7 and 7.8 of CTM-030. In addition to the interference checks required by CTM-030, an interference check of the NO and NO₂ sensors shall be performed for combustion devices which have a potential to emit SO₂ concentrations greater than 10 ppm. The interference check procedure should consider both positive and negative biases in the data, and shall be submitted for District written approval prior to performing the source test. Results of the interference check shall be included with the final report, and shall demonstrate (using an equation similar to that in Section 6.3.1 of CTM-030) that the combined NO and NO₂ interference responses due to SO₂ contribute less than 5% of the NO_x concentration in the stack.

6.0 TESTING UNDER NON-IDEAL CONDITIONS

The following is a discussion of some common non-ideal conditions and their solutions in source testing:

If there are fluctuations in the process or operating conditions, such as, changes in load, the testing may continue as long as the operating conditions are recorded to show each fluctuation. If the combustion equipment shuts down completely or if there are severe fluctuations during sampling, testing must be repeated for a minimum of 15 continuous minutes during acceptable operating conditions. The severity of fluctuations will be determined on a case-by-case basis. If the device has the tendency to operate at a non-ideal condition for testing, a protocol for testing at such condition should be submitted for review and approval prior to testing.

If there are dampers or bypass stacks present, testing should be conducted as follows:

- 1) If excess air is frequently introduced to the exhaust stack at a variable rate, concentration testing shall be performed while no excess air is introduced to the exhaust stack.
- 2) When all or part of the exhaust gas is frequently restricted to or bypassed from the exhaust stack, concentration testing shall be performed while the total exhaust gas is routed to the exhaust stack (except for flue gas recirculation).

All changes in process and operating conditions and test interruptions must be noted with the beginning and ending times of each occurrence on the field data sheets.

For multiple stacks, perform sampling at each of the stacks during acceptable testing conditions. Sampling at each stack need not be performed simultaneously.

If other non-ideal testing conditions exist, the facility must submit a source test protocol for review and approval prior to testing.

7.0 CALCULATIONS

Calculations should be carried out to at least one significant digit beyond that of the acquired data and then should be rounded off after final calculation to three significant digits for each run. All rounding off of numbers should be in accordance with the ASTM E380-82 procedures.

7.1 HEATING VALUE

If heating value of the fuel is measured continuously using an instrument, such as a calorimeter, this value shall be reported in units of Btu/dscf for gaseous fuels, and Btu/gal for liquid fuels. The density (lb/gal) shall also be reported for liquid fuels.

Note that if the heating value is determined from a fuel gas composition analysis, the calculated heating value must be corrected to a standard temperature of 60 degrees Fahrenheit. The following methods may be used for fuel analysis:

Gaseous Fuel	ASTM D1945
Refinery Fuel	ASTD D1946
Liquid Fuel	ASTM D2382-88

Other ASTM methods may be used upon approval by the District. Fuel analysis must be conducted by an independent laboratory.

7.2 F-FACTOR CALCULATION

For fuels not listed in Table 19-1 (EPA 40 CFR 60 Appendix A, Method 19), an F-Factor may be determined using procedures described in paragraph 3.2 of Method 19. An explanation of the analytical procedures used for determining the ultimate analysis for the fuel shall be provided with the test report.

The F-Factor shall not be used in cases where enriched oxygen is used, non-fuel sources of carbon dioxide are present (e.g., lime kilns and calciners), and the oxygen content of the stack is 19 percent or greater.

7.3 EMISSION RATE OF NO_x AND CO

7.3.1 CONCENTRATION CORRECTIONS

All emission concentrations determined using District Method 100.1 or hand-held portable analyzers shall be corrected according to Section 2.7 of District Method 100.1, or per Section 8.1 of CTM-030.

For hand-held portable analyzers, NO_x data less than 10 ppm shall be reported as "< 10 ppm", and then corrected per Section 8.1 of CTM-030. However, if the correction reduces the concentration value (i.e. < 9 ppm), the result shall be reported as "< 10 ppm".

7.3.2 CONCENTRATION ADJUSTMENT OF PORTABLE ANALYZER DATA

In addition to the correction provided in Section 7.3.1 above, the measured concentrations shall be adjusted as follows:

$$\text{Adjusted Concentration} = \text{Raw Concentration} + (0.01) (T_{\text{Err}}) (\text{High Span Conc.})$$

Where:

Raw Concentration = Averaged 15 minute data of the NO_x, CO, or O₂ gas measured during stack traverse;

High Span Conc. = the high span calibration gas used to calibrate the NO_x, CO, or O₂ analyzer;

$$T_{\text{Err}} = \sqrt{E_{\text{Cal}}^2 + E_{\text{Lin}}^2 + E_{\text{Stabl}}^2}$$

E_{Cal} = higher of the pre-test and post-test calibration errors (%);

E_{Lin} = higher of the pre-test and post-test linearity errors (%); and,

E_{Stabl} = the stability check error (%), adjusted for a 30-minute period.

7.3.3 NO_x OR CO EMISSION (Concentration at 3% O₂)

Compute using the following:

$$N = (P) \frac{20.9 - 3}{20.9 - \%O_2}$$

Where:

N = NO_x or CO Concentration (ppm) corrected to 3% O₂;

P = NO_x or CO Concentration in flue gas, (ppm), dry basis, adjusted per Sections 7.3.1 and 7.3.2, as applicable;

%O₂ = Oxygen concentration (%), dry basis, measured in the flue gas and adjusted per Sections 7.3.1 and 7.3.2, as applicable.

7.3.4 NO_x OR CO EMISSION (lb/MMBtu)

Compute using the following:

$$E = (K) \frac{20.9}{20.9 - \%O_2} (F_d) (P)$$

Where:

E = NO_x or CO Emission Rate (lb/MMBtu);

K = 1.195×10^{-7} for NO_x, and 7.27×10^{-8} for CO;

%O₂ = Oxygen concentration (%), dry basis, measured in the flue gas and adjusted per Sections 7.3.1 and 7.3.2, as applicable.

F_d = Oxygen based F-Factor, dry basis; and,

P = NO_x or CO Concentration in flue gas, (ppm), dry basis, adjusted per Sections 7.3.1 and 7.3.2, as applicable.

8.0 HIGH OXYGEN CONCENTRATIONS

Because the oxygen correction term is in the denominator of the equations presented in Section 7.3.3 and 7.3.4, errors in oxygen measurements at concentrations approaching 20.9% are greatly magnified in the calculation of NO_x or CO emissions. Hence, an alternative procedure for accurately determining NO_x and CO emissions at high oxygen concentrations is necessary. This section defines a procedure for determining Rule 1146 compliance (in units of lb/MMBtu) at stack oxygen concentrations exceeding 19%.

8.1 EMISSION SAMPLING

The procedures outlined in Section 4.2 shall be strictly followed. Sampling shall be conducted over a continuous period of not less than 30 minutes.

8.2 FLOW DETERMINATION

8.2.1 VELOCITY TRAVERSES

The exhaust flow rate shall be determined in order to calculate the pollutant mass emission rate. For stacks, or exhaust pipes, without existing sampling ports, determine the best sampling port location using SCAQMD Method 1.1. In order to satisfy Method 1.1, the exhaust pipe may need to be lengthened. If so, a temporary alteration may be made by attaching to the exhaust pipe an extension of the same size and shape as the original pipe, where the connecting seam does not cause a flow disturbance and is leak free.

Traverse sampling is necessary for the determination of stack gas velocity. The number of traverse points must be determined in accordance with SCAQMD Method 1.1 for stacks with diameters greater than 12 inches and SCAQMD Method 1.2 for stacks with diameters 4 to 12 inches. All traverse points, as determined by Method 1.1 or 1.2, must be used for measuring stack gas velocity.

If conditions such as negative flow and/or severe cyclonic flow exist that reduce the accuracy and precision of velocity measurements, an alternative location must be chosen where this condition does not exist.

8.2.2 STACK FLOW RATE

To determine the average stack flow rate in the stack, use one of the following methods:

Method	Description	Usage	Limitations
Method 2.1	Determination of Stack Gas Velocity and Volumetric Flow Rate	Most commonly used. For stacks greater than 12 inches in diameter and velocity head greater than 0.05 in. H ₂ O. ¹	Cyclonic flow will result in inaccurate velocity measurements. ²
Method 2.2	Direct Measurement of Gas Volume Through Pipes and Small Ducts	For pipes or ducts where velocity traverse cannot be performed due to physical limitations, i.e., low flow and size restrictions.	For gas temperature range of 32 to 120 °F.
Method 2.3	Determination of Gas Velocity and Volumetric Flow Rate From Small Stacks or Ducts	For velocity measurement in stacks 4 to 12 inches in diameter using a standard Pitot tube.	Standard Pitot tube may pose a problem in high moisture-laden or high particulate-laden gas streams.

¹ If the velocity head is below 0.05 in. H₂O, the source tester may propose an alternate method for measuring low velocity, such as, using a flow measuring device manufactured especially for low velocity. Use of such instruments must be approved prior to source testing.

² If cyclonic flow is present, perform the velocity traverse as specified in the District Source Test Manual Chapter X, Section 1.2.

The velocity traverse shall be performed simultaneously with concentration testing. If pulsating flow is detected during the velocity traverses, use a dampening method as described in Chapter X Section 9 of the SCAQMD Source Test Manual.

8.2.3 EXHAUST GAS MOLECULAR WEIGHT DETERMINATION

Carbon dioxide (CO₂) and O₂ concentrations are needed to determine the molecular weight of the exhaust gas in order to calculate the standard volumetric flow rate. These gases shall be measured using Method 100.1. All testing procedures and QA/QC requirements must satisfy those as stated in the method. Molecular weight sampling shall be performed simultaneously with the NO_x and CO sampling.

8.2.4 MOISTURE DETERMINATION

In conjunction with the velocity traverses, conduct moisture sampling using Method 4.1. A single sampling point near the stack center is acceptable for moisture sampling. If low moisture is suspected (< 5%), increase the sampling time to acquire at least twice the required sample volume.

8.3 **HEAT INPUT**

8.3.1 HEATING VALUE

The higher heating value of the fuel shall be determined according to the procedures in Section 7.1 of this protocol.

8.3.2 FUEL USAGE

The fuel usage shall be determined using a totalizing fuel meter by calculating the difference between the facility's fuel meter volume readings (in cubic feet) prior to, and following the Method 100.1 sampling period. This fuel usage must be corrected to standard pressure (14.7 psia) and temperature (60°F). The time between fuel meter readings (in minutes) must also be accurately recorded.

Meter calibration data must be submitted with the final report.

8.4 CALCULATIONS

8.4.1 POLLUTANT MASS EMISSIONS

The NO_x and CO mass emissions rate (in units of lb/hr) shall be determined according to the following equation:

$$E_R = (1.583 \times 10^{-7}) (MW) (\text{Flow}) (P)$$

Where:

E_R = NO_x or CO Emission Rate (lb/hr);

MW = Molecular weight, 46 for NO_x, and 28 for CO;

Flow = Exhaust stack flow rate (dscfm), determined using calculation procedures in District Method 2.1;

P = NO_x or CO Concentration in flue gas, (ppm), dry basis, corrected according to Section 2.7 of District Method 100.1.

8.4.2 HEAT INPUT CALCULATION

The heat input shall be calculated from data collected in Section 8.3 as follows:

$$HI = (U) (HHV) (60) (10^{-6})$$

Where:

HI = Heat input rate (MMBtu/hr);

U = Fuel input rate (scf/min), calculated by dividing the fuel usage (Section 8.3.2) by the minutes between fuel meter readings;

HHV = Higher heating value of the fuel, determined in Section 8.3.1.

8.4.3 NO_x OR CO EMISSION CALCULATION

The NO_x or CO emissions, for Rule 1146 compliance determination is calculated by:

$$C_{R1146} = (E_R) / (HI)$$

Where:

C_{R1146} = NO_x or CO emissions (lb/MMBtu) for Rule 1146 compliance determination;

E_R = NO_x or CO Emission Rate (lb/hr), determined from Section 8.4.1;

HI = Heat input rate (MMBtu/hr), determined from Section 8.4.2.

9.0 QUALITY ASSURANCE/ QUALITY CONTROL (QA/QC)

The test and analytical equipment shall meet all QA/QC requirements cited throughout this Protocol, and where not cited, shall be calibrated as a minimum within the requirements set forth in SCAQMD Source Test Methods Chapter III, Calibrations. The facility's equipment fuel meter shall be calibrated in accordance with procedures outlined in EPA 40 CFR Part 75, Appendix E.

9.1 PRESSURE MEASUREMENT

Pressure measurement instruments shall have an error no greater than the following values:

<u>Measurement</u>	<u>Accuracy</u>	<u>Precision</u>
Gas and Liquid Pressure	$\pm 5\%$ (of Reading)	$\pm 2\%$ (of Range)

9.2 TEMPERATURE MEASUREMENTS

Temperature measuring instruments shall have an error no greater than the following values:

<u>Accuracy</u>	<u>Precision</u>
$\pm 1.5\%$ of Reading	$\pm 1.5\%$ of Reading

9.3 FUEL GAS FLOW

The gas meter reading shall be corrected to dry standard conditions (60°F, 14.7 psia), and adjusted for instrument calibration. Calibration curves shall be used if required. The overall accuracy of gas flow measurements shall be equal to or less than 2% of the actual value.

9.4 NON-GASEOUS FUEL USAGE MEASUREMENTS

The accuracy of non-gaseous fuel use measurements, using calibration curves if required, shall be equal to or less than $\pm 2\%$ of the actual value. Calibration data shall be submitted with the test report.

9.5 TIME

The elapsed time measurement shall be measured with an instrument that is accurate within ± 0.5 seconds per hour.

10.0 REPORT

10.1 REPORT INFORMATION

The following information is to be included in the test report:

1. All printouts, must be included in the final report and must be clearly identified as to:

-location/source	-range changes
-operator initials	-range of measurement
-date/running times	-calibrations
-actual test interval	-cal gas concentration/cyl. no.
-contaminant/diluent	-range of calibration
2. A summary of the Source Test results.
3. A brief process description. Indicate equipment operation during testing; as well as any other information which may influence the final report.
4. A simple schematic diagram of the process, showing the sampling location, with respect to the upstream and downstream flow disturbances.
5. The sampling and analytical procedures. Be specific about all aspects of sampling and analysis. Include diagrams of test equipment and methods.
6. Complete raw field data, including production data indicative of the testing interval, analyses, and the test results (show all calculations).
7. Calibration data regarding all sampling and measuring equipment utilized during testing (see District Source Testing Manual, Chapter III or "Quality Assurance Handbook For Air Pollution Measurement Systems", Vol. III, U.S. EPA-600/4-77-0276).

10.2 REPORT FORMAT

The information in the final report shall be formatted as follows:

- I. Table of Contents
- II. Executive Summary
- III. Introduction
- IV. Equipment Description Including Fuel Meter(s), if applicable. This section shall include a statement that verifies acceptability of the method test location and the operating condition during the test.
- V. Test Critique
- VI. List of Sampling and Analytical Methods Used. This section shall include a list of the test methods used. Do not include copies or descriptions of the source test methods if the methods were adhered to as written. If exceptions were made to the methods, submit only an explanation of the exceptions.
- VII. Appendices
 - A. SCAQMD Method Results
 - B. Portable Analyzer Results
 - C. Schematics of Stack Sampling Locations
 - D. Field Data Sheets - testing method and process data sheets shall be compiled in separate sections
 - E. QA/QC
 - F. Laboratory Analytical Data
 - G. Calibration Data and Calculations
 - H. Calibration Gas Certificates
 - I. Chain of Custody Information (as necessary)
 - J. Process Operating Data Including Fuel Usage During Test
 - K. Calculations Using Fuel F-Factor (if necessary)
 - L. Additional QA/QC for Non-LAP Approved Source Test Laboratories (if necessary)

ATTACHMENT I

**ADDITIONAL QA/QC FOR
NON-LAP APPROVED
SOURCE TEST LABORATORIES**

GENERAL INFORMATION

CERTIFICATE OF NON CONFLICT OF INTEREST

**SOURCE TEST LABORATORIES
GENERAL INFORMATION**

- 1) Name of Laboratory: _____
- 2) Street Address: _____
- 3) P.O. Box: _____
- 4) City: _____ State: _____
- 5) Zip: _____ County: _____
- 6) Phone No.: _____ Fax No.: _____
- 7) e-mail Address: _____
- 8) Individual Owner's Name (if applicable): _____
- 9) Name of Parent Company (if applicable): _____
- 10) Authorized Representative responsible for ensuring that the laboratory complies with the conditions and criteria for testing approval: _____
Title: _____
- 11) Technical Director or Manager of Laboratory: _____
Title: _____
- 12) List of methods for which testing was conducted:

- 13) **Organization**
Please attach a personnel organization chart. Please show all management staff and personnel who will be performing work under LAP, including technicians, student workers, and temporaries, and describe their duties. Be sure to identify the following functions however named:

Authorized contact. This person has the authority to initiate or terminate the Approval, or to change its scope. This person has the authority to commit to meeting LAP requirements.

Technical director. This person has overall technical responsibility for the test facility.

QA Officer. This person guarantees that the data quality meets test plan or project specifications, has overall responsibility to review implementation of QA policies, takes corrective action, and is not involved in production of data. This person keeps complaint records for one year.

Supervisor(s), project manager(s), or team leader(s): These people have day to day oversight of work and data quality.

14) **Personnel**

Please provide a curriculum vita or resume for all management staff and personnel who perform LAP work, including technicians, student workers, temporaries etc. Specifically address education, relevant experience, special training, title and normal duties.

15) **Integrity**

Please attach a copy of your policies on individual conflict of interest and sign the conflict of interest form included with this Application.

Please attach a copy of your policies on data falsification and data selection/ rejection, and a copy of your policies on QC.

16) **Chain of Custody, Evidence Retention**

Please attach your Chain of Custody policy and provide a copy of the Chain of Custody form. Please describe your procedures for identifying, receiving, and storing samples, and for assuring the security of samples, data, equipment etc. Please describe your procedures for authorizing sample and data disposal.

17) **Equipment, Instruments, Reagents and Standards**

Please attach a copy of your procedure for determining equipment, instrument, reagent, and standard suitability and for maintaining traceability.

18) **Documentation**

Please attach a copy of your procedure for generating and validating data forms, formats, procedures, logbooks, SOPs, spreadsheets, data, reports, etc. and for maintaining revision control.

19) **Quality Assurance**

Please describe how deviations or problems in facilities, data integrity, chain of custody, security, methods, equipment, instruments, reagents, standards, quality control, documents, reports, etc are detected, flagged and corrected. Describe the role of each staff level for maintaining data quality, and describe the duties and authority of the Quality Assurance officer in detail.

20) Please include with this application:

- 20a. General description of the laboratory facility, physical location, and scope of operation. This should include size, remote sites or sub-facilities, mobile units and major equipment. This should also include an indication of the amount of testing that is performed in fields similar to the one performed in this test report. Provide a brief overview of other testing services offered by the laboratory.
- 20b. Year laboratory established.
- 20c. List of major equipment and instrument.
- 20d. Standard Operating Procedures (SOPs) for each method used (except portable analyzer methods). These SOPs should include step-by-step instructions for performance of the particular method.
- 20e. Conflict of interest statement (example supplied herewith).
- 20f. Signed agreement to participate in site/ proficiency audit for Rule 1146 testing.

Authorized Signature

Date

CONFLICT OF INTEREST STATEMENT

1. The test facility shall have no financial interest in the company or facility being tested, or in the parent company or any subsidiary thereof.
2. The company or facility being tested, or parent company or subsidiary thereof, shall have no financial interest in the test facility.
3. Any company or facility responsible for the emission of significant quantities of pollutants to the atmosphere, or parent company or subsidiary thereof, shall have financial interest in the test facility.
4. The test facility shall not be in partnership with, own or be owned by, in any part or in full, the contractor who has provided or installed equipment (basic or control), or monitoring systems, for the company being tested.

I certify that the test facility meets the above criteria to the best of my knowledge and belief.

Signature, authorized contact

Date

ATTACHMENT II

**DISTRICT METHOD 100.1 GENERAL REQUIREMENTS AND
NO₂ TO NO CONVERTER EFFICIENCY CHECK PROCEDURE**

GENERAL CONTINUOUS GAS MONITORING REQUIREMENTS

The District requires continuous gas monitoring equipment employing sample extraction and conditioning, and electronic detection, to be conducted strictly according to District Method 100.1, with the emphasis upon representativeness, documentation, and quality assurance. This includes, in part:

1. Gas analyzers must meet minimum acceptable standards for method of detection, sensitivity, noise, precision, linearity, and interference (see TABLE 100.1-1 for details). Also, the gas sample extraction and conditioning equipment (probe, filter, pump, conditioner, connective plumbing, etc., and data acquisition and logging equipment shall meet minimum acceptable specifications, as described in Method 100.1.
2. It is recommended that the entire sampling system for continuous gas monitoring instruments should be leak checked before and after each test run by evacuating the system to a minimum of 20 in. Hg vacuum, and plugging for a period of 5 minutes. The resultant loss of vacuum can not exceed 1 in. Hg during this period.
3. Calibration of all analyzers must be accomplished at zero, mid span (40-60% of full scale range), and high span (80-95% of full scale range). The lowest practicable range should be selected for monitoring, so that the measured emission values are within 20-95% of the range. If a significant amount of the data are outside of this range, the data may be rejected, depending upon the application.
4. The calibration gases must be certified according to EPA Protocol Number 1, or certified to an analytical accuracy of $\pm 1\%$ and be NIST traceable (except cal gases used for system bias check), following EPA-600/R93/224, "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards", TABLE 2-2. Superblend or multi-component blend gas recertifications are based upon the individual component(s) with the shortest recertification date.
5. Field calibrations employing gas dilution systems (mass flow or critical orifice) must be performed in accordance with EPA Method 205, "Gas Dilution Verification Protocol", or available "Draft" SCAQMD Method, and supported with appropriate documentation. A 5-point check is required.
6. A calibration error check, and zero/span drift check must be performed before and after each test run. Calibration error must be less than $\pm 2\%$ of the range of measurement for zero, mid, and high range calibration gases. Zero/span drift must be less than $\pm 3\%$ of the range of measurement.
7. A system bias check must be performed before and after each test run by alternately introducing cal gases to the entire sampling system, then to the gas analyzer(s), for comparison. The difference can not exceed $\pm 5\%$ of the analyzer range.
8. Semi-annual analyzer certifications consisting of linearity plot, calibration curve, response time, and interference response must be furnished with the other calibrations to satisfy Q/A documentary requirements.
9. NO_x measurement must be performed in the NO_x mode of the analyzer. An NO_2 to NO converter is required if NO_2 constitutes 5% or more of the total NO_x in the sample stream, or

the rule or permit condition requires "NO_x" monitoring. The NO₂ to NO converter must be at least 90% efficient (use the NO₂ to NO converter efficiency procedure). The converter should be high temperature (650°C) stainless steel, if no NH₃ is present. If NH₃ is present in the sample stream, then a low temperature (350°C) molybdenum catalyst must be used in the converter. This check must be done at the beginning of the test.

10. The connective tubing from the probe to the sample conditioner must be heated above the dewpoint and the dewpoint reported. The sample conditioner must be able to maintain a dewpoint temperature of 37°F or less.
11. Data recorder resolution must be at least 0.5% of the range of measurement. A data point for each contaminant/diluent monitored must be recorded at least once/minute. Analog chart recorders must have a minimum 10-inch chart width, with 100 minor divisions.
12. All facets of testing must be continuously recorded. This includes the 3-point calibration, system bias, calibration error, and zero/span drift checks, which must precede and conclude each test run.
13. All chart traces, or digital printouts, must be included in the final report and must be clearly identified as to:
 - location/source
 - operator initials
 - date/running times
 - actual test interval
 - contaminant/diluent
 - range changes
 - range of measurement
 - calibrations
 - cal gas concentration/cyl. no.
 - range of calibration
14. When more than one gas trace is shown on a chart, the individual traces must be distinguishable by color coding or some other means (original charts may be submitted, and returned following evaluation). If a gas measurement range has been "offset" from zero, or zero has been "transposed" to the right side of the recorder chart, it must be clearly identified. this offset should not be more than 5-small divisions of the chart. This data must be corrected using the ratio of the offset.
15. Gaseous measurements must be conducted a minimum of 15 continuous minutes at each load or specified condition, after the readings have stabilized (RECLAIM reference method sampling requires a minimum of 30 continuous minutes at each load for RATAs).
16. Sampling locations not meeting the minimum site selection standards for District Method 1 must be tested for absence of stratification. (A gaseous constituent concentration profile differing more than 10% between any two monitoring points within the same cross-sectional plane of a stack or duct indicates stratification.) If stratification is present, and alternate approved site selection or modification is not possible, then special monitoring (see Chapter X of the District Source Testing Manual) will be required.

NO₂ TO NO CONVERSION TEST PROCEDURE
(Alternative to O₃ Titration Method-40 CFR 50.1, Appendix F)

1. NO_x Analyzer Requirements

- a. Full span range 0-20 ppm or 0-25 ppm
- b. Equipped with NO and NO_x modes

2. Auditing Gas Requirements

- a. NO₂ in air (or N₂): Use NO₂ in air for a stainless steel converter.
- b. Concentration of NO₂: 15 to 18 ppm ± 0.5 ppm (*C_o, ppm*)
- c. Recertification: An audit gas should be recertified after six months.

3. Calibration Gas Requirements

- a. Concentration: NO (17 to 19 ppm) with less than 0.1 ppm NO₂ - High Span
NO (10 to 13 ppm) with less than 0.1 ppm NO₂ - Mid Span
- b. Zero Gas: High purity N₂

4. Calibration of Analyzer:

- a. Calibrate NO mode with the NO calibration gases.
- b. Calibrate NO_x mode with the same gases without any gain adjustment.
- b1. If the analyzer is equipped with two independent gain adjusting circuits, skip 4.b., then repeat 4.a. for the NO_x mode.

5. Conversion Efficiency (CE) Test

- a. Analyze the audit gas with NO mode. Read and standardize concentration. (*C₁, ppm*)
- b. Analyze the audit gas with NO_x mode. Read and standardize concentration. (*C₂, ppm*)

6. Calculation for Conversion Efficiency:

$$\%CE = \frac{|C_2 - C_1|}{C_0} \times 100$$

7. Criteria for Acceptability of CE

- a. %CE must be larger than 90%.
- b. C₁ must be less than 5 % of total NO_x (NO + NO₂) in the NO₂ audit gas (Section 2b).

NOTE: NO₂ audit gas concentration of higher value than what is specified in *Section 2*. may be required where NO₂ present in the exhaust gas being measured is greater than 30 ppm. Select the NO₂ gas within 10% of the expected NO₂ concentration in the exhaust.

ATTACHMENT III

**EPA CONDITIONAL TEST METHOD
CTM-030**